

# BLZZRD™

## PORTABLE SAMPLER

### Installation and Operation Guide



Part #69-2973-063

Revision D, December 2022

Copyright © 2022. All rights reserved, Teledyne ISCO

Use and Disclosure of Data: Information contained herein is classified as EAR99 under the U.S. Export Administration Regulations. Export, reexport or diversion contrary to U.S. law is prohibited.



## Foreword - Water and Wastewater Products

This instruction manual is designed to help you gain a thorough understanding of the operation of the equipment. Teledyne ISCO recommends that you read this manual completely before placing the equipment in service.

Although Teledyne ISCO designs reliability into all equipment, there is always the possibility of a malfunction. This manual may help in diagnosing and repairing the malfunction.

If a problem persists, call or e-mail Teledyne ISCO technical support for assistance. Simple difficulties can often be diagnosed over the phone. For faster service, please have your serial number ready.

If it is necessary to return the equipment to the factory for service, please follow the shipping instructions provided by technical support, including the use of the Return Merchandise Authorization (RMA) specified. Be sure to include a note describing the malfunction. This will aid in the prompt repair and return of the equipment. **No item may be returned for service without a Return Merchandise Authorization (RMA) number issued by Teledyne.**

Teledyne ISCO welcomes suggestions that would improve the information presented in this manual or enhance the operation of the equipment itself.

**Teledyne ISCO is continually improving its products and reserves the right to change product specifications, replacement parts, schematics, and instructions without notice.**

### Contact Information

#### *Customer Service*

Phone: (800) 228-4373 (USA, Canada, Mexico)  
(402) 464-0231 (Outside North America)  
Fax: (402) 465-3022  
Email: [isco.orders@teledyne.com](mailto:isco.orders@teledyne.com)

#### *Technical Support*

Phone: Toll Free (866) 298-6174 (Samplers and flowmeters)  
Email: [iscowatersupport@Teledyne.com](mailto:iscowatersupport@Teledyne.com)

Return equipment to: 4700 Superior Street, Lincoln, NE 68504-1398

#### *Other Correspondence*

Mail to: P.O. Box 82531, Lincoln, NE 68501-2531

**Warranty and Operation Manuals can be found on our website at:**

[www.teledyneisco.com](http://www.teledyneisco.com)



*BLZZRD Safety  
Information*

The Teledyne ISCO BLZZRD Transportable Sampler is a “definite purpose” device, intended for use only with compatible Teledyne ISCO equipment. Do not use this product with any other manufacturers’ equipment, or for any other purpose. Use for any purpose not described in this manual could cause personal injury or property damage.

*Electrical Requirements*

The sampler requires 12 VDC at 6 Amperes. The DC power input is through the controller back panel 12V input ports. The primary power port (3-pin) takes priority over the secondary power port (2-pin). If power is removed from the primary port, the secondary port will be engaged after a momentary switch-over period, resuming normal operation. Battery back-up configurations should connect the battery to the secondary input. If only one 12V source is connected, it should be connected to the primary port.

The Teledyne ISCO-supplied DC connect cables provide over-current protection through the use of an in-line 8A, 32V fuse. The sampler may also be powered by 90 to 264 VAC, 47 to 63 Hz, 2 Ampere power source using the attached AC line cord. The internal AC power supply is protected by an internal 4A 250V input fuse. Additionally, the controller circuitry is protected by an internal 5.0 ampere PTC (Positive Temperature Coefficient) device.

 **WARNING**

**Electrocution hazard. Never manipulate electrical switches or power connections with wet hands or when your feet are in contact with water.**

 **WARNING**

**AC electrical power must meet the applicable electrical code requirements for your installation and must be provided with an earth ground connection. If necessary, consult with a certified electrician to ensure that AC power is provided in accordance with the local electrical code.**

*AC Powered BLZZRDs*

The AC power cord of this device is equipped with a three-prong grounding plug designed to mate with a grounded power outlet. Grounding minimizes the possibility of electrical shock.

It is the user’s responsibility to ensure that the AC power source is properly grounded. If in doubt, have the outlet checked by a qualified electrician.

If the available AC power outlet only accepts two prongs, or if it is determined that the outlet is improperly grounded, the outlet must be replaced by a qualified electrician before attempting to power this device.

 **WARNING**

**Never modify the AC power cord or use a three-prong to two-prong adapter.**

 **WARNING**

If this device's power cord is frayed or otherwise damaged, discontinue its use immediately. The AC power cord is not user-serviceable and must be returned to the factory for repair or replacement.

*General Warnings*

Before installing, operating, or maintaining this equipment, it is imperative that all hazards and preventive measures are fully understood. While specific hazards may vary according to location and application, take heed in the following general warnings:

 **WARNING**

This instrument has not been certified for use in "hazardous locations" as defined by the National Electrical Code.

 **WARNING**

Avoid hazardous practices! If you use this instrument in any way not specified in this manual, the protection provided by the instrument may be impaired; this will increase your risk of injury.

 **AVERTISSEMENT**

Éviter les usages périlleux! Si vous utilisez cet instrument d'une manière autre que celles qui sont spécifiées dans ce manuel, la protection fournie de l'instrument peut être affaiblie; cela augmentera votre risque de blessure.

Additional safety information can be found in Appendix C.

*Hazard Severity Levels*

This manual applies *Hazard Severity Levels* to the safety alerts. These three levels are described in the sample alerts below.

 **CAUTION**

Cautions identify a potential hazard, which if not avoided, may result in minor or moderate injury. This category can also warn you of unsafe practices, or conditions that may cause property damage.

 **WARNING**

















Warnings identify a potentially hazardous condition, which if not avoided, could result in death or serious injury.

 **DANGER**

**DANGER** – limited to the most extreme situations to identify an imminent hazard, which if not avoided, will result in death or serious injury.

*Hazard Symbols*

The equipment and this manual use symbols used to warn of hazards. The symbols are explained below.

<b>Hazard Symbols</b>	
<b>Warnings and Cautions</b>	
	The exclamation point within the triangle is a warning sign alerting you of important instructions in the instrument's technical reference manual.
	The lightning flash and arrowhead within the triangle is a warning sign alerting you of "dangerous voltage" inside the product.
 	Pinch point. These symbols warn you that your fingers or hands will be seriously injured if you place them between the moving parts of the mechanism near these symbols.
<b>Symboles de sécurité</b>	
	Ce symbole signale l'existence d'instructions importantes relatives au produit dans ce manuel.
	Ce symbole signale la présence d'un danger d'électocution.
 	Risque de pincement. Ces symboles vous avertit que les mains ou les doigts seront blessés sérieusement si vous les mettez entre les éléments en mouvement du mécanisme près de ces symboles
<b>Warnungen und Vorsichtshinweise</b>	
	Das Ausrufezeichen in Dreieck ist ein Warnzeichen, das Sie darauf aufmerksam macht, daß wichtige Anleitungen zu diesem Handbuch gehören.
	Der gepfeilte Blitz im Dreieck ist ein Warnzeichen, das Sei vor "gefährlichen Spannungen" im Inneren des Produkts warnt.
 	Vorsicht Quetschgefahr! Dieses Symbol warnt vor einer unmittelbar drohenden Verletzungsgefahr für Finger und Hände, wenn diese zwischen die beweglichen Teile des gekennzeichneten Gerätes geraten.
<b>Advertencias y Precauciones</b>	
	Esta señal le advierte sobre la importancia de las instrucciones del manual que acompañan a este producto.
	Esta señal alerta sobre la presencia de alto voltaje en el interior del producto.
 	Punto del machacamiento. Sus dedos o manos serúyos dañados si usted los coloca entre las piezas móviles cerca de estos símbolos.





# BLZZRD™

## Portable Refrigerated Sampler

---

### *Table of Contents*

#### **Section 1 Introduction**

---

1.1 About this Manual . . . . .	1-1
1.2 About 700 Series Modules . . . . .	1-2
1.3 SDI-12 Sondes . . . . .	1-2
1.4 Telephone Communication . . . . .	1-3
1.4.1 6712 LTE Modem . . . . .	1-3
1.4.2 Software Compatibility . . . . .	1-3
1.5 Memory to Store Monitoring Data . . . . .	1-4
1.6 Pump Requirements . . . . .	1-5

#### **Section 2 Installation/Preparation**

---

2.1 Receiving Inspection . . . . .	2-1
2.1.1 Case Ventilation . . . . .	2-1
2.2 Installation . . . . .	2-1
2.3 Positioning a Sampler . . . . .	2-2
2.4 Inspecting the Pump Tube . . . . .	2-4
2.4.1 Pump Tube Warning . . . . .	2-4
2.4.2 Visual Inspection . . . . .	2-4
2.5 Installing the Bottle Kit . . . . .	2-5
2.5.1 Installing the 14-Bottle Kit . . . . .	2-5
2.5.2 Installing the 4-Bottle Kit . . . . .	2-8
2.5.3 Installing 2.5 Gallon Composite Bottle Kits . . . . .	2-10
2.5.4 Installing the 5 Gallon Composite Bottle Kit . . . . .	2-13
2.6 Installing a Power Source . . . . .	2-14
2.6.1 Installing the Battery Box . . . . .	2-14
2.6.2 AC Power Sources . . . . .	2-15
2.6.3 12 VDC Power Sources . . . . .	2-15
2.6.4 Battery Recommendations . . . . .	2-15
2.7 Attaching the Suction Line . . . . .	2-16
2.7.1 Attaching Suction Line to Pump Tube . . . . .	2-16
2.8 Attaching a Strainer . . . . .	2-17
2.8.1 Alternative to Strainers . . . . .	2-17
2.8.2 Tips for Routing Suction Line and Strainer . . . . .	2-18
2.8.3 Pressurized Lines . . . . .	2-18
2.9 Connecting Instruments . . . . .	2-19
2.9.1 Connecting 700 Series Modules . . . . .	2-19
2.9.2 Connecting External Instruments . . . . .	2-20
2.10 Programming the BLZZRD . . . . .	2-21
2.11 Locking the Sampler . . . . .	2-21
2.11.1 Program Lock . . . . .	2-21
2.12 Servicing the Sampler . . . . .	2-22
2.13 Recovering the Sampler . . . . .	2-23
2.14 Collecting Sample Bottles . . . . .	2-23

#### **Section 3 Programming Introduction**

---

3.1 Initial Operation . . . . .	3-1
---------------------------------	-----

3.1.1	Selecting a Language . . . . .	3-3
3.2	Using Menus and Entering Numbers . . . . .	3-3
3.2.1	Selecting Menu Options . . . . .	3-3
3.2.2	Entering Numbers . . . . .	3-3
3.3	Quick View Screens . . . . .	3-4
3.3.1	Paging Through Quick View Screens . . . . .	3-4
3.3.2	Changing Settings in a Quick View Screen . . . . .	3-5
3.4	Clock and Calendar . . . . .	3-6
3.5	Clock Start Time Menu Screens . . . . .	3-6
3.6	Site Description and Program Name Menu Screens . . . . .	3-7
3.7	Messages . . . . .	3-9
3.7.1	Information Messages . . . . .	3-9
3.7.2	Help Notes . . . . .	3-9
3.7.3	Warning Messages . . . . .	3-10
3.7.4	Operation Messages . . . . .	3-10
3.8	Menu Trees . . . . .	3-10
3.9	Memory . . . . .	3-12

## **Section 4 Standard Programming**

---

4.1	Switching between Standard and Extended Modes . . . . .	4-1
4.2	Selecting Language and Units of Length . . . . .	4-1
4.3	Programming Examples . . . . .	4-2
4.4	Pacing . . . . .	4-5
4.4.1	Time Pacing . . . . .	4-5
4.4.2	Flow Pacing . . . . .	4-5
4.4.3	Trigger Pacing . . . . .	4-6
4.5	Distribution . . . . .	4-6
4.5.1	Sequential . . . . .	4-8
4.5.2	Bottles Per Sample . . . . .	4-8
4.5.3	Samples Per Bottle . . . . .	4-8
4.5.4	Composite . . . . .	4-8
4.5.5	Continuous Sampling . . . . .	4-8
4.6	Start Times . . . . .	4-8
4.6.1	How Do Start Times Work? . . . . .	4-9
4.6.2	Sampler Enable/Disable . . . . .	4-10
4.7	Programming for 700 Series Modules . . . . .	4-10
4.8	Other Functions . . . . .	4-15

## **Section 5 Extended Programming**

---

5.1	Switching Between Standard and Extended Modes . . . . .	5-1
5.2	One-Part and Two-Part Programs . . . . .	5-1
5.3	Storage for Extended Programs . . . . .	5-2
5.3.1	Selecting a Stored Program . . . . .	5-2
5.4	Programming for 700 Series Modules . . . . .	5-4
5.5	Programming Examples . . . . .	5-4
5.5.1	Storm Water Runoff Sampling . . . . .	5-5
5.6	Suction Head . . . . .	5-11
5.7	Rinses and Retries . . . . .	5-11
5.8	Pacing . . . . .	5-12
5.8.1	Event Pacing . . . . .	5-12
5.8.2	Nonuniform Time Pacing . . . . .	5-14
5.8.3	Nonuniform Clock Time Pacing . . . . .	5-14
5.8.4	Nonuniform Interval Pacing . . . . .	5-14
5.8.5	Random Interval Pacing . . . . .	5-14
5.9	Distribution . . . . .	5-16
5.9.1	Sequential . . . . .	5-16

5.9.2	Bottles Per Sample	5-16
5.9.3	Samples Per Bottle	5-16
5.9.4	Composite	5-16
5.9.5	Multiple Bottle Compositing	5-16
5.9.6	Time Switching	5-16
5.9.7	Continuous Sampling	5-17
5.10	Flow Proportional Sample Volumes	5-19
5.10.1	Calculating Variable Sample Volume Settings	5-20
5.10.2	Calculations when Using an External Flow Meter	5-21
5.10.3	Calculations when Using a 700 Series Flow Module	5-22
5.11	Sampler Enable	5-23
5.11.1	Sampler Enable Responses	5-25
5.12	Pauses/Resumes	5-27
5.13	Running Programs	5-27
5.13.1	Run Time Screens	5-28
5.13.2	Interrupting a Running Program	5-28
5.14	Other Functions	5-28
5.15	Software Options	5-28
5.15.1	Liquid Detector Enable/Disable	5-28
5.15.2	Programming Style	5-29
5.15.3	Measurement Interval	5-29
5.15.4	Dual Sampler Mode	5-29
5.15.5	Bottle Full Detection	5-30
5.15.6	Display Backlighting	5-30
5.15.7	Event Marks	5-30
5.15.8	Pre-sample and Post-sample Purge Counts	5-31
5.15.9	Serial Output	5-32
5.15.10	Interrogator Connector Power	5-34
5.16	Hardware Setup	5-35
5.16.1	SDI-12 Sonde Setup	5-35
5.16.2	Rain Gauge Setup	5-35
5.16.3	I/O Pin Programming	5-36
5.16.4	Dialout Alarms	5-38
5.16.5	Digital Modem Text Messaging	5-38
5.16.6	Refrigerator Temperature	5-39

## Section 6 Running Programs

---

6.1	Starting a Program	6-1
6.2	Run Time Screens	6-1
6.2.1	Module Readings	6-2
6.3	How Does the Sampler Work?	6-3
6.3.1	A Typical Sampling Cycle	6-4
6.3.2	Measuring Sample Volume	6-4
6.4	How Does the Refrigerator Work?	6-5
6.5	Interrupting a Running Program	6-5
6.5.1	Stop Program	6-6
6.5.2	Resume Program	6-6
6.5.3	View Data	6-6
6.5.4	Grab Sample	6-6
6.5.5	Pump Tube Alarm	6-6
6.5.6	Calibrate Volume	6-6
6.5.7	Cal/Adj Parameters	6-6
6.5.8	Adjust Pacing	6-6
6.5.9	Adjust Volume	6-7
6.5.10	Power Used	6-7
6.6	Manual Functions	6-7
6.6.1	Grab Samples	6-8

6.6.2 Calibrate Volume . . . . .	6-10
6.6.3 Operating the Pump . . . . .	6-12
6.6.4 Moving the Distributor Arm . . . . .	6-13
6.7 Reports . . . . .	6-14
6.7.1 Collecting Reports . . . . .	6-14
6.7.2 Viewing the Data . . . . .	6-14
6.7.3 Configuring Reports . . . . .	6-15
6.7.4 Sampling Results Report . . . . .	6-16
6.7.5 Combined Results Report . . . . .	6-16
6.7.6 Summary Report . . . . .	6-17
6.8 System IDs . . . . .	6-24

## **Section 7 SDI-12 Sondes**

---

7.1 Teledyne ISCO AQ700 Sonde . . . . .	7-1
7.2 SDI-12 Data Parameters . . . . .	7-2
7.3 Scanning for Sondes . . . . .	7-3
7.4 ISCO-Ready Sondes . . . . .	7-5
7.5 Non-ISCO-Ready SDI-12 Sondes . . . . .	7-6
7.6 Sonde Setup - Storing Parameter Data . . . . .	7-7
7.7 Sonde Calibration and Validation . . . . .	7-8
7.7.1 Calibration Screens . . . . .	7-8
7.7.2 Validation Screen . . . . .	7-9
7.8 Disconnecting Sondes . . . . .	7-9

## **Section 8 Remote Operation**

---

8.1 Computer Operation . . . . .	8-1
8.1.1 Menu Control . . . . .	8-2
8.1.2 External Program Control . . . . .	8-4
8.1.3 Remote Control of Sampler Keypad . . . . .	8-8
8.2 SMS Commands . . . . .	8-9

## **Section 9 Maintenance**

---

9.1 Maintenance Checklist . . . . .	9-1
9.2 Cleaning Guidelines . . . . .	9-1
9.2.1 Cleaning the BLZZRD Exterior . . . . .	9-1
9.2.2 Cleaning the Refrigerated Compartment . . . . .	9-2
9.2.3 Defrosting the Refrigerated Compartment . . . . .	9-2
9.2.4 Sample Bottles . . . . .	9-2
9.2.5 Cleaning or Replacing Wetted Parts . . . . .	9-3
9.3 Maintenance Screens . . . . .	9-3
9.3.1 Set Clock . . . . .	9-3
9.3.2 Pump Tube Alarm . . . . .	9-5
9.3.3 Internal Battery . . . . .	9-5
9.3.4 Diagnostics . . . . .	9-6
9.4 Replacing the Pump Tube . . . . .	9-8
9.4.1 Pump Tube Life . . . . .	9-9
9.5 Opening the Controller Case . . . . .	9-11
9.6 Replacing the Desiccant . . . . .	9-12
9.7 Replacing the Internal Battery . . . . .	9-13
9.8 Error Messages . . . . .	9-16
9.8.1 Fatal Error Message . . . . .	9-16
9.8.2 Pump Warning Message . . . . .	9-16

## **Appendix A Menu Flowcharts**

---

## Appendix B Material Safety Data Sheets

---

## Appendix C General Safety Procedures

---

C.1 Hazards . . . . .	C-1
C.2 Planning . . . . .	C-2
C.3 Adverse Atmospheres . . . . .	C-2
C.4 Entering Manholes . . . . .	C-2
C.4.1 Traffic Protection . . . . .	C-3
C.4.2 Removing the Covers . . . . .	C-3
C.4.3 Other Precautions . . . . .	C-3
C.4.4 Emergencies . . . . .	C-4
C.4.5 Field Equipment . . . . .	C-4
C.5 Lethal Atmospheres in Sewers . . . . .	C-4

## Appendix D Replacement Parts

---

D.1 Overview . . . . .	D-1
------------------------	-----

## Appendix E Accessories

---

E.1 Order Information . . . . .	E-1
---------------------------------	-----

## Appendix F Battery Selection Guide

---

F.1 Introduction . . . . .	F-1
F.2 Determining the Power Consumption . . . . .	F-1

## List of Figures

---

1-1 Software Update window . . . . .	1-4
2-1 14-Bottle kit . . . . .	2-6
2-2 Attaching the distributor arm and discharge tube . . . . .	2-6
2-3 14-Bottle adapter plate installed . . . . .	2-7
2-4 Carrier alignment . . . . .	2-7
2-5 14-Bottle configuration completed . . . . .	2-7
2-6 4-Bottle kits . . . . .	2-8
2-7 4-Bottle adapter plate installed . . . . .	2-9
2-8 4-Bottle configurations completed . . . . .	2-9
2-9 Composite 2.5 gallon glass kit . . . . .	2-10
2-10 Composite 2.5 gallon Nalgene (polyethylene) kit . . . . .	2-10
2-11 Attaching the composite tube guide and discharge tube . . . . .	2-11
2-12 Composite bottle adapter plate installed . . . . .	2-12
2-13 Composite 2.5 gallon bottle kits completed . . . . .	2-12
2-14 Composite 5 gallon bottle kit . . . . .	2-13
2-15 Composite 5 gallon bottle kit completed . . . . .	2-13
2-16 . . . . . Mounting holes for battery box	2-14
2-17 . . . . . Battery box strap installed	2-14
2-18 Battery box secured . . . . .	2-14
2-19 Attaching the suction line to the pump tubing . . . . .	2-17
2-20 Optional Password Protected Functions . . . . .	2-22
3-1 Menu Tree for Standard Programming . . . . .	3-11
4-1 Sample Distribution . . . . .	4-7
4-2 Start Time Diagram . . . . .	4-10
5-1 Sample Distribution . . . . .	5-18
5-2 Factors affecting flow-weighted variable-volume samples . . . . .	5-20

5-3	Variable-volume scenario	5-21
5-4	Start Time Diagram for Sampler Enable Settings	5-23
5-5	Event Mark Timing Diagram	5-31
6-1	Running a Program	6-2
6-2	Identifying sampler components	6-3
6-3	Taking a grab sample	6-8
6-4	Report: Program Settings	6-19
6-5	Report: Sampling Results	6-20
6-6	Report: Combined Results	6-21
6-7	Report: Combined Results (continued)	6-22
6-8	Report: Summary	6-23
7-1	Teledyne ISCO AQ700 Multi-Parameter Sonde	7-1
8-1	Press and hold ?? to connect to the sampler	8-2
8-2	Example list of remote menu commands	8-2
9-1	Align arrows when installing the bottom plate	9-2
9-2	Removing and replacing the pump tube	9-10
9-3	Placing the pump tube in the liquid detector	9-10
9-4	Controller mounting screw (1 of 4)	9-11
9-5	Desiccant location	9-12
9-6	Removing the desiccant bag	9-13
9-7	Main Circuit Board Connections	9-14
9-8	Lithium battery	9-15
A-1	Menu Tree for Standard Programming	A-1
A-2	Standard Programming: Programming Screens	A-2
A-3	Standard and Extended Programming: Start Times	A-3
A-4	Standard Programming: Quick View; Programming Screens	A-4
A-5	Standard Programming: Quick View; Start Times	A-5
A-6	View Report	A-6
A-7	Maintenance Screens	A-7
A-8	Manual Functions Screens	A-8
A-9	Extended Programming: Programming Screens	A-9
A-10	Extended Programming: Equipment Setup	A-10
A-11	Extended Programming: Pacing and Distribution	A-11
A-12	Extended Programming: Event Conditions	A-12
A-13	Extended Programming: Programmed Sampler Enable	A-13
A-14	Extended Programming: Quick View; Programming Screens	A-14
A-15	Extended Programming: Quick View; Equipment Set-Up	A-15
A-16	Extended Programming: Quick View; Pacing and Distribution	A-16
A-17	Extended Programming: Quick View; Programmed Sampler Enable	A-17
A-18	Extended Programming: Normal View; Software Options	A-18
A-19	Extended Programming: Quick View; Software Options	A-19
A-20	Extended Programming: Normal View; Hardware	A-20
A-21	Extended Programming: Quick View; Hardware	A-21
A-22	SDI-12 Sonde Screens	A-22
F-1	Amp-hours required to cool 6 liters of sample liquid from 20 °C (A <sub>c</sub> )	F-3
F-2	Amp-hours per day required to maintain sample liquid at 3 °C (A <sub>m</sub> )	F-3

## List of Tables

---

1-1 BLZZRD Sampler Features . . . . .	1-5
1-2 BLZZRD Construction Materials . . . . .	1-8
1-3 Technical Specifications for the BLZZRD Sampler . . . . .	1-9
1-4 Connector Specifications . . . . .	1-15
2-1 Selecting the Right Strainer . . . . .	2-18
2-2 Connecting External Instruments to the Sampler . . . . .	2-20
3-1 About the Keypad . . . . .	3-2
5-1 Serial Data Codes . . . . .	5-33
6-1 Reports: Program Events, Source Codes, and Error Codes . . . . .	6-17
7-1 SDI-12 Sonde Parameters . . . . .	7-3
8-1 BLZZRD Remote Menu Commands . . . . .	8-3
8-2 Serial Data Codes . . . . .	8-6
8-3 Remote Control of Sampler Keypad . . . . .	8-8
8-4 6712 LTE SMS Commands . . . . .	8-9
C-1 Hazardous Gases . . . . .	C-7
E-1 Accessories . . . . .	E-2

## List of Examples

---

Entering Numbers . . . . .	3-4
Using Quick View Screens . . . . .	3-5
Setting the Clock and Calendar . . . . .	3-6
Entering a Site Description . . . . .	3-8
Viewing Help Notes . . . . .	3-9
Standard Program –	
Sample every 15 Minutes, One Sample per Bottle . . . . .	4-2
Standard Program –	
Flow-Paced Sampling, Two Bottles Per Sample . . . . .	4-4
Start Times . . . . .	4-9
Programming with 730 Bubbler Flow Module Installed . . . . .	4-12
Programming with 750 Area Velocity Module Installed . . . . .	4-14
Selecting a Stored Extended Program . . . . .	5-3
Storm Water Runoff Sampling Program . . . . .	5-6
Event Paced Sampling . . . . .	5-13
Nonuniform Time Pacing . . . . .	5-15
Time Switching Distribution . . . . .	5-17
Sampler Enable Programming . . . . .	5-24
I/O Pin Programming . . . . .	5-37
Dialout Alarms . . . . .	5-38
Taking a Grab Sample . . . . .	6-9
Calibrating Sample Volume . . . . .	6-11
Operating the Pump Manually . . . . .	6-12
Moving the Distributor Arm . . . . .	6-13
Viewing Reports . . . . .	6-15
Configuring Reports . . . . .	6-16
Viewing System IDs . . . . .	6-24
Sonde Setup screens . . . . .	7-5
Setting the Clock . . . . .	9-3
Resetting the Pump Tube Alarm . . . . .	9-4
Resetting the Internal Battery Countdown Timer . . . . .	9-5
Self-diagnostics . . . . .	9-6





# BLZZRD™

## Portable Refrigerated Sampler

---

### *Section 1 Introduction*

The BLZZRD is a multi-bottle, sequential/composite water sampler featuring a transportable refrigeration system and an advanced, multifunction controller. Its unique design provides easy transport, quick setup, and on-site cooling—without reliance on ice or utility power. An optional Mobility Kit with pneumatic tires lets you wheel the sampler over rough terrain to remote sampling sites.

The BLZZRD is ideal for storm water, TMDL, pretreatment compliance, and other mobile sampling applications requiring accurate sample preservation, versatile control, and flexible, reliable data handling.

In the field, the sampler delivers 48 hours or more of refrigeration from a 12 volt deep cycle battery, which mounts conveniently on the optional wheeled cart. Its power-saving system begins cooling when the first sample is drawn. The sampler can wait patiently for days or weeks to collect event-triggered samples, dial out an alarm/notification on its optional modem, and then preserve the samples until they are retrieved. For even greater versatility, the sampler runs on either 12 VDC or AC line voltage.

The BLZZRD controller offers all of the control, expansion, and data management options available with Teledyne ISCO's 6700 Series Samplers, including:

- Plug-in flow measurement modules eliminate the need for a separate flow meter.
- Direct-connect rain gauge, and SDI-12 connection to multi-parameter sondes enable comprehensive runoff and non-point-source monitoring.
- High-capacity data logging and Teledyne ISCO's Flowlink® Software provide the tools for sophisticated analysis and reports.

#### **1.1 About this Manual**

This manual contains the information you will need to install, program, operate, and maintain the BLZZRD sampler. There are nine sections and six appendices:

- Section 1, *Introduction* lists the features and specifications of the BLZZRD.
- Section 2, *Installation / Preparation* steps you through installing the sampler and preparing it for use.
- Section 3, *Programming Introduction* describes features that are common to both standard and extended programming modes.

- Section 4, *Standard Programming* shows how to program your sampler using standard programming mode.
- Section 5, *Extended Programming* shows how to program your sampler using extended programming mode.
- Section 6, *Running Programs* describes how to run, pause, and stop programs, and activities after a program has run.
- Section 7, *SDI-12 Sondes* provides information on setting up sondes to work with the sampler.
- Section 8, *Remote Operation* explains how the sampler can be controlled remotely.
- Section 9, *Maintenance* covers the basic maintenance required to keep your sampler in top operating condition.

The appendices are:

- Appendix A, *Menu Flowcharts*
- Appendix B, *Material Safety Data Sheets*
- Appendix C, *General Safety Procedures*
- Appendix D, *Replacement Parts*
- Appendix E, *Accessories*
- Appendix F, *Battery Selection Guide*.

## **1.2 About 700 Series Modules**

The bay on the controller's side accepts any of Teledyne ISCO's 700 Series Modules. The 700 Series includes:

- 701 pH Parameter Module for monitoring pH and temperature.
- 720 Submerged Probe Module for monitoring a flow stream's level and flow rate with a submerged probe.
- 730 Bubbler Flow Module for monitoring a flow stream's level and flow rate with a bubbler system.
- 750 Area Velocity Module for monitoring a flow stream's level, velocity, and flow rate.
- 780 4-20 mA Input Module for interfacing to non-Teledyne ISCO devices with 4 to 20 milliampere output signals.

The modules are optional accessories. The samplers do not require them for operation. However, the modules offer a number of advantages: they are an economical way to combine flow-rate or parameter monitoring with sampling, and you can program the samplers and modules as a single unit. Additionally, the samplers store the readings in memory.

## **1.3 SDI-12 Sondes**

The data parameters used by the sonde and sampler must match. Some sondes have "plug and play" setup. Known as "ISCO Ready," these sondes can tell the sampler what values

they have, their order, and units of measure. Other sondes can still be used, but require manual setup to identify proper data types for each data value reported.

## 1.4 Telephone Communication

This sampler is capable of cellular communication when a modem is installed. Remote phone communication enables alarms, data retrieval, and sampler control for remote sites. Cellular modems allow the sampler to send and receive SMS text messages.

### 1.4.1 6712 LTE Modem

The 6712 LTE is a modem for wireless communication using the Long-Term Evolutions (LTE) standard. It requires a customer-supplied cell phone service SIM from Verizon or another provider. Refer to Section 1.4.2 for details.

### 1.4.2 Software Compatibility

The 6712 LTE modem works with 6712 Series/Avalanche/BLZZRD samplers running software version 3.05 or later. The RTD (Rapid Transfer Device) works with 6700 Series/BLZZRD samplers running software version 3.05 or later. The following instructions detail the steps for identifying your sampler's software version, downloading the required software, and installing it in the sampler controller.

#### *Identifying Your Software Version*

To determine if your sampler has the correct software installed for your application:

1. Turn the sampler on. From the home screen, select VIEW REPORT.
2. Select SYSTEM IDs.

The software version will appear in the list of IDs.

#### *Downloading Software*

To download the correct version of software:

1. Go to Teledyne ISCO's public website at <http://www.teledyneisco.com>
2. Click WHAT WE DO on the menu ribbon near the top of the page. On the panel that opens, select WATER AND WASTE-WATER > SOFTWARE AND FIRMWARE UPDATES.
3. Scroll down to find the 6712 / BLZZRD section.
4. Select and download the appropriate language version you require.
5. Rev. D 6712 – BLZZRD units require version 3.05.004 or above.
6. Save the .zip file to a directory of your choice and extract the contents.

#### *Installing Software*

 <b>CAUTION</b>
--

Updating the sampler's software will reset the program and memory partition settings to factory default settings and will erase stored data. Record settings and download data before proceeding.

1. Turn the sampler on and connect your computer to the interrogator port using either USB cable 60-2004-508 or serial port cable 60-2544-044.
2. Open the Update Software application, found in the program -> Flowlink 5.1 menu.
3. Follow the instructions provided in the PDF document Updating Software Procedures. The Help file in the application also contains these instructions.

**Note**

While there are other code versions, ONLY the Rev. 3.05.004 or above is compatible with the Rev. D 6712-BLZZRD model.

For Path to Files:, use the Select... button to browse to the files you downloaded. The software filename will have a .bin extension.




Figure 1-1 Software Update window

## 1.5 Memory to Store Monitoring Data

The samplers contain enough memory to store five sampling programs, sampling data, Teledyne ISCO 700 Series Module readings, and SDI-12 parameter readings. You can view the readings on your sampler's display.

You can also retrieve the readings and reports so the information can be processed on a personal computer. Readings and reports may be collected with a computer running Teledyne ISCO's Flowlink software. Flowlink can download the information through a direct connection, a modem connection (when the sampler is equipped with the optional dial-out modem or connected to the optional digital cellular modem), or from a 581

Rapid Transfer Device (RTD). The RTD is a quick and simple way to transfer the data from the field to your computer. Reports are easily collected with Teledyne ISCO's Samplink software.

## 1.6 Pump Requirements

The sampler's peristaltic pump is extremely powerful. This pump requires a high capacity power source such as deep cycle marine battery or AC power. More information on power requirements can be found in section 2.6, *Installing a Power Source*.

The pump also requires a pump tube made specifically for the BLZZRD and Teledyne ISCO's 6700 Series samplers. These pump tubes are easily recognized by their blue alignment collars. The pump tubing is the same as that for Teledyne ISCO's 6712 Samplers, but different from Teledyne ISCO's earlier model samplers, such as the 3700 Series. Other types of pump tubing will not work in the BLZZRD. Refer to section 9.4, *Replacing the Pump Tube*.

 **CAUTION**

This sampler has a high performance pump. As with all such pumps, it relies upon liquid to cool working components. If the sampler is programmed to pump in the absence of liquid in excess of 5 minutes, excessive heat buildup may damage the paddles, rollers, and housing. Ensure that the liquid inlet is completely immersed.

**Table 1-1 BLZZRD Sampler Features**

<b>General Features</b>	
Top Cover	<ul style="list-style-type: none"> <li>• Protects pump and controller.</li> </ul>
Controller	<ul style="list-style-type: none"> <li>• Control panel sloped 15 degrees for easy reading.</li> <li>• Keys labeled with large, vivid icons.</li> <li>• 80-character display (4 lines by 20 columns).</li> <li>• Display has selectable backlight:                             <ul style="list-style-type: none"> <li>Always on or always off.</li> <li>Timed, switching off when keypad is inactive for 60 seconds.</li> </ul> </li> <li>• Memory for program and data storage.</li> <li>• Flash memory for easy software upgrades.</li> <li>• NEMA 4X and 6 (IP67) ratings.</li> <li>• Sample compartment temperature logging.</li> <li>• Contains a rechargeable desiccant to prevent moisture damage to the electronics, pump, and distributor systems.</li> </ul>
Refrigerator Section	<ul style="list-style-type: none"> <li>• Insulated.</li> <li>• Accepts 1, 4, or 14 bottle configurations.</li> <li>• Air-circulation fan quickly cools samples.</li> <li>• Mounting points for optional locking cable.</li> </ul>

**Table 1-1 BLZZRD Sampler Features (Continued)**

Distributor Arm	<ul style="list-style-type: none"> <li>• Directs liquid to multiple bottles. Easily removed for composite sampling.</li> </ul>
Discharge Tube	<ul style="list-style-type: none"> <li>• Routes sample liquid from pump tube, through distributor arm to sample bottle.</li> </ul>
Composite Tube Guide for Composite Sampling	<ul style="list-style-type: none"> <li>• Keeps discharge tube in place over composite bottles.</li> </ul>
Optional Mobility Cart	<ul style="list-style-type: none"> <li>• Pneumatic tires for rugged terrain.</li> <li>• Comfortable handle for easy maneuvering.</li> </ul>
Compatible Teledyne ISCO Products	<ul style="list-style-type: none"> <li>• 581 Rapid Transfer Device</li> <li>• 674 Rain Gauge</li> <li>• 700 Series Modules</li> <li>• 2100 Series Flow Meters</li> <li>• Signature Series Flow Meters</li> <li>• Flowlink 4, Samplink</li> <li>• 1640 Liquid Level Actuator</li> <li>• SDI-12 Sondes</li> <li>• 4100 Series Flow Loggers</li> <li>• 4200 Series Flow Meters</li> </ul>
Real-Time Displays	As the sampler runs a sampling program, it displays the program's status. The status display may include such information as the time of the next sample, the number of the next bottle, whether the sampler is disabled or stopped, and the refrigerated compartment temperature. If the sampler encounters an error while running the program, it displays a message alerting you to the problem.
Programming for 700 Modules	Program the modules from the sampler's control panel. The module's program settings become part of the sampling program.
Setup for SDI-12 Sondes	Program SDI-12 sondes from the sampler's control panel. The program settings become part of the sampling program. See Section 7.
Memory for Stored Programs and Readings	512 kilobytes of battery-backed RAM (Random Access Memory), to store: <ul style="list-style-type: none"> <li>• Five sampling programs.</li> <li>• A sampling report from the most recently run program. It records as many as 1000 sampling events; events can be the program start time, enable time, sample event information, etc.</li> <li>• 700 Series module readings. The readings can be: level, flow rate, velocity, pH, temperature, or the data collected by the 4-20 mA module.</li> <li>• Rain gauge and refrigerator temperature readings.</li> <li>• SDI-12 sonde readings.</li> </ul>
Five Reports Available	<ul style="list-style-type: none"> <li>• The Program Settings report, listing current program settings.</li> <li>• The Sampling Results report, listing the events occurring during the program.</li> <li>• The Combined Results report, combining sampling events with readings from a rain gauge, module, or SDI-12 sonde.</li> <li>• The Module Summary report, summarizing flow rate or parameter readings.</li> <li>• The Rainfall Summary report, listing a summary of rainfall readings.</li> </ul>
Serial Data Output	ASCII data output from the interrogator port.
Units of Measure	A variety of metric and English units of measure for length, flow rate, flow volume, and temperature.
External Sampler Enable	Teledyne ISCO flow meters and flow loggers have a programmable sampler enable feature that lets them send an electronic signal to a BLZZRD that enables (starts) or disables (stops) a running sampling program.
Dual Sampler Mode	Dual Sampler Mode operates two samplers, running programs sequentially.
Command Driven Operation	Operate sampler functions using RS-232 communications.

**Table 1-1 BLZZRD Sampler Features (Continued)**

Warning Messages	Pump Tube Warning. The BLZZRD displays a warning to inspect the tube.
On-Line Help Notes	When programming the sampler, press the [?] (Help) key for a brief help note. All help topics appear in the index.
Two Programming Levels	Standard programming lets you set up typical sampling programs quickly. Extended programming includes all features available in standard programming plus additional features.
<b>Sample Delivery System Features</b>	
Peristaltic Pump	Benefits: Liquid moves continuously under pumped flow. The pump has no metering chambers or gravity fed internal tubing to trap sediment or residual liquid. Sample liquid contacts only the strainer, suction line, tube coupling, pump tube, bulkhead fitting, and sample bottles.
LD90 Liquid Detector	Non-wetted Detection: Sample liquid never touches the detector.
Two Line Purges in Sampling Cycle	The sampling cycle always includes a pre-sample purge and post-sample purge that clears the suction line of residual liquid.
Easy Grab Samples	Simply disconnect the pump tube from the bulkhead fitting on the refrigerated section and place the pump tube over your sample container.
Vinyl and Teflon® Suction Lines	<ul style="list-style-type: none"> <li>• 3/8-inch ID (Inside Diameter) vinyl line.</li> <li>• 3/8-inch ID Teflon® lined with polyethylene jacket.</li> </ul>
<b>Standard Programming Features</b>	
Pacing	<ul style="list-style-type: none"> <li>• <b>Uniform Time Pacing:</b> Sampling at regular time intervals.</li> <li>• <b>Flow Pacing:</b> Sampling at regular flow-volume intervals.</li> </ul>
Distribution Methods	<ul style="list-style-type: none"> <li>• <b>Composite:</b> Samples deposited in a single large bottle.</li> <li>• <b>Sequential:</b> Only one sample placed in each bottle.</li> <li>• <b>Samples Per Bottle:</b> Multiple samples placed in each bottle.</li> <li>• <b>Bottles Per Sample:</b> One sample deposited in multiple bottles.</li> </ul>
Three Flexible Start-Time Settings	<ul style="list-style-type: none"> <li>• <b>Start Immediately:</b> Starts the sampling program immediately.</li> <li>• <b>Delayed Start:</b> Starts the sampling program after a user-definable delay of 1 to 999 minutes.</li> <li>• <b>Clock Time:</b> Starts the sampling program at a user-definable time on one or more days of the week.</li> </ul>
Option for a Continuous Running Program	<ul style="list-style-type: none"> <li>• <b>Continuous Sampling:</b> When sample bottles are regularly replaced, the sample distribution can restart with the first bottle set after the last bottle set is filled, without interrupting the running program.</li> </ul>
<b>Extended Programming Features</b>	
Pacing	<ul style="list-style-type: none"> <li>• <b>Uniform Time Pacing:</b> Sampling at regular time intervals.</li> <li>• <b>Nonuniform Time Pacing:</b> Sampling at irregular time intervals.</li> <li>• <b>Random Interval Pacing:</b> Sampling at unique random time intervals generated by the controller.</li> <li>• <b>Flow Pacing:</b> Sampling at regular flow-volume intervals.</li> <li>• <b>Event Pacing:</b> Sampling each time a user-definable event occurs.</li> </ul>
Flow Proportional Sample Volumes	Allows for sample sizes to be based on flow. (This option is available only with uniform time pacing.)

**Table 1-1 BLZZRD Sampler Features (Continued)**

Distribution Methods	<ul style="list-style-type: none"> <li>• <b>Composite:</b> Samples deposited in a single large bottle.</li> <li>• <b>Sequential:</b> Only one sample placed in each bottle.</li> <li>• <b>Samples Per Bottle:</b> Multiple samples placed in each bottle</li> <li>• <b>Bottles Per Sample:</b> One sample deposited in multiple bottles</li> <li>• <b>Multiple Bottle Compositing:</b> A combination of samples per bottle and bottles per sample distribution methods.</li> <li>• <b>Time Switched Bottles or Bottle sets:</b> Control the sample distribution using clock times.</li> </ul>
Programmable Sampler Enable	An BLZZRD sampler can be programmed to enable or disable a running sampling program when readings received from a connected rain gauge, module, or SDI-12 Sonde meet certain conditions.
Pauses and Resumes	Create intermittent sampling schedules.
Two-Part Programming	Two-part programming lets you set up a sampling program that divides the bottles into two groups, filling each group according to separate pacing, distribution, sampler enable, and pause and resume settings. This is ideal for storm-water run-off sampling.
Auto Suction Head or Fixed Suction Head	<p>The suction head, is the vertical distance from the flow stream to the liquid detector. Extended programming has two settings.</p> <ul style="list-style-type: none"> <li>• <b>Auto-Suction Head:</b> The head is automatically determined.</li> <li>• <b>Fixed Suction Head:</b> A user-definable measurement for the head.</li> </ul>
Suction Line Rinses	Program setting for the number of times (0 to 3) that the BLZZRD rinses the suction line before drawing a sample.
Sampling Retries	Program setting for the number of times (0 to 3) that the BLZZRD attempts to sample if it fails to deliver the entire sample volume.
Three Flexible Start Time Settings	<ul style="list-style-type: none"> <li>• <b>Run Immediately:</b> Starts the sampling program immediately.</li> <li>• <b>Delayed Start:</b> Starts the sampling program after a user-definable delay of 1 to 999 minutes.</li> <li>• <b>Clock Time:</b> Starts the sampling program at a user-definable time on one or more days of the week.</li> </ul>
Option for a Continuous Running Program	<b>Continuous Sampling:</b> When sample bottles are regularly replaced, the sample distribution can restart with the first bottle set after the last bottle set is filled, without interrupting the running program.

**Table 1-2 BLZZRD Construction Materials**

Item	Material
Top Cover	ABS (Acrylonitrile Butadiene Styrene) plastic
Bottle Rack	Powder coated steel
Controller Case	Noryl®
Distributor Arm	Polypropylene and stainless steel
Distributor Arm Nut Pump Rollers	Delrin®
Control Panel Connector Labels	Polyester



**Table 1-2 BLZZRD Construction Materials (Continued)**

Distributor Shaft Pump Shaft Pump Paddles Pump Band Bulkhead Fitting Latches Handles Metal Retaining Rings	Stainless steel
Draw Catches:	EPDM (Ethylene Propylene Diene Monomer) rubber
Amphenol Connectors	Cadmium-plated Aluminum

**Table 1-3 Technical Specifications for the BLZZRD Sampler**

General Notes:	
1. All weights may vary $\pm 1$ lb ( $\pm 1/2$ kg).	
2. All dimensions may vary $\pm 1/4$ inch ( $\pm 0.64$ cm).	
3. Sample delivery specifications valid for the following conditions and ranges, unless otherwise stated: 75 °F (24 °C) ambient, sample liquid is tap water at 50 to 80 °F (10 to 27 °C), sample volumes from 50 ml to 1000 ml, suction line lengths of 25 ft (7.6 m) or less, suction heads from 1 to 20 ft (0.3 to 6.1 m), atmospheric pressure between 29.92 and 31.89 inHg (760 and 810 mmHg), and a power source of a Teledyne ISCO High Capacity Power Pack at 12.5 volts DC no load output.	
<b>Controller, Pump, and Tubing: Mechanical Specifications</b>	
Weight of Controller:	Controller only: 13.0 lbs (5.9 kg) With pump tube: 13.2 lbs (6.0 kg)
Controller Dimensions:	Length: 10.3 in (26.0 cm) Width: 12.5 in (31.7 cm) Height: 10 in (25.4 cm)
Temperature Ranges:	Operational: 32 °F to 120 °F (0 °C to 49 °C) Storage: -20 °F to 140 °F (-29 °C to 60 °C)
Maximum Altitude:	2,000 Meters
Installation Category:	II
Pollution Degree:	2
Humidity:	95% RH Maximum
Enclosure:	NEMA 4X and 6 Enclosure: IP67 Pump: IP17
Typical Delivered Volume Accuracy:	(The ability to deliver the programmed sample volume.) $\pm 10$ ml or $\pm 10\%$ of programmed value, whichever is greater.
Typical Repeatability:	(The ability to repeat the delivered volume for a set of samples collected under the same conditions.) $\pm 5$ ml or $\pm 5\%$ of the average of the maximum and minimum sample volume in a sample set, whichever is greater, at lifts up to 25 feet.

**Table 1-3 Technical Specifications for the BLZZRD Sampler (Continued)**

Typical Pump Flow Rate and Line Transport Velocity:	Suction Head	Flow Rate	Line Transport Velocity
	3 feet	3.9 liters per minute	3.0 ft/s (0.91 m/s)
	5 feet	3.9 liters per minute	3.0 ft/s (0.90 m/s)
	10 feet	3.7 liters per minute	2.9 ft/s (0.87 m/s)
	15 feet	3.6 liters per minute	2.7 ft/s (0.83 m/s)
	20 feet	3.3 liters per minute	2.5 ft/s (0.77 m/s)
	23 feet	3.1 liters per minute	2.3 ft/s (0.71 m/s)
	25 feet	2.8 liters per minute	2.2 ft/s (0.66 m/s)
Maximum Suction Head:	28 ft (8.5 m) at 30 inHg (762 mmHg). (The maximum suction head is the vertical height at which a sample can be taken.)		
Pump Tube Life:	Tube life may vary depending on the abrasiveness of the sample liquid. Recommended maximum 1,000,000 pump counts. This equates to 912 standard samples. A standard sample is 200 ml at 5 ft (1.5 m) suction head using a 10 ft (3 m) vinyl suction line.		

**Table 1-3 Technical Specifications for the BLZZRD Sampler (Continued)**

<b>BLZZRD Controller: Electrical Specifications</b>	
Controller Internal Battery:	5 years minimum (maintains internal logic, program settings & stored data)
Real Time Clock Accuracy:	1 minute per month
Program Memory:	Nonvolatile programmable Flash. Can be field updated via interrogator connector.
Sampler Power Requirements:	Nominal: 12 volts DC by battery or AC power converter Operational: 11 to 13 Volts DC Maximum: 14.4 Volts DC
<b>Power Consumption<sup>a</sup></b>	
<b>BLZZRD Controller in Standby</b>	10 mA
<b>Sampler Controller Operating Current</b>	Approximately 30 mA based on 200 mL sample every hour, 10 ft (3 m) suction line, and 5 ft (1.5 m) suction head.
<b>Sampler Refrigeration Operating Current</b>	Approximately 1 Amp maintenance current at 23°C ambient, with full sample bottles. <sup>d</sup>
<b>Average Current of Accessories</b>	
Controller Display Backlight	228 mA
701 pH Module <sup>b</sup>	11 mA
710 Ultrasonic Module <sup>b</sup>	11 mA
720 Submerged Probe Module <sup>b</sup>	11 mA
730 Bubbler Module <sup>b</sup> — at 2 ft of liquid level — at 10 ft of liquid level	24 mA 31 mA
750 Area Velocity Module <sup>b</sup>	12 mA
780 4-20 mA Module <sup>b</sup>	10 mA
YSI 600 with pH, DO, Conductivity, and Temperature <sup>b</sup>	11 mA
Programmable Analog 4-20 mA Output Option <sup>c</sup>	124 mA
6712 LTE modem — standby current — current while communicating	19.1 mA 837.6 mA
<p>a. Current ratings when input power is 12.5 VDC and ambient temperature is 77 °F (25 °C)</p> <p>b. Average current when programmed to take readings at 15 minute intervals.</p> <p>c. Rating for three outputs.</p> <p>d. For more information and calculation of power requirements, see Appendix F - Battery Selection Guide.</p>	


**Table 1-3 Technical Specifications for the BLZZRD Sampler (Continued)**

<b>BLZZRD Controller: Software Specifications</b>				
Suction Line:	Programmable 3 to 99 ft (1 to 30 m) lengths of: <ul style="list-style-type: none"> <li>• 3/8" (0.95 cm) I.D. Vinyl</li> <li>• 3/8" (0.95 cm) I.D. Teflon®-lined with polyethylene jacket</li> </ul>			
Sample Frequency:	<ul style="list-style-type: none"> <li>• From 1 minute to 99 hours and 59 minutes in 1 minute increments between consecutive samples.</li> <li>• Nonuniform times in minute intervals or clock time</li> <li>• Random time intervals between consecutive samples</li> <li>• From 1 to 9,999 flow pulses in single-pulse intervals</li> <li>• Flow paced in volume with attachable flow module</li> </ul>			
Rainfall Reading Units:	Inches or millimeters			
Module Readings:	<ul style="list-style-type: none"> <li>• <b>pH</b></li> <li>• <b>Ultrasonic<sup>a</sup></b></li> <li>• <b>Submerged Probe</b></li> <li>• <b>Bubbler</b></li> <li>• <b>Area Velocity</b></li> <li>• <b>4–20 mA</b></li> </ul>	1 byte representing pH 2 bytes representing temp. 2 bytes representing level 2 bytes representing level 2 bytes representing level 2 bytes representing level 2 bytes representing velocity 2 bytes representing percent 2 bytes representing level 4 bytes representing flow (1 byte = 8 bits)	0.1 pH storage resolution 0.1 °C storage resolution 0.0001 m storage resolution 0.0001 m storage resolution 0.0001 m storage resolution 0.0001 m storage resolution 0.001 ft/s storage resolution 0.1% storage resolution	
SDI-12 Sonde Readings:	<b>Parameter</b>	<b>Range</b>	<b>Resolution</b>	<b>Storage Bytes</b>
	Temperature	-40–100 °C	0.1 °C	2
	Conductivity	0–100 mS/cm	0.01 mS/cm	2
	Specific Conductance	0–100 mS/cm	0.01 mS/cm	2
	Total Dissolved Solids	0–90 g/l	0.01 g/l	2
	Salinity	0–70 ppt	0.1 ppt	2
	Dissolved Oxygen	0–20 mg/l	0.1 mg/l	1
	pH	0–14	0.1	1
	ORP	-999–999 mv	0.1 mv	2
	Level	0–6.5279 m	0.0001 m	2
	Ammonium-Nitrogen	0–200 mgN/l	0.1 mgN/l	2
	Ammonia-Nitrogen	0–200 mgN/l	0.1 mgN/l	2
	Nitrate-Nitrogen	0–200mgN/l	0.1 mgN/l	2
	Turbidity	0–5000 NTU	0.1 NTU	2
	Chloride	0v1000 mg/l	0.1 mg/l	2
	Chlorophyll	0–400 ug/l	0.1 ug/l	2
	Other	-99999.0–99999.0	0.0001	4

**Table 1-3 Technical Specifications for the BLZZRD Sampler (Continued)**

Module Reading Conversions:	<ul style="list-style-type: none"> <li>• Level and 4–20 mA readings converted to flow rate units: <ul style="list-style-type: none"> <li>gallons per second      gallons per minute</li> <li>million gallons per day      cubic feet per second</li> <li>cubic meters per second      cubic meters per hour</li> <li>cubic meters per day      liters per second</li> </ul> </li> <li>• Velocity units:      feet per second      meters per second</li> <li>• Totalized flow units: <ul style="list-style-type: none"> <li>gallons      million gallons</li> <li>cubic feet      cubic meters</li> <li>liters</li> </ul> </li> <li>• Devices supported in flow conversion: <ul style="list-style-type: none"> <li>Weirs:      V-notch—22.5, 30, 45, 60, 90, and 120 degrees</li> <li>                    Rectangular—with and without end contractions</li> <li>                    Cipoletti.</li> <li>Flumes:      Parshall— 1", 2", 3", 6", 9", 1.5', 2', 3', 4', 5', 6', 8', 10', 12'</li> <li>                    Palmer-Bowlus— 4", 6", 8", 12", 15", 18", 21", 24", 27",</li> <li>                    30", 48"</li> <li>                    Trapezoidal— LG 60 V, 2" 45 WSC, 12" 45 SRCRC</li> <li>                    H— 0.5', 0.75', 1', 1.5', 2', 2.5', 3', 4.5'</li> <li>Manning Formula:      Round, U-channel, Rectangular, Trapezoidal</li> <li>Area Velocity:      Round, U-channel, Rectangular, Trapezoidal</li> <li>Data Points:      50 level-flow rate points, 50 level-area points</li> <li>Flow Equation:      <math>Q = a \cdot H^b + c \cdot H^d</math> where: <ul style="list-style-type: none"> <li>Q = flow</li> <li>H = head</li> <li>a, b, c, d, = entered values</li> </ul> </li> </ul> </li> </ul>
Refrigerated Compartment Temperature Sensor	Range: -40.0 °C to 100 °C Accuracy specification: ±0.3 °C over a range of 0.0 to 10.0 °C. Reading data is 2 bytes, with a storage resolution of 0.1 °C.
<b>Refrigerant Information</b>	
Type	R134a
Global Warming Potential	1430
Charge	0.055kg
CO <sub>2</sub> Equivalent	0.0787t

**Table 1-3 Technical Specifications for the BLZZRD Sampler (Continued)**









<b>BLZZRD Sampler: Physical Specifications</b>	
Weight:	74.0 lbs (34 kg) Dry weight, less battery and bottle configurations
Dimensions:	Length: 28 in (71 cm) Width: 18.5 in (47 cm) Height: 31.5 in (80 cm)
Cooling:	CFC-free refrigeration system maintains sample temperature at 37.4 °F ±1.8 (3 °C ±1) at ambient temperatures ranging from 32 to 104 °F (0 to 40 °C).
Power Requirements:	DC mode: 12 VDC, 6 amperes (from external battery) AC mode: 90–264 Volt, 47–63 Hz, 2 amperes
 The mains line cord is the disconnect device.	
Operational Temperature:	32 to 122 °F (0 to 50 °C)
Bottle Configurations:	<ul style="list-style-type: none"> <li>• 14 wedge polypropylene bottles, 950 mL(1 quart)</li> <li>• 4 polypropylene bottles, 5.0 liter (1.3 gallon)</li> <li>• 4 glass bottles, 1.8 liter (.5 gallon)</li> <li>• 1 round polyethylene bottle, 9.4 liters (2.5 gallon)</li> <li>• 1 round glass bottle, 9.4 liters (2.5 gallon)</li> <li>• 1 square polypropylene bottle 19 liters (5 gallon)</li> </ul>

a. The 710 Ultrasonic Module and sensor have been discontinued as of August 17, 2012.




**Note**

The refrigerator contains hermetically sealed fluorinated greenhouse gases.  
 See Table 1-3, Refrigerant Information

**Table 1-4 Connector Specifications**

Connector Icon	Pin-Location Diagram	Pin Function
<p><b>Rain Gauge</b></p> 	<p><b>9-Pin Female, Sealed</b></p> 	<p>A +12 volts DC            B Ground            C 5V Programmable I/O1 pin            D Rain gauge input            E 12V switched            F SDI-12 Data recv            G SDI-12 Transmit            H 5V Programmable I/O2 pin            I 5V Programmable I/O3 pin</p>
<p><b>Interrogator</b></p> 	<p><b>6-Pin Female, Sealed</b></p> 	<p>A Switched +12 volts DC            B Ground            C Connection Sense            D Transmit            E Receive            F NC</p>
<p><b>Flow Meter</b></p> 	<p><b>6-Pin Male, Sealed</b></p> 	<p>A +12 volts DC            B Ground            C Flow Pulses In            D Bottle Number Out            Dual Sampler Out            E Event Mark Out            F Inhibit In</p>
<p><b>Module (in module bay)</b></p>	<p><b>8-Pin Male, Sealed</b></p> 	<p>1 NC            2 +12 volts DC            3 Ground            4 Module Sense            5 Receive            6 Transmit            7 Clock            8 Control</p>
<p><b>12 V PRIMARY POWER</b></p>	<p><b>3-Pin Male, Sealed</b></p> 	<p>A Ground            B +12 volts DC            C unassigned</p>

**Table 1-4 Connector Specifications (Continued)**

<p><b>12 V BACKUP POWER</b></p>	<p>2-PIN MALE, SEALED</p> 	<p>A Ground                  B +12 volts DC</p>
 <p>REFRIGERATION</p>	<p>5-PIN MALE, SEALED</p> 	<p>A +12 volts DC                  B unassigned                  C Enable signal                  D Ground                  E unassigned</p>



# BLZZRD™

## Portable Refrigerated Sampler

---

### *Section 2 Installation/Preparation*

#### **2.1 Receiving Inspection**

Carefully unpack the shipment and inspect the contents. If there is any damage to the shipping carton or any components, contact the shipping agent and Teledyne ISCO (or its authorized representative) immediately.

If there is any evidence that the system has been damaged in shipping, do not plug it into the power line. Contact Teledyne ISCO or its authorized representative for advice.

Compare the contents of the boxes with the enclosed packing slip. If there are any shortages, contact Teledyne ISCO immediately.

##### **2.1.1 Case Ventilation**

If your sampler was shipped via air transportation, the controller case ventilation through the Flowmeter connector will be open. Before placing the sampler into service, tighten the protective cap on this Flowmeter connector, or attach a mating cable to it. This connector must be sealed to prevent moisture from entering the case.

The controller case ventilation prevents differences between internal case and atmospheric pressures. Extreme differences in pressures will affect the keypad label. If bubbles form under the keypad label, momentarily open this vent by removing the cap or cable and allow the case pressure to equalize. If needed, push down on the bubbles to force the air out.

#### **2.2 Installation**

The following sections describe the steps necessary to place the sampler into operation. In most applications, the steps are:

1. *Positioning a Sampler.* (Section 2.3)
2. *Inspecting the Pump Tube.* (2.4)
3. *Installing the Bottle Kit.* (2.5)
4. *Installing a Power Source.* (2.6)
5. *Attaching the Suction Line.* (2.7)
6. *Attaching a Strainer.* (2.8)
7. *Connecting Instruments.* (2.9)
8. *Programming the BLZZRD.* (2.10)
9. *Locking the Sampler.* (2.11)

An overview of post sampling activities (2.12) follow the installation instructions.

### 2.3 Positioning a Sampler

There are a few considerations when selecting a site for the sampler. The foremost concern should be personal safety.

 **WARNING**

The installation and use of this product may subject you to hazardous working conditions that can cause you serious or fatal injuries. Take any necessary precautions before entering the worksite. Install and operate this product in accordance with all applicable safety and health regulations, and local ordinances.

 **WARNING**

If this product is used in a manner not specified in this manual, the protection provided by the equipment may be impaired.

 **WARNING**

The BLZZRD has not been approved for use in hazardous locations as defined by the National Electrical Code.

 **WARNING**

If you must enter a manhole or other dangerous location to install the sampler or suction line, observe standard safety precautions. Refer to Appendix C for a discussion of safety considerations.

The following points should also be considered:

- *Power*—The only means to totally remove power from the sampler is by disconnecting the mains line cord from the power outlet, or the cable to the battery. Position the battery or connect to the mains outlet in a location where power may be disconnected easily in an emergency.

  **WARNING**

**AC Powered samplers— Never defeat or modify the mains plug earth ground connection.**

- *Level surface*—The sampler should be placed on a level surface to prevent tipping or spills.

 **CAUTION**

Do not tip the sampler on its side or completely upside-down. Positions other than the sampler's normal upright position may cause oil to run into the compressor inlet, which can permanently damage the cooling system. If the sampler is tipped or turned over, the sampler's refrigeration system must not be operated for at least one hour after returning the sampler to its upright position.

- *Support*—The surface must be able to support the BLZZRD at full capacity. This weight would include the sampler, the full sample bottles, and the battery (if used).
- *Ventilation*—The sampler requires at least 1 foot (0.3 m) of air space around the refrigeration components. Inadequate ventilation will reduce the cooling capacity and significantly increase power consumption.
- *Environmental*—The sampler is designed for outdoor use. However, you should avoid installing the sampler in locations where its components are subject to chemical attack. Also, prolonged exposure to direct sunlight will eventually damage the exterior. If the sampler is subject to chemical attack or prolonged UV exposure, consider using a protective enclosure. Keep in mind that positioning the sampler in direct sunlight will also increase power consumption, a factor worth considering when using battery power.

 **CAUTION**

Do not install the BLZZRD in corrosive atmospheres. Exposure to hydrogen sulfide gas, marine atmospheres, or salt spray will damage the refrigeration system.

- *Avoid submersion*—Although the controller will resist damage (rated NEMA 4x, 6), the refrigerator system and bottle compartment cannot prevent the liquid from entering. Liquid entering the refrigerated system will damage the cooling system; liquid entering the bottle compartment will contaminate the collected samples.
- *Accessibility*—The sampler must be installed in a location where it can be recovered easily without tipping or difficult maneuvering.
- *Security*—The location may need to provide some degree of security to prevent tampering or vandalism. You can read more about securing the sampler in section 2.11.

Place the sampler on a flat, horizontal surface. Placing the sampler on an incline may cause the sample to miss the bottle mouth. When installing the sampler, be sure the vertical distance between the level of the liquid and the pump is as small as possible.

 **Note**

When using the BLZZRD with a 750 or 710 module, transmitters such as cell phones or walkie talkies must not be operated within 3 meters of the sampler. If water level readings are fluctuating due to TV/radio station transmitter towers in the area, the sampler must be relocated.

## 2.4 Inspecting the Pump Tube

The sampler is shipped from the factory with a new pump tube installed. An inspection for first-time use is not necessary. However, the pump tube should be inspected or replaced before running all subsequent sample-collection programs.

If your sampling protocol mandates that you replace the pump tube for each sampling program, refer to the replacement instructions in Section 9.4. Otherwise, the pump tube can remain until one of the two following conditions are present:

- The sampler controller displays a pump tube warning, or—
- A pre-sampling program visual inspection identifies a worn or damaged tube.

The pump tube must be replaced when the first of either condition exists.

### **Note**

The importance of regular tubing inspection and replacement cannot be overstated. The key is to replace the tube before failure, not after. When a pump tube ruptures, grit and other abrasive debris can be driven into the pump shaft seal. Over time, this abrasive material will degrade the pump seal, jeopardizing the NEMA 4x 6 rating of the controller.

Failure to maintain the pump tube may result in permanent damage to the sampler. Check the condition of the pump tube regularly and if the tube shows signs of fatigue or wear, replace it immediately. A properly maintained sampler will provide years of reliable service that is expected of a Teledyne ISCO Sampler.

Section 9.4 of this manual describes the pump tube removal and replacement steps. Afterwards, be sure to reset the pump counter (section 9.3.2).

### 2.4.1 Pump Tube Warning

The sampler displays a pump tube warning at the recommended replacement interval. The warning display will alternate with the run screens, and is part of the VIEW LOG screens. Regardless of the visual condition of the pump tube, it should be replaced as soon as possible after the warning.

The warning appears after the controller reaches the factory set value of 1,000,000 pump counts. This value will deliver approximately 500 samples of 200 ml each, using a  $\frac{3}{8}$ -inch by 10-foot suction line at a 5-foot suction head. The pump tube replacement interval of 1,000,000 pump counts should be sufficient for most applications. If you are sampling abrasive liquids or liquids with a high content of suspended solids, you may find that the pump tube requires replacement more frequently.

### 2.4.2 Visual Inspection

The pump tube must be inspected before running each program. Pump tube failures can prevent the sampler from collecting the samples—or worse—may even damage the controller.

To inspect the pump tube:

 **CAUTION**

Moving parts can cause injuries. Remove power before inspecting pump tube.

1. Disconnect power from the sampler.
2. Unlatch the pump band and swing it away to expose the pump tube. (The band is the rounded metal cover that holds the tube in the pump.)
3. Visually inspect the pump tube for cracks where it is compressed by the rollers. If the tube is cracked, it must be replaced. Section 9.4 provides detailed instructions on replacing the pump tube and resetting the pump tube counter.
4. Visually inspect the inside of the pump housing. The housing, drain holes, and rollers should be free from debris. Clean if necessary.
5. Latch the pump band.

## 2.5 Installing the Bottle Kit


The sampler is shipped from the factory with sample bottles in place. When using the sampler for the first time, you will only need to remove the bottle lids. When installing cleaned bottles or a new kit, keep these guidelines in mind:

- Install all bottles to ensure that they remain in place.
- Align the bottles correctly. If the bottles are misaligned, the sampler may miss the bottle mouth, or the sampler will deposit samples in the “wrong” bottle.
- Each time you change bottle configurations, check the program settings for the number of bottles and bottle volume.

### 2.5.1 Installing the 14-Bottle Kit


To install the 14-Bottle Kit (68-2970-001):

1. Attach the distributor arm to the distributor shaft underneath the sampler controller. Note that the metal shaft and arm are keyed to ensure correct alignment. Secure the distributor arm with the knurled nut.
2. Refer to Figure 2-2. Attach the 13 inch (33 cm) discharge tube (Teledyne ISCO Part Number 60-2973-054) to the bulkhead fitting. Route the tube through the stainless-steel tube holder.
3. Route the other end of the tube through the spring on the distributor arm. Then, place the end of the discharge tube through the opening in the end of the distributor arm. The end of the tube should protrude below the arm  $\frac{1}{16}$  to  $\frac{1}{8}$  inch (see Figure 2-2). If too much tube protrudes from the arm, it can catch on bottles and jam the distributor arm.



- 1 Bottle, 950 ml (×14)
- 2 Bottle Carrier
- 3 Bottle Lid (×14)
- 4 Adapter Plate
- 5 Discharge tube (×2)


*Figure 2-1 14-Bottle kit*



- 1 Bulkhead Fitting
- 2 Discharge Tube
- 3 Tube Holder
- 4 Distributor Arm
- 5 Nut


*Figure 2-2 Attaching the distributor arm and discharge tube*

- 4. Install the 14-bottle adapter plate in the refrigerated compartment. The bottom of the compartment has two posts that ensure the adapter plate is oriented correctly. See Figure 2-3.
- 5. Load the bottles into the carrier. Note that the bottles fit between the handles of the carrier. The handles keep the bottles positioned correctly in the carrier.
- 6. Place the carrier into the refrigerated compartment. The carrier must be aligned with the alignment post on the plate. See Figure 2-4. Once it is aligned, it should seat fully against the adapter plate.
- 7. Close and latch the refrigerated compartment.



1 Alignment Posts

Figure 2-3 14-Bottle adapter plate installed



1 Bottle Carrier  
2 Adapter Plate  
3 Alignment Post

Figure 2-4 Carrier alignment




Figure 2-5 14-Bottle configuration completed

**2.5.2 Installing the 4-Bottle Kit**

To install the 4-Bottle Kit (68-2970-002):

1. Attach the distributor arm to the distributor shaft underneath the sampler controller. Note that the metal shaft and arm are keyed to ensure correct alignment. Secure the distributor arm with the knurled nut. See Figure 2-2.
2. Refer to Figure 2-2. Attach the 13 inch (33 cm) discharge tube (Teledyne ISCO Part Number 60-2973-054) to the bulkhead fitting. Route the tube over the stainless-steel tube holder.
3. Route the other end of the tube through the spring on the distributor arm. Then, place the end of the discharge tube through the opening in the end of the distributor arm. The end of the tube should protrude below the arm  $\frac{1}{16}$  to  $\frac{1}{8}$  inch (see Figure 2-2). If too much tube protrudes from the arm, it can catch on bottles and jam the distributor arm.
4. Install the 4-bottle adapter plate in the refrigerated compartment. The bottom of the compartment has two posts that ensure the plate is oriented correctly. See Figure 2-7.
5. Place the bottles into the refrigerated compartment. The bottles should seat fully against the adapter plate.
6. Close and latch the refrigerated compartment.





- 1 Bottle, 5 Liter (×4)
- 2 Bottle Lid (×4)
- 3 Adapter Plate
- 4 Discharge Tube (×2)



- 1 Glass Bottle, 1.8 Liter (×4)
- 2 Bottle Lid (×4)
- 3 Adapter Plate
- 4 Discharge Tube (×2)
- 5 Bottle Stabilizer

*Figure 2-6 4-Bottle kits*





1 Alignment Posts

Left: Four 5 L Plastic Bottle Adapter Plate

Right: Four 1.8 L Glass Bottle Adapter Plate

*Figure 2-7 4-Bottle adapter plate installed*




Left: Four 5 L Plastic Bottle Configuration

Right: Four 1.8 L Glass Bottle Configuration

*Figure 2-8 4-Bottle configurations completed*

**2.5.3 Installing 2.5 Gallon Composite Bottle Kits**

There are two 2.5 gallon composite bottle kits for the sampler, glass (68-2970-006), and Nalgene (68-2970-009). Refer to Figures 2-9 and 2-10. Both 2.5 gallon composite bottle kits use the same adapter plate.




- 1 Glass Bottle
- 2 Composite Tube Guide
- 3 Adapter Plate
- 4 Discharge Tube (×2)
- 5 Bottle Lid (not shown)

*Figure 2-9 Composite 2.5 gallon glass kit*




- 1 Nalgene Bottle
- 2 Composite Tube Guide
- 3 Adapter Plate
- 4 Discharge Tube (×2)
- 5 Bottle Lid (not shown)

*Figure 2-10 Composite 2.5 gallon Nalgene (polyethylene) kit*



*Figure 2-11 Attaching the composite tube guide and discharge tube*

1. Attach the composite tube guide to the underside of the controller. The tube guide will snap in place.
2. Attach the 11<sup>1</sup>/<sub>4</sub> inch (28.5 cm) discharge tube (Teledyne ISCO Part Number 60-2963-033) to the bulkhead fitting. Route the tube over the stainless-steel tube holder (Figure 2-11).
3. Place the other end of the tube into the side opening of the composite tube guide. The end of the tube should be positioned just below the bottom opening of the composite tube guide.
4. Install the composite bottle adapter plate in the refrigerated compartment. The bottom of the compartment has two posts to ensure that the adapter plate is oriented correctly. See Figure 2-12.
5. Place the bottle into the refrigerated compartment. The bottle should seat fully against the adapter plate.
6. Close the refrigerated compartment. Figure 2-13 shows the completed configurations.
7. Latch the refrigerated compartment.



1 Alignment Posts

*Figure 2-12 Composite bottle adapter plate installed*



*Figure 2-13 Composite 2.5 gallon bottle kits completed*

#### 2.5.4 Installing the 5 Gallon Composite Bottle Kit

1. Attach the 11¼ inch (28.5 cm) discharge tube (Teledyne ISCO Part Number 60-2963-033) to the bulkhead fitting. Route the tube over the stainless-steel tube holder.
2. Install the adapter plate and 5 gallon composite bottle into the refrigerated compartment. The bottle opening should be oriented toward the left side of the compartment.

**Note**

The 5-gallon configuration does not use a composite tube guide.

3. While closing the refrigerated compartment, place the end of the discharge tube into the bottle opening.
4. Close and latch the refrigerated compartment.



- 1 Bottle, 5 gallon
- 2 Discharge Tube (×2)
- 3 Bottle Lid (not shown)

Figure 2-14 Composite 5 gallon bottle kit





Figure 2-15 Composite 5 gallon bottle kit completed

## **2.6 Installing a Power Source**

The BLZZRD allows you to power the refrigerated sampler using either AC (90–264 Volt, 47–63 Hz), DC (12 Volt), or both. If both power sources are connected, the sampler will use AC power unless it is interrupted, at which time it will switch to DC power.


### **2.6.1 Installing the Battery Box**

To fasten the battery to the tray, use the battery box strap that is supplied with the 948 battery box to retain the battery in place. Pre-drilled holes are in the battery tray to allow for the fastening of the battery box strap. Alternatively, use an elastic bungee and place the hooks into the battery tray holes.



- 1 Holes for battery box strap
- 2 Holes for bungee tie down

*Figure 2-16 Mounting holes for battery box*



*Figure 2-17 Battery box strap installed*



*Figure 2-18 Battery box secured*

### 2.6.2 AC Power Sources



**AC Powered samplers– Never defeat or modify the mains plug earth ground connection.**

To connect the sampler to AC power, use the attached AC power cord. The sampler may be ordered with a power cord for North American outlets, or with a power cord for most European outlets. Users in other regions may need to purchase an appropriate plug adapter for use with the available power outlets. Because the sampler includes an internal universal AC power converter, only outlet adapters are required to configure the sampler for the power source. No other hardware or wiring changes are necessary.

### 2.6.3 12 VDC Power Sources

To connect the sampler to a DC power source, use the supplied 12 VDC connect cables. Connect cable 60-1394-023 is used to connect the BLZZRD to an automotive or deep-cycle marine battery with heavy-duty clips. These cables provide over-current protection through the use of an in-line 8A Slo-Blo fuse.



Only use Teledyne ISCO battery cables to connect the sampler to a DC power source. The cable length and fusing protect you and the equipment from over-current conditions and the risk of fire.



Never use a DC extension cable without first consulting with a Teledyne ISCO Service Technician. They will advise you of the proper wire gauge for the length you require.



Be sure to attach the positive and negative clips to the correct battery terminals. If polarity is reversed, the sampler may be permanently damaged.

The refrigerator compressor is equipped with a battery monitor that will cut out to protect the compressor and the battery when the voltage drops to 10.4–10.8 VDC.



Never charge the battery while it is connected to the sampler. Over-voltages could damage internal electronic components.

### 2.6.4 Battery Recommendations

Before each sampling program, the battery should be exchanged with a fully-charged battery.

Power consumption is mostly determined by the ambient temperature. Higher ambient temperatures will in turn create a higher demand for power. A battery selection guide is provided in Appendix F to help you determine the required battery capacity.

## **2.7 Attaching the Suction Line**

The suction line is the tubing from the sampling point to the pump intake. The BLZZRD uses a  $\frac{3}{8}$ -inch ID suction line of lengths 3 to 99 feet. Teledyne ISCO offers vinyl or PTFE suction lines. The PTFE tubing has a polyethylene jacket to protect it from kinks and abrasions.

*Guidelines for Measuring and Cutting the Suction Line:*

- Cut the line to the desired length but use the shortest length feasible for the installation.
- Cut the line in 1 foot increments. For instance, 4 feet, not  $3\frac{1}{2}$ . If using metric units of measure, cut the line in increments of 0.1 meter. Do not include the length of the strainer in the measurement.

The vinyl line contains a very low parts-per-million level of phenols. If phenol content affects your sample analysis, use the PTFE suction line.

When installing the sampler, be sure the vertical distance between the liquid level and the pump is as small as possible.

### **2.7.1 Attaching Suction Line to Pump Tube**

Attach the vinyl suction line to the pump tube with the  $\frac{3}{8}$ -inch tubing coupler. First, screw the threaded end into the suction line until the flat surface is flush against the suction line (Figure 2-19). Then, push the other end of the coupler into the end of the pump tube until the other flat surface is flush against the tubing.






Figure 2-19 Attaching the suction line to the pump tubing

## 2.8 Attaching a Strainer

Teledyne ISCO offers three styles of strainer that help prevent solids from clogging the suction line:




- $\frac{3}{8}$ -inch standard weighted polypropylene strainer for routine and priority-pollutant sampling
- $\frac{3}{8}$ -inch stainless steel low flow strainer
- $\frac{3}{8}$ -inch weighted CPVC-coated strainer for acidic liquid sources

To select the right strainer for your application, see Table 2-1. To install the strainer, carefully screw the strainer's threaded connector into the suction line. If attaching the strainer to a PTFE suction line, heat the end of the PTFE suction line before threading the strainer into the line. Warming the line will make the line more pliable to avoid damage.

### 2.8.1 Alternative to Strainers

When sampling from high velocity streams with heavy suspended solids, some field investigations suggest that more representative samples are obtained without the strainer. Consider attaching a short piece of thin walled aluminum tubing to the end of the suction line; anchor the tubing so that the inlet opens upstream. The aluminum tubing's thin wall will not disturb the flow stream, and most sample analyses disregard aluminum ions. Under most conditions, the pre-sample purge removes any debris over the tubing entrance.

**Table 2-1 Selecting the Right Strainer**

Strainer	Dimensions	Application	Maximum Depth
<b>3/8 Stainless Steel Ends, Polypropylene Center</b>		Routine applications.	Vinyl Line: 22 feet (6.7m). PTFE Line: 15 feet (4.5m).
<b>3/8 Stainless Steel Low Flow</b>		Low flow applications	Vinyl Line: 22 feet (6.7m). PTFE Line: 15 feet (4.5m).
<b>3/8 CPVC</b>		Highly acidic liquids.	Vinyl Line: 4 feet (1.2m).

**Note**

- The suction lines will float when filled with air during the purge cycles and when depths exceed the listed ratings. Secure the suction lines when sampling at depths exceeding the maximum ratings.
- All strainer dimensions are in inches.
- PTFE suction line is compatible only with the 3/8 stainless steel strainer.

**2.8.2 Tips for Routing Suction Line and Strainer**

Route the line so that it runs continuously downhill from the sampler to the liquid source. This helps drain the line during pre-sample and post-sample purges. When the sampler is used in below freezing temperatures, there is a risk of the suction line being frozen. A suitably warm sampling source can usually prevent this, provided there are no loops in the suction line. Some situations may require more protective measures, such as insulation of the suction line, or heat tape. Thoroughly draining the suction line minimizes the possibility of frozen liquid clogging the line.

For representative samples, place the strainer in the main current of the flow stream, not in an eddy or at the edge of flow. Placing an intake at the bottom may produce samples with excess heavy solids and no floating materials, while placement at the top may produce the opposite conditions.

**2.8.3 Pressurized Lines**

The sampler can obtain samples from pressurized lines, as long as the line pressure remains below 15 PSI (pounds per square inch). Pressures greater than 15 PSI may prevent the sampler from purging the suction line; moreover, extreme pressures can force liquid past the pump, even when the pump is not running.

For more consistent sample volumes from lines under constant pressure, calibrate the samples.

 **Note**

Teledyne ISCO does not recommend sampling from pressurized lines.

## 2.9 Connecting Instruments

### 2.9.1 Connecting 700 Series Modules

The BLZZRD sampler can be used in conjunction with the 700 Series Modules and several types of external instruments.

The bay on the controller's side accepts any of Teledyne ISCO's 700 Series Modules. These modules are optional accessories that are not required for operation of the sampler. However, the modules offer an economical way to combine flow-rate or parameter monitoring with sampling.




The 720 Submerged Probe, 730 Bubbler, and 750 Area-Velocity Flow Modules have a side-mounted desiccant cartridge. Before installing one of these modules, remove the desiccant cartridge and the two retaining clips. This is necessary for the module to fit under the controller cover. After installing the module, insert the desiccant cartridge into the sampler's two retaining clips located behind the controller.

Refer to the Module's manual for detailed installation instructions. Generally, to install a module:

1. If desired, retrieve stored data. (See Section 6.7.)
2. Turn the sampler off.
3. Remove the connector cap in the module bay and move it aside.
4. Slide the module into the bay. Push against the module to be sure the connector is firmly seated.
5. Complete the module installation by attaching wire or sensor connectors, or tubing to the module as required.
6. Turn on the sampler controller. The controller updates its configuration to add the features of the module and adds a memory partition (section 3.9) for the parameter data.

**2.9.2 Connecting External Instruments**

The BLZZRD is compatible with other Teledyne ISCO parameter and flow measurement instruments, as well as SDI-12 sondes from various manufacturers.

Table 2-2 Connecting External Instruments to the Sampler	
CONNECTOR ICON	CONNECT THESE INSTRUMENTS:
<p><b>Rain Gauge</b></p> 	<ul style="list-style-type: none"> <li>• 674 Rain Gauge</li> <li>• Programmable I/O Pins (pins C, H, and I)</li> <li>• SDI-12 Sondes</li> </ul>
<p><b>Interrogator</b></p> 	<ul style="list-style-type: none"> <li>• 581 RTD (Rapid Transfer Device)</li> <li>• IBM PC or compatible computer running FLOWLINK or SAMPLINK</li> <li>• External Modem</li> <li>• External Data Logger/Controller</li> </ul>
<p><b>Flow Meter</b></p> 	<ul style="list-style-type: none"> <li>• 1640 Liquid Level Actuator</li> <li>• Signature Flow Meters</li> <li>• Dual Mode Sampler</li> <li>• Pulse Duration Input Interface</li> <li>• 4-20 mA Input Interface</li> <li>• Closed-pipe Flow Meters via 4-20 mA Interface</li> </ul>

**Rain Gauge Connector** – Attach the rain gauge, SDI-12 Sonde, or Refrigerator Temperature Sensor cable connector to the nine-pin Rain Gauge port. When connecting a rain gauge and an SDI-12 Sonde, or rain gauge and a Refrigerator Temperature Sensor, use the appropriate ‘Y’ cable. For part numbers and ordering information, see the Accessories appendix at the end of this manual.

To operate with an SDI-12 sonde, the sampler must be placed in Extended Programming mode, and SDI-12 Sonde Setup must be selected from the Hardware Setup menu (refer to Section 5.16).

For complete SDI-12 sonde setup instructions, refer to Section 7.

**Interrogator Connector** – This connector is the primary port for data collection and external sampler control.

**Flow Meter Connector** – Connect the sampler to a Teledyne ISCO 4200 Series Flow Meter, 4100 Series Flow Logger, 2100 Series Flow Module, 1640 Liquid Level Actuator, or input interface by attaching their connect cable to the Flow Meter Connector. When using the sampler with an external instrument, remember these guidelines:

- Both the sampler and the flow instrument must be on and running a program.
- If using a flow meter, logger, or module for flow pacing, the sampler and the flow instrument must both run programs with flow pacing settings.

- If using a flow meter or flow logger for trigger pacing, the sampler must run a program with flow pacing settings, and the flow instrument must run a program with trigger pacing settings.
- When the sampler runs an event-paced program, it disregards pacing pulses from a flow meter or flow logger. However, the sampler continues to monitor for enable signals.

 **WARNING**

**Most Teledyne ISCO flow meters are not certified for use in hazardous locations as defined by the National Electrical Code. Contact your Teledyne ISCO representative for your equipment's status. Never operate equipment in hazardous locations that has not been certified as safe for such use. Refer to Appendix C in the back of this manual for more safety information.**

## 2.10 Programming the BLZZRD

Due to the many features supported by the BLZZRD sampler, programming instructions are not addressed in this section. Refer to the following sections for programming instructions:

- Section 3 introduces the keypad and display and provides general programming instructions.
- Section 4 provides instructions for the sampler's *Standard* programming features.
- Section 5 provides instructions for the sampler's *Extended* programming features.

## 2.11 Locking the Sampler

The BLZZRD sampler contains two locking latches by which the sampler may be mechanically locked. Use padlocks or a locking cable mechanism to tamper-proof the sampler.

### 2.11.1 Program Lock

A program lock software option can be purchased from Teledyne ISCO. This option secures access to the sampler's program settings. The program lock is ordered separately as Teledyne ISCO Part Number 60-9003-532.

Once activated, the option is always available, even if the controller software is upgraded at a later time. The option must be installed at the factory. Consult the factory for further details.

To see if this software option has been activated, select VIEW REPORT>SYSTEM IDs. A software options screen will list a code for each active option. If active, the Program Lock will be listed as option code 1108.

If the program lock is enabled, a user-selected password (up to five digits) must be entered before any program changes or run time modifications can be made. Run time modifications include accessing the Manual Pause menu options and shutting off the sampler. You will still be able to run the current program, view reports, and interrogate the sampler without the knowledge of the password.

If three consecutive attempts to enter the password fail, a five minute “time-out” will follow in which most keys are deactivated.




Figure 2-20 Optional Password Protected Functions

The user-selected password defaults to “6712” and may be changed in the Software Options that are available in the Extended Program mode. While the password may only be changed in the Extended Program mode, the program lock works for both Standard and Extended programs.

Should you forget your password, a key code can be obtained from Teledyne ISCO’s Customer Support department for a nominal fee.

## 2.12 Servicing the Sampler

Servicing the sampler includes:

- Recovering the sampler from the monitoring site
- Collecting the filled sample bottles and replacing them with clean bottles
- Replacing a battery, if required
- Collecting the data recorded by the sampler during the program.

Samplers are usually serviced after they complete a sampling program. When working up a service schedule, you will need to estimate the program completion time. You should also be familiar with the program’s settings for the start time or if the sampler is programmed for continuous sampling.

Continuous sampling is controlled by the distribution setting entered at the “RUN CONTINUOUSLY?” screen. A “YES” response directs the sampler to reset the distribution sequence and continue the countdown to the next sample event. Appropriate service intervals will prevent overfilled bottles.

### **2.13 Recovering the Sampler**

When recovering the sampler, keep it level to prevent spilling the collected samples.

If the battery’s charge is questionable, replace it with a fully-charged battery.

The sampler has a “fuel gauge” that gives an indication of power usage. The controller keeps track of how much power has been consumed since the last time it lost power. The current power consumption, as well as the previous power consumption, can be accessed by pressing the STOP key while in the main menu.

Keep in mind that the sampler has no idea as to how much charge was originally stored in the battery (or even if a battery is being used), and therefore has no idea as to the remaining capacity of its power source.

### **2.14 Collecting Sample Bottles**

If you return the entire sampler back to the office, place caps on the bottles and drain any water from the carrier or tub. If you leave the sampler at the site and only need to collect the bottles, follow these guidelines for bottle collection.

To collect the bottles:

1. Release the latches and open the refrigerated compartment.
2. Place caps on the bottles.
3. Remove the bottles. If configured for 14-bottle sampling, you can lift the bottle carrier from the compartment.
4. Attach a label to the bottle(s) detailing the location, date, etc.
5. Clean the inside of the refrigerated compartment if necessary.





# BLZZRD™

## Portable Refrigerated Sampler

---

### *Section 3 Programming Introduction*

Before programming the sampler, you should become familiar with its keys and how to use the screens.

The sampler has two sets of programming modes. The first mode, *Standard Programming*, lets you set up typical sampling programs easily and efficiently. The second mode, *Extended Programming*, lets you create sophisticated programs for complex sampling applications.


All of the sampling features available in standard programming mode are available in extended programming mode. However, extended programming provides many additional features not found in standard programming.

Standard and extended programming features are explained separately in sections 4 and 5 respectively. This section focuses on some basic features that are common to both programming modes, such as using the menus and setting the clock.

#### **CAUTION**

This sampler has a high performance pump. As with all such pumps, it relies upon liquid to cool working components. If the sampler is programmed to pump in the absence of liquid in excess of 5 minutes, excessive heat buildup may damage the paddles, rollers, and housing. Ensure that the liquid inlet is completely immersed.

### **3.1 Initial Operation**

Turn the sampler on by pressing the Standby button (  ). The start-up screen appears first.


```
BLZZRD
STANDARD PROGRAMMING
For HELP at any
screen press ? key.
```

It remains on the display for about eight seconds or until you press a key. It tells you the type of programming screens the sampler is currently using—standard or extended—and gives you a tip about the on-line help. The main menu screen, shown below, appears next.

```
RUN
PROGRAM
VIEW REPORT
OTHER FUNCTIONS
```

(If a 700 Series module is attached, a different screen may appear first. See the module manual for more information.)

**Table 3-1 About the Keypad**



KEY	NAME	FUNCTION
	Standby	Turns sampler on or off.
	Stop	Stops the pump or distributor. Pauses a running sampling program. In programming screens, returns to a previous screen.
	Enter	Accepts a menu choice or number entry and goes to next screen.
	Help	In programming screens, displays a brief help message.
	Down-Right Arrow	Selects the menu option right or below the current choice.
	Up-Left Arrow	Selects the menu option left or above the current choice.
	Numbers	Types a number.
	Decimal Point	Types a decimal point.
	Pump Reverse	Press when at the main menu to run the pump in reverse.
	Pump Forward	Press when at the main menu to run the pump forward.

The sampler is programmed at the factory so that the standard programming screens appear when you first turn the sampler on. You can switch between programming modes by typing in a numerical command at the main menu:

- At the main menu type **6712.2** to show the extended programming screens.
- At the main menu type **6712.1** to show the standard programming screens.

The sampler will continue to use the program mode you selected, even if the sampler is turned off and on again, until you manually switch to the other programming mode.

With the display backlight activated, the screen will glow for several seconds following your last keystroke.

If you press a key and the screen text does not change, it is probably because the keystroke turned the backlight on, instead of performing its function. Simply press the same key again.

### 3.1.1 Selecting a Language


You have the option of having screens display in English or another language. The default is English. At the main menu, type **6712.8**:



Use the arrow keys to move between the choices (languages other than those shown above could appear on your unit). When the language choice you want is blinking, press ↵ (Enter). The ↵ key always accepts the blinking option.

## 3.2 Using Menus and Entering Numbers

A menu is a list of options. The main menu has four options:



The options identify the operations that can be performed from the menu. For example, to run the sampling program, select RUN; to modify the program, select PROGRAM.

### 3.2.1 Selecting Menu Options

In menu screens, one menu option always blinks.

- Press ↵ when the blinking option is the one you want; it will accept your choice and go to the next screen.
- Select a different option by pressing an arrow key until the option you want blinks. Then press ↵.

#### **Note**

This instruction manual identifies the active blinking option by using a bold typeface.

### 3.2.2 Entering Numbers

Number-entry screens prompt you to enter a value. Enter numbers using the keypad buttons, then press ↵ to accept the value. See the example *Entering Numbers* for more details.


### Example: Entering Numbers

To enter a number:

Press the number keys to type the number. Then, press ↵. As soon as you press ↵, the sampler saves your number and moves to the next screen. In some screens, you can use the . (Decimal Point) key in a number.

Also note that some screens, such as the Suction Line Length screen, display the range of acceptable numbers between parentheses.

If an entry is too low or high, the controller beeps and erases the entry. Type a new number and continue.



SUCTION LINE LENGTH  
IS \_\_\_\_\_ ft  
(3-99)

### 3.3 Quick View Screens

Quick view screens are a special type of menu screen. They show the current program settings and let you move quickly through the program.

To see the Quick View screens, change the programming style to QUICK VIEW/CHANGE, as shown below:

1. From the Main Menu, select OTHER FUNCTIONS.



RUN  
PROGRAM  
VIEW REPORT  
**OTHER FUNCTIONS**

2. Select PROGRAMMING STYLE.



MAINTENANCE  
MANUAL FUNCTIONS  
**PROGRAMMING STYLE**

3. Select QUICK VIEW/CHANGE.



PROGRAMMING STYLE  
NORMAL  
**QUICK VIEW/CHANGE**

#### 3.3.1 Paging Through Quick View Screens

The arrows in the corners of each quick view screen are menu options that let you move from one quick view screen to another.


- Select the up arrow (↑) to go to the previous screen.
- Select the down arrow (↓) to go to the next screen.
- Press Stop to return to the main menu.

Because the forward arrow is always blinking when the screen appears, you can simply press ↵ to go to the next quick view screen, making it easy to page through the screens.

### Example: Using Quick View Screens


This example shows you how to change the sample volume.

1. Begin by paging through the quick view screens until you see the screen containing the sample volume setting:



RUN  
**PROGRAM**  
VIEW REPORT  
OTHER FUNCTIONS


2. The down arrow will be blinking. Continue paging through the screens by pressing the ↓ key. The Sample Volume screen will appear after several pages.



200 ml SAMPLES


3. Press either Arrow key until the 200 ml SAMPLES line blinks and press the ↓ key. The sampler

displays the normal view screen so you can view and modify the setting.



SAMPLE VOLUME:  
**400** ml (10-1000)

4. After modifying the setting, press the ↓ key. The sampler stores the new settings and returns to the updated quick view screen.



400 ml SAMPLES

You can repeat this process until all settings shown in the quick view screens have been changed. When you are done, select the screen's forward arrow to move to the next quick view screen, or press Stop to return to the main menu.

Note that you can step forward or backwards through the Quick View screens by selecting either the Up or the Down arrow.

#### 3.3.2 Changing Settings in a Quick View Screen

Although the quick view screens offer you a quick way to see the program settings, they also provide you with a way to change settings. Using quick view screens to change settings is sometimes a faster way to change a program because you can go quickly to the setting or settings that needs updating.

To change the program settings in a quick view screen, press an arrow key until the setting blinks. Then press ↓. The sampler displays the screen used to change the setting.

### Example: Setting the Clock and Calendar

This example shows you how to set the clock and calendar.

1. From the Main Menu, select OTHER FUNCTIONS.

```
  RUN
PROGRAM
VIEW REPORT
OTHER FUNCTIONS
```

2. Select MAINTENANCE.

```
MAINTENANCE
MANUAL FUNCTIONS
PROGRAMMING STYLE
```

3. Select SET CLOCK.

```
SET CLOCK
PUMP TUBE ALARM
INTERNAL BATTERY
DIAGNOSTICS
```

4. Use the number keys to type the current time and date.

```
ENTER TIME AND DATE:
HH:MM DD-MON-YY
14:__ __-__-__
```

Time is entered using a 24-hour format.

Note that you can skip fields by pressing the arrow keys. When you are done, press the Stop key to return to the main menu.

### 3.4 Clock and Calendar

You may have to set the internal clock or calendar. The samplers are shipped with their clocks set to U.S. Central Standard time. If your sampler operates in a different time zone, you will have to reset the clock. The example *Setting the Clock and Calendar* shows you how to find the time and date screen.

When setting the clock and calendar, use a 24-hour clock for times and the day-month-year format for dates. The example shows blanks for each position that accepts an entry. The positions are called fields. When this screen appears on your sampler, the fields will contain the sampler's current settings for the time and date.

- To move from one field to another without changing the setting, press the arrow keys. Use this technique when you want to change only one or two settings.
- Change the setting by typing a new number. Press  $\downarrow$  to accept the new setting.

For example, to enter 2:00 p.m. (2:00 p.m. is 14:00 on a 24-hour clock), type 14. Press  $\downarrow$ . Next, type 0 (zero) for the minutes, and press  $\downarrow$ . To enter a date, such as January 22, 2004, type: 22  $\downarrow$  01  $\downarrow$  04  $\downarrow$ .

### 3.5 Clock Start Time Menu Screens

There are three start time settings you can use in a sampling program. This section discusses one of these settings, the clock start time. A clock start time lets you set up a sampling program that starts at the same time on one or more days of the week. For example, you may want to begin sampling at 6:00 AM on Monday through Friday.

To use a Clock Start Time:

1. Step through the programming screens until you see the program delay screen.

NO DELAY TO START  
DELAYED START  
**CLOCK TIME**

2. Select CLOCK TIME. You will see the FIRST SAMPLE AT screen.

FIRST SAMPLE AT:  
**06:00**

3. Use the number keys to enter the Start time using the 24 hour clock format. Then, press ↵. The screen advances to the select the days of the week.

SELECT DAYS:  
SU **MO TU WE TH FR SA**  
DONE

4. To select a day or days from the menu:
  - a. Press an arrow key until the cursor is on the correct day. Then, press ↵.
  - b. Repeat these steps until each day you want is blinking.

**Note**

To remove days from the list, press an arrow key until the cursor is on the day you want to remove. Press ↵ to deselect it.

5. When finished, select DONE and press ↵.

### 3.6 Site Description and Program Name Menu Screens

The sampler has two text-entry menus, one for site description and another for an extended-program name. They work in the same way. A site description is commonly a number, address, or other short note that helps identify the monitoring site. Extended programming has a similar screen that lets you enter program names for stored programs.

The example *Entering a Site Description* shows how to change the site description from FACTORY01 to SITE 29. On the text entry screen line 1 contains text between two quotation marks. Lines 2 and 3 contain the menu options—numbers, letters, and punctuation marks—used to spell out the description. In line 3, the space between the ampersand (&) and the double quotes (") is a space character. Line 4 contains two additional options:

- BACK-UP, an option that moves the cursor on the text line to the left one character.
- DONE, an option that tells the sampler to save the text.

### Example: Entering a Site Description

The steps below show how to change the site description from *FACTORY01* to *SITE 29*. These steps are similar to changing an Extended Program Name.

1. From the Main Menu, select PROGRAM.

```
      RUN
    PROGRAM
  VIEW REPORT
OTHER FUNCTIONS
```

2. Select YES to change the Site Description.

```
SITE DESCRIPTION:
"FACTORY01"
  CHANGE?
  YES      NO
```

3. The far-left character is blinking. Use the arrow keys to move the cursor to the desired letter. For this example select "S." Press the  $\downarrow$  key and the display advances one character.

```
SITE: FACTORY01
ABCDEFGH IJKLMN OPQRST
UVWXYZ-& "0123456789
BACK-UP  DONE
```

4. Use the arrow keys to select "I" and press the  $\downarrow$  key.

```
SITE: SICTORY01
ABCDEFGHIIJKLMN OPQRST
UVWXYZ-& "0123456789
BACK-UP  DONE
```

5. Continue by entering T, E, and a space. Enter a 2 and a 9.

```
SITE: SITE 2901
ABCDEFGHIJKLMN OPQRST
UVWXYZ-& "0123456789
BACK-UP  DONE
```

6. Enter two more spaces to replace the "01." Move the cursor to "DONE" and press the  $\downarrow$  key.

```
SITE: SITE 29
ABCDEFGHIKL MN OPQRST
UVWXYZ-& "0123456789
BACK-UP  DONE
```

The cursor is the blinking rectangle that identifies the current field. When the screen first appears, the cursor blinks on the first character of the text in line 1. The matching character in line 2 or 3 also blinks.

Some additional tips for using the text-entry menus:

- Hold an arrow key down for faster cursor movement.
- If your description uses all ten spaces in the text line, the sampler automatically moves to the next screen as soon as you add the tenth character.
- You can enter numbers by pressing the number keys.
- Erase characters by replacing them with a space character.
- Select the BACK-UP option to move the cursor to the left.
- To leave the menu before completing the entry, press Stop. The sampler will display this menu:

```
SAVE CHANGES?
  YES      NO
```

Select YES to save any changes made, NO to leave the original description or program name unchanged.



### 3.7 Messages

The sampler has four sets of messages that supplement the programming screens. *Information messages* tell you about programming status. *Help Notes* are brief comments associated with programming screens describing the setting or offering assistance. *Warning messages* tell you something is different than the teledyne ISCO's standard. *Operation messages* report the sampler's status as it runs a program.

#### 3.7.1 Information Messages

These are brief messages about the programming status, such as "ROM PASSED TEST" or the status message that displays when you first start up the sampler. Information messages are displayed for four seconds, unless they are terminated by pressing any key on the keypad.

#### 3.7.2 Help Notes

To see a Help note:

1. Press the ? (Help) key.
2. When the note requires several screens, the word "more" appears in the lower-right corner. Press ↓ for the next note.
3. Move back and forth between screens by pressing the arrow keys.
4. Press Stop at any time to return to the programming screen.

Most help notes refer to a related section in this manual. Screens that require extensive explanation have notes that contain only a reference to the manual. Sections mentioned in the notes appear in the index and the table of contents.

### Example: Viewing Help Notes

This example shows how to view the Help note for the Suction Line Length screen. These steps apply to many programming screens.

1. From a programming screen, press the ? (Help) key.

```
SUCTION LINE LENGTH
IS 7 ft
(3-99)
```

2. The sampler displays the Help Note for current programming screen. Press ↓ or the Down-Right arrow key to see the next screen.

```
Enter the length of
suction line only,
excluding strainer
and pump tube. <more>
```

3. Press ↓ or the Down-Right arrow key to see additional screens.

```
See "Installing
Suction Line" in
your manual. <more>
```

4. While viewing Help Notes, the ↓ or the Down-Right arrow key displays the next screen. The Up-Left arrow will display the previous screen. And as the final screen indicates, pressing the Stop key will return you to the programming screen.

```
Press the red STOP
key to exit HELP.
```

### 3.7.3 Warning Messages

Warning messages appear when the sampler determines something is out of the ordinary. For example, the following warning appears when you type a number that is larger than the standard bottle volume:

```
BOTTLE VOLUME IS  
35000 ml (300-30000)
```

```
WARNING!  
STANDARD BOTTLE  
VOLUME EXCEEDED!
```

The sampler does accept nonstandard volumes because you might use a nonstandard container occasionally. For a list of recommended volumes for Teledyne ISCO bottles press the ? (Help) key at the “Bottle Volume Is” screen, or refer to the Technical Specifications.

The sampler uses the bottle and sample-volume settings to determine the maximum number of samples that can be deposited without overflowing the bottles. Entering a volume that exceeds the standard volume may cause the sampler to overflow the bottle.

### 3.7.4 Operation Messages

An operation message reports the sampler’s status as it runs a program. An example is shown below.

```
PROGRAM WILL START  
AT 06:00 TU 20-FEB  
05:56:22 TU 20-FEB
```

## 3.8 Menu Trees

Selecting a menu option will take you to a number-entry screen or another menu screen. The screens are organized in a branching structure that resembles a tree. Refer to Figure 3-1 for a simple chart of the menu tree for standard programming. The standard programming structure is different—much simpler—than the extended programming structure. Charts that show you most of the sampler’s screens appear in Appendix A, *Menu Flowcharts*.

The entire structure changes a little when a module is attached because the sampler adds a set of screens for a module as soon as it detects the module’s presence. The manuals for each module contain menu trees for its programming screens.




Figure 3-1 Menu Tree for Standard Programming

### 3.9 Memory

The sampler controller contains 512 kilobytes of battery-backed RAM (Random Access Memory) that stores:

- Five sampling programs
- The sampling results from the most recent program
- The rainfall data collected from rain gauges
- The data from a refrigerator temperature sensor
- The data collected from a 700 Series module
- Parameter data collected from an SDI-12 Sonde
- Calculated flow information from the current flow module

The sampling results memory stores up to approximately 1004 events (the actual number is dependent on the number of SDI-12 parameters). The program run and start time information is logged as four events. These four entries are held for the duration of the program. The remaining 1000 are for logging sample events, program events, and errors from the currently running program. If the running program exceeds 1000 events, the data will “roll over.” That is, the sampler will replace the oldest stored event with the newest event. When using an SDI-12 sonde, the sampler logs additional entries at each sample event.

Module, sonde, refrigerator temperature, and rain gauge readings are stored in memory partitions. The storage capacity depends on the reading types, the number of bytes per reading, and the data storage interval.

The sampler clears the sampling results memory when the site description changes. It is important to collect the sampling results before changing the site description—the data can not be retrieved afterwards. Keep in mind that the description may change when selecting an extended program.

The partition data is cleared when the sampler automatically creates and sizes the memory partitions. This occurs when you:

- Select a new program with a different storage interval
- Change the Data Storage Interval in the current program
- Change the programming mode
- Install a new 700 Series Module
- Change 700 Series Module mode of operation
- Add or remove a rain gauge with the Hardware Setup
- Add or remove a SDI-12 sonde parameter with the Hardware Setup
- Reinitialize the sampler

The stored data is lost as the sampler reformats its memory. As a precaution, before clearing the memory, the sampler display will show the following message:

```
INTERVAL CHANGED --  
DOWNLOAD DATA NOW  
OR LOSE ALL DATA!  
DONE
```

If you want to keep the stored data, download it before proceeding. When DONE is selected, the sampler clears the stored data and reformats the memory partitions. The sampler then displays the new capacity of the partition memory.

```
PARTITION MEMORY  
WILL LAST 13 DAYS
```

If the duration is too short, try changing the data storage interval or eliminating unnecessary reading types in the Hardware or Module Setup. For example, a sampler configured with a 730 Bubbler Module and a rain gauge storing data at a one-minute interval, creates level and rainfall partitions that hold 100 days of readings. Changing the storage interval to two minutes extends the storage capacity to 201 days.

The sampler does not have to be running a program to store readings. When the sampler is on, it stores readings at the selected data storage interval. Once the partition is filled, the readings roll over.

When the sampler is running a program, the partitions become “triggered” by the programmed start time or the first enable time, whichever is later. A triggered partition continues to log readings at the selected interval, but will not replace any readings taken after the trigger event. The sampler fills the partition and then stops recording to preserve the data. The first reading preserved in a full partition is one hour before the trigger event. A full partition can be reset by starting another program or changing storage parameters.

If you need continuous readings for analysis, be sure to collect the readings before the partition fills or rolls over to avoid gaps in the data



# BLZZRD™

## Portable Refrigerated Sampler

### Section 4 Standard Programming

The BLZZRD has two sets of programming screens. One set, called standard programming screens, lets you set up typical sampling programs easily and efficiently. The other set, extended programming screens, lets you create sophisticated programs for complex sampling applications.

All of the sampling features available within standard programming are available in extended programming. However, extended programming provides many additional features which are discussed in Section 5. Menu flowcharts for both sets of programming screens are in Appendix A.

#### 4.1 Switching between Standard and Extended Modes

As a factory default, the first time you turn the sampler on it will begin in standard programming mode. The start-up screen tells you which programming mode the sampler is currently using. You can switch between programming modes by entering a numerical command at the main menu:

```
      RUN
    PROGRAM
VIEW REPORT
OTHER FUNCTIONS
```

- Type **6712.1** to enter standard programming mode.
- Type **6712.2** to enter extended programming mode.

#### 4.2 Selecting Language and Units of Length

You have the option of displaying screens in English or Spanish (or possibly another language). The default is English. At the main menu, type **6712.8** to display:

```
ENGLISH
SPANISH
```

Use the arrow key to move between the two choices. When the desired choice is blinking, press ↵.

If you select the English option, you will advance to a second screen that asks you to select your units of length:

```
SELECT UNITS FOR
LENGTH:
ft      m
```

Use the arrow key to move between the two choices, feet or meters. When the desired choice is blinking, press ↵. The default unit of length for the Spanish option is meters.

### 4.3 Programming Examples

The following examples show you simple time-paced and flow paced sampling programs. Details about many of the menu items are provided in later sections.

#### Example: Standard Program – Sample every 15 Minutes, One Sample per Bottle

This example shows you how to program the controller to take a time-paced sample at 15 minute intervals. The 1000 ml samples will be placed in separate bottles. The example displays are shown in the Normal programming style.

1. Turn the sampler on by pressing the On/Off key. Press ↵. The start-up screen disappears after eight seconds.

```

BLZZRD
STANDARD PROGRAMMING
For HELP at any
screen press ? key.
  
```

2. The PROGRAM option will be blinking. Press ↵.

```

RUN
PROGRAM
VIEW REPORT
OTHER FUNCTIONS
  
```

3. The option NO will be blinking. Press ↵. For the purposes of this example, it isn't necessary to change the description.

```

SITE DESCRIPTION:
"FACTORY051"
CHANGE?
YES NO
  
```

4. Select the desired temperature units.

```

SELECT UNITS FOR
TEMPERATURE
°F °C
  
```

5. Select the number of bottles in your bottle kit by pressing either arrow key until the correct number blinks. Press ↵. For this example, select 14.

```

NUMBER OF BOTTLES:
1 4 14
  
```

6. Type the volume for the bottles in your kit. For this example, 950 is correct, so simply press ↵.

```

BOTTLE VOLUME IS
950 ml (300-3000)
  
```

7. Type the length of the suction line, then press ↵. If you change the length, the sampler will display a message, "PLEASE WAIT!...GENERATING PUMP TABLES."

```

SUCTION LINE LENGTH
IS 10 ft
(3-99)
  
```

8. Because this example requires samples every 15 minutes, select TIME PACED by pressing an arrow until the option TIME PACED blinks. Then, press ↵.

```

TIME PACED
FLOW PACED
  
```

9. Type 0 for hours and press ↵. Type 15 for minutes and press ↵. *Tip:* Move back and forth between hours and minutes by pressing an arrow key.

```

TIME BETWEEN
SAMPLE EVENTS
0 HOURS, 15 MINUTES
  
```

10. Because this program requires one sample in each bottle, select SEQUENTIAL by pressing an arrow until the option blinks. Then, press ↵.

```

SEQUENTIAL
BOTTLES/SAMPLE
SAMPLES/BOTTLE
  
```

Continued...



*Standard Program – Sample every 15 Minutes, One Sample per Bottle Continued*

11. For this example, select NO by pressing an arrow until the option blinks. Then, press ↵. Selecting YES allows the program to run indefinitely by repeating the sample distribution. Continuous sampling assumes that filled bottles are replaced with empty bottles at regular service intervals.

RUN CONTINUOUSLY?  
YES **NO**

12. Type the volume of the sample you want deposited in each bottle. Then, press ↵.

SAMPLE VOLUME:  
**200** ml (10-1000)

13. For this example, select DELAYED START by pressing an arrow until the option blinks. Then, press ↵.

NO DELAY TO START  
**DELAYED START**  
CLOCK TIME

14. Type the delay period you want between the time you run the program and the time the sampler takes the first sample. Then, press ↵.

FIRST SAMPLE  
AFTER A  
**5** MINUTE DELAY  
(1-999)

15. Run the program immediately by selecting YES. Select NO if you want to run the program later by selecting RUN from the main menu. Press ↵ after making your choice. In this example, NO is selected.

PROGRAMMING COMPLETE  
RUN THIS PROGRAM  
NOW?  
YES **NO**

16. Run the program by selecting RUN and pressing ↵.

**RUN**  
PROGRAM  
VIEW REPORT  
OTHER FUNCTIONS

## Example: Standard Program – Flow-Paced Sampling, Two Bottles Per Sample

This example shows you how to program the controller to take flow-paced samples with the following settings: The example displays are shown in the Normal programming style.

- Program type: Standard
- Site description: Default
- Bottle kit: 14, 950 milliliter bottles
- Suction line length: 7 feet
- Pacing: Flow pacing, every two pulses
- Distribution: 2 bottles per sample
- Sample Volume: 250 milliliters
- Start time: Clock Time, 6:00 AM on Monday, Wednesday, Friday
- Sampling duration: 24 hours
- Module: No module installed

1. At the Main Menu, select PROGRAM and press ↵.

```

    RUN
    PROGRAM
    VIEW REPORT
    OTHER FUNCTIONS
    
```

2. The option NO will be blinking. Press ↵.

```

    SITE DESCRIPTION:
    "FACTORY01"
    CHANGE?
    YES  NO
    
```

3. Select the desired temperature units.

```

    SELECT UNITS FOR
    TEMPERATURE
    °F °C
    
```

4. Select 14. Press ↵.

```

    NUMBER OF BOTTLES:
    1 4 14
    
```

5. Type 950 and press ↵.

```

    BOTTLE VOLUME IS
    950 ml (300-30000)
    
```

6. Type 7, then press ↵.

```

    SUCTION LINE LENGTH
    IS 7 ft
    (3-99)
    
```

7. Select FLOW PACED, then, press ↵.

```

    TIME PACED
    FLOW PACED
    
```

8. Enter the number of pulses between sample events. Press ↵.

```

    FLOW BETWEEN
    SAMPLE EVENTS:
    2 PULSES (1-9999)
    
```

9. Select BOTTLES/SAMPLE.

```

    SEQUENTIAL
    BOTTLES/SAMPLE
    SAMPLES/BOTTLE
    
```

10. Enter 2.

```

    2 BOTTLES PER
    SAMPLE EVENT (1-14)
    
```

11. For this example, select NO to disable the Run Continuously option. Selecting YES allows the program to run indefinitely by repeating the sample distribution and assumes that filled bottles are replaced at regular service intervals.

```

    RUN CONTINUOUSLY?
    YES  NO
    
```

12. Type 250. Then, press ↵.

```

    SAMPLE VOLUME:
    250 ml (10-1000)
    
```

*Continued...*

*Standard Program – Flow-Paced Sampling, Two Bottles Per Sample Continued*

13. Select CLOCK TIME.

NO DELAY TO START  
 DELAYED START  
**CLOCK TIME**

14. Enter 6, then enter 0. This is the clock time at which the sampler will start the pacing countdown.

START FLOW COUNT AT:  
 06:00

15. Select MO, WE, and FR.  
Use the arrow keys and ↵ to select the days. When the correct days are blinking, select DONE and press ↵.

SELECT DAYS:  
 SU MO TU **WE** TH FR SA  
 DONE

16. Enter 48. This setting will stop the sampler after 48 hours. This program may stop sooner if all bottles are full.

MAXIMUM RUN TIME:  
**48** HOURS

17. Select NO.

PROGRAMMING COMPLETE  
 RUN THIS PROGRAM  
 NOW?  
 YES **NO**

## 4.4 Pacing

Sample pacing is the rate at which the sampler takes samples. Depending on the type of pacing you select, the rate is controlled by the sampler's internal clock or by inputs received from connected instruments.

Standard programming provides time pacing and flow pacing (refer to Step 7 in the previous programming example). When programming, you will be asked to select time or flow paced sampling. Depending upon your choice, you will see one of the following screens:

TIME BETWEEN  
 SAMPLE EVENTS:  
 \_\_\_ HOURS, \_\_\_ MINUTES

OR

FLOW BETWEEN  
 SAMPLE EVENTS:  
 \_\_\_\_\_ PULSES (1-9999)

### 4.4.1 Time Pacing

In time-paced sampling, the interval between samples is a constant time interval. When you program the sampler for time pacing, the sampler prompts you to enter the time between sample events in hours and minutes. Time paced programs always take a sample at the start time.

### 4.4.2 Flow Pacing

Flow paced sampling requires a flow meter, flow logger, or a 700 Series flow module. A flow meter or flow logger paces a sampler by sending an electronic signal to the sampler after measuring a specified volume of liquid. Because each pulse represents a volume interval, flow pacing rates are proportional to the volume of water flowing through the channel.

When you program the sampler for flow pacing and are using a flow meter or flow logger, the sampler prompts you to enter the interval between sample events in pulses. The sampler initiates a sample event when the set number of pulses is received. Programs that are flow paced do not take a sample at the start time.

After setting the program start time for a flow paced program, you will be prompted to enter a Maximum Run Time. Setting a maximum run time allows you to stop the program after a duration of 1 to 999 hours. To run a flow paced program indefinitely, enter 0 (zero) hours.

The flow pacing screens change when you attach a flow module.

FLOW BETWEEN SAMPLE EVENTS : _____ Mgal (0.001 -99999)
---

Because the sampler is more closely integrated with the 700 Series flow modules, the standard flow pacing screen prompts you for the flow volume between sample events instead of pulses between events. The flow volume units displayed are what you had previously programmed.

#### 4.4.3 Trigger Pacing

Teledyne ISCO 2100 and 4200 Series Flow Meters and 4100 Series Flow Loggers send pulses for trigger pacing. The flow meter or flow logger sends pulses at two different time intervals when trigger pacing a sampler. Trigger pacing, for example, lets you sample less frequently when the trigger condition—a condition based on level, flow, or rainfall—remains below a set point. The intervals are determined by the flow instrument's program settings for trigger pacing.


When conditions change and readings pass the set point, triggering the new rate, you can sample more frequently. Or, when the trigger condition is a clock setting, the instrument changes the pacing interval according to the time of day. Using a clock trigger condition, for instance, the flow meter or flow logger can pace the sampler slowly at night and more frequently during the day. (For more information about trigger pacing, see your flow meter or Flowlink Help.)

The flow meter or flow logger sends the same pulse for both flow and trigger pacing, and the sampler cannot distinguish between them. Therefore, to program a sampler for trigger pacing, simply program it for flow pacing.

#### 4.5 Distribution


Distribution describes how the sampler is to deposit samples. A sample is the volume of liquid deposited in a bottle. A sample event includes the full sampling cycle and may deposit a sample into more than one bottle.

**Sequential**




Each bottle receives one sample from one sample event.

**Bottles-per-Sample**



Multiple bottles receive a sample from one sample event.

**Samples-per-Bottle**



Each bottle receives a sample from multiple sample events.

Figure 4-1 Sample Distribution

In standard programming, you can program the sampler for these different distribution methods:

- Sequential
- Bottles Per Sample
- Samples Per Bottle
- Composite

Figure 4-1 illustrates the programming screens for sequential, bottles per sample, and samples per bottle distribution.

#### **4.5.1 Sequential**

In sequential distribution, the sampler deposits one sample in each bottle. A sequential sample represents a “snapshot” of the flow stream at a point in time.

#### **4.5.2 Bottles Per Sample**

In bottles per sample distribution, the sampler deposits a sample in each of a set of bottles. A bottle set includes at least two bottles but may include all bottles. Use bottles per sample when the volume to be collected is larger than the amount one bottle can hold or when you need identical samples.

#### **4.5.3 Samples Per Bottle**

In samples per bottle distribution, the sampler deposits samples from several sample events in a single bottle before moving to the next bottle. Use samples per bottle distribution to collect a series of small composite samples.

#### **4.5.4 Composite**

For single bottle configurations, samples per bottle distribution is known as composite sampling. A composite sample represents an average of the flow stream’s characteristics during the sampling period.

#### **4.5.5 Continuous Sampling**

Sample programs can be run indefinitely by selecting “YES” at the “RUN CONTINUOUSLY?” screen. Continuous sampling resets the distribution when the distribution sequence is complete. That is, when the last bottle/set is reached, the next bottle/set is the first bottle/set. All pacing modes except RANDOM are supported.

When running a program continuously, the bottles must be serviced at regular intervals to prevent overfilling the bottles. The sampler assumes that the next bottle/set is empty and ready to receive samples.

### **4.6 Start Times**

It is important to understand the difference between the time at which you run a program and the program’s start time. Running a program simply means selecting RUN from the sampler’s main menu. The start time is the time at which the sampler begins the program’s first sample interval countdown. The start time is controlled by your selections from the start time screens.

Each program contains start-time settings that tell the sampler when to begin the program. When programming the sampler, you can select one of three start time options: NO DELAY TO START, DELAYED START, or CLOCK TIME.

## Example: Start Times

**Delayed Starts** count down before running a sampling program. To enable a Delayed Start:

1. Select DELAYED START.

```
NO DELAY TO START
DELAYED START
CLOCK TIME
```

2. If the program is time-paced, enter the number of minutes to delay the first sample.

```
FIRST SAMPLE
AFTER A
— MINUTE DELAY
(1-999)
```

If the program is flow-paced, enter the number of minutes to delay the flow pulse counting.

```
START FLOW COUNT
AFTER A
— MINUTE DELAY
(1-999)
```

**Clock Times** delay a sampling program until a specified time and day of week. To enable this feature:

1. Select CLOCK TIME.

```
NO DELAY TO START
DELAYED START
CLOCK TIME
```

2. Enter the time of day using the 24-hour clock format.

```
FIRST SAMPLE AT
06:00
```

3. Select the day or days of the week that the program can begin on.

```
SELECT DAYS:
SU MO TU WE TH FR SA
DONE
```

Press an arrow key until the cursor is on the correct day. Press ↵. Repeat until each day you want is blinking. Then, press an arrow until DONE blinks. Press ↵.

- Select NO DELAY TO START when you want the sampler to start as soon as you select RUN.
- Select DELAYED START when you want the sampler to delay from 1 to 999 minutes before starting the program.
- Select CLOCK TIME when you want the sampler to begin the program at a specific time on at least one day of the week.

### 4.6.1 How Do Start Times Work?

As soon as you select RUN from the main menu, the sampler checks the program's start time settings. With a "DELAYED START" setting, the sampler starts a countdown to the start time. The period between the time you select RUN and the start time you've specified in the program is called the "delay to start time."

When you run a program with "CLOCK TIME" settings, the sampler also starts a countdown to start time until the day of week and clock times are met.

The sampler may not collect a sample at the start time. The sampler must be enabled, and the pacing selected must call for a sample at start. In any event, the sampler checks the pacing settings and begins the pacing-interval countdown.

#### 4.6.2 Sampler Enable/Disable

Teledyne ISCO flow meters and flow loggers have a programmable sampler-enable feature that lets them enable (start) or disable (stop) a program according to certain monitored conditions. For example these conditions can be level, flow rate, pH, temperature, percent, rainfall, I/O; or a combination of two conditions.

The Model 1640 Liquid Level Actuator can also be used, as well as any other instrument that can ground the input to pin F. Grounding the input disables the sampler.

The sampler disregards disable signals during the delay to start time. However, when the sampler detects a disable signal at the start time, it suspends the program until it is enabled. The diagram in Figure 4-2 shows the sampler's responses.




Figure 4-2 Start Time Diagram

#### 4.7 Programming for 700 Series Modules

The bay on the controller's side accepts any of Teledyne ISCO's 700 Series Modules. These modules are optional accessories that are not required for operation of the sampler. However, the modules offer an economical way to combine flow-rate or parameter monitoring with sampling.

The programming of these modules is integrated with the programming of the sampler, so attachment of a module will cause different screens to be displayed, depending on the type of module and the selections you make while programming.



Two modules that are commonly used are the 730 Bubbler Module, which monitors a flow stream's level and flow rate, and the 750 Area Velocity Module for monitoring a flow stream's level, velocity, and flow rate. The following examples step through screens you might see while programming the sampler and module. Refer to the example *Standard Program – Flow-Paced Sampling, Two Bottles Per Sample* (which does not have a module attached) at the beginning of this section to compare the sequence of screen displays.

 **CAUTION**

Most Teledyne ISCO accessories and interfacing equipment are not certified for hazardous locations as defined by the National Electrical Code. Check applicable guidelines, codes, and regulations of federal, state, city, and county agencies. Refer to Appendix C in the back of this manual for more safety information.

### Example: Programming with 730 Bubbler Flow Module Installed

1. With the sampler off, insert the module in the bay on the controller.
2. Turn the sampler on by pressing the On/Off key. Press ↵. Download stored data, if you want to keep it. Pressing DONE erases it and advances to the next screen.

```

MODULE INSERTED---
DOWNLOAD DATA NOW
OR LOSE ALL DATA!
DONE
  
```

3. The startup screen appears for 8 seconds.

```

BLZZRD
STANDARD PROGRAMMING
For HELP at any
screen press ? key.
  
```

4. The option PROGRAM will be blinking. Press ↵.

```

RUN
PROGRAM
VIEW REPORT
OTHER FUNCTIONS
  
```

5. The option NO will be blinking. Press ↵.

```

SITE DESCRIPTION:
"FACTORY051"
CHANGE?
YES NO
  
```

6. Select the desired temperature units.

```

SELECT UNITS FOR
TEMPERATURE
°F °C
  
```

7. Select the desired units for the flow rate. For this example, choose cfs. Press ↵.

```

SELECT UNITS FOR
FLOW RATE:
cfs gps gpm Mgd
lps m3s m3h m3d
  
```

8. Select the desired units for the flow volume. For this example, choose Mgal. Press ↵.

```

SELECT UNITS FOR
FLOW VOLUME:
cf gal Mgal
m3 lit
  
```

9. Select YES to program the module.

```

PROGRAM MODULE?
YES NO
  
```

10. If you choose FLOWMETER, proceed to step #11. If you choose LEVEL ONLY, proceed to step #14.

```

MODE OF OPERATION:
FLOWMETER
LEVEL ONLY
  
```

11. When you select Flowmeter, you will have several options for performing the flow conversion. For this example choose Weir. Press ↵.

```

FLOW CONVERSION:
WEIR FLUME
DATA POINTS
MANNING FLOW-INSERT
  
```

...Continued

Programming with 730 Bubbler Flow Module Installed, Continued...

12. For this example, choose V-Notch. Press ↵.

V-NOTCH  
RECTANGULAR  
CIPOLLETTI

13. For this example, choose a Weir angle of 90°.

V-NOTCH WEIR ANGLE:  
120° 90° 60°  
45° 30° 22.5°

14. The current level will be displayed. Enter an adjustment if needed. If the adjustment is different by more than 0.50 ft, you will be prompted "Are you sure? Yes No" Choose no to re-enter; select yes if correct.

CURRENT LEVEL IS  
\_\_ . \_\_ ft  
ADJUST LEVEL TO  
\_\_ . \_\_ ft

15. Select the number of bottles in your bottle kit by pressing either arrow key until the correct number blinks. Press ↵. (For this example, select 14.)

NUMBER OF BOTTLES:  
1 4 14

16. Type the volume for the bottles in your kit. For this example, 950 is correct, so simply press ↵.

BOTTLE VOLUME IS  
950 ml (300-30000)

17. Type the length of the suction line, then press ↵.

SUCTION LINE LENGTH  
IS 10 ft  
(3-99)

If you change the length, the sampler will briefly display a message, "PLEASE WAIT!... GENERATING PUMP TABLES."

18. Select FLOW PACED. If the module is being used as a flowmeter, you will see screen #19a. If the module is being used for level only, you will see screen #19b.

TIME PACED  
FLOW PACED

19. If the module is being used as a flowmeter, choose the amount of flow between sample events. If level

only, enter the number of pulses between sample events. Flow pulses must be provided by an external flow meter.

FLOW BETWEEN  
SAMPLE EVENTS:  
0.075 Mgal  
(0.001 - 99999)

FLOW BETWEEN  
SAMPLE EVENTS:  
2 PULSES (1-9999)

20. Select BOTTLES/SAMPLE.

SEQUENTIAL  
BOTTLES/SAMPLE  
SAMPLES/BOTTLE

21. Enter 2.

2 BOTTLES PER  
SAMPLE EVENT (1-14)

22. Select NO.

RUN CONTINUOUSLY?  
YES NO

23. Enter 250.

SAMPLE VOLUME:  
250 ml (10-1000)

24. Select NO DELAY TO START.

NO DELAY TO START  
DELAYED START  
CLOCK TIME

25. Enter 0. This setting will cause the sampler to run indefinitely until stopped by user intervention or event such as a full-bottle condition.

MAXIMUM RUN TIME:  
0 HOURS

26. Select NO.

PROGRAMMING COMPLETE  
RUN THIS PROGRAM  
NOW?  
YES NO

### Example: Programming with 750 Area Velocity Module Installed

- With the sampler off, insert the module in the bay on the controller.
- Turn the sampler on by pressing the On/Off key. Press ↵. Download stored data, if you want to keep it. Pressing DONE erases it and advances to the next screen.

```

MODULE INSERTED---
DOWNLOAD DATA NOW
OR LOSE ALL DATA!
DONE
    
```

- The startup screen appears for 8 seconds.

```

BLZZRD
STANDARD PROGRAMMING
For HELP at any
screen press ? key.
    
```

- The option PROGRAM will be blinking. Press ↵.

```

RUN
PROGRAM
VIEW REPORT
OTHER FUNCTIONS
    
```

- The option NO will be blinking. Press ↵.

```

SITE DESCRIPTION:
"FACTORY051"
CHANGE?
YES NO
    
```

- Select the desired temperature units.

```

SELECT UNITS FOR
TEMPERATURE
°F °C
    
```

- Select the desired units for the flow rate. For this example, choose cfs. Press ↵.

```

SELECT UNITS FOR
FLOW RATE:
cfs gps gpm Mgd
lps m3s m3h m3d
    
```

- Select the desired units for the flow volume. For this example, choose Mgal. Press ↵.

```

SELECT UNITS FOR
FLOW VOLUME:
cf gal Mgal
m3 lit
    
```

- Select the desired units for the flow rate. For this example, choose fps. Press ↵.

```

SELECT UNITS FOR
VELOCITY:
fps mps
    
```

- Select YES to program the module.

```

PROGRAM MODULE?
YES NO
    
```

- The 750 has two modes of operation: FLOWMETER or LEVEL AND VELOCITY. For this example, choose FLOWMETER.

```

MODE OF OPERATION:
FLOWMETER
LEVEL AND VELOCITY
    
```

- When you select Flowmeter, you will have two options for performing the flow conversion: AREA \* VELOCITY or LEVEL TO FLOW RATE. For this example, choose AREA \* VELOCITY.

```

FLOW CALCULATION:
AREA*VELOCITY
LEVEL TO FLOW RATE
    
```

- For this example, choose Channel Shape.

```

LEVEL TO AREA:
CHANNEL SHAPE
DATA POINTS
    
```

- For the channel shape, select Round Pipe.

```

ROUND PIPE
U-CHANNEL
RECTANGULAR CHANNEL
TRAPEZOIDAL CHANNEL
    
```

- Enter the diameter of the round pipe. This example shows a two-foot diameter pipe.

```

ROUND PIPE
DIAMETER = 2.00 ft
    
```

Continued...

*Programming with 750 Area Velocity Module Installed, continued...*

16. The current level will be displayed. Enter an adjustment if needed. If the adjustment is different by more than 0.50 ft, you will be prompted “Are you sure? Yes No”. Choose NO to re-enter; select YES if correct.

```
CURRENT LEVEL IS
  _ . _ ft
ADJUST LEVEL TO
  _ . _ ft
```

17. Enter the amount of zero level offset. Refer to the 750 Area Velocity manual for more information on zero level offsets.

```
ZERO LEVEL OFFSET:
  _ . _ ft
```

18. If you are using a standard-size sensor, select the maximum depth for the velocity measurement. This screen does not appear if using a low-profile sensor which is automatically set to 1 inch.

```
MAXIMUM DEPTH FOR
VELOCITY MEASUREMENT
  2"   3"   4"
```

19. From this point forward, the screens will be the same as steps 15 through 26 in the example “Programming with 730 Bubbler Flow Module Installed.”

```
NUMBER OF BOTTLES:
  1   4   14
```

## 4.8 Other Functions

In Standard Programming, the “Other Functions” menu screen includes options for:

- Maintenance (see Section 9).
- Manual Functions (see Section 6.6).
- Programming Style (see Section 3.3).

You can access these options by selecting OTHER FUNCTIONS from the Main Menu.

```
RUN
PROGRAM
VIEW REPORT
OTHER FUNCTIONS
```

```
MAINTENANCE
MANUAL FUNCTIONS
PROGRAMMING STYLE
```



# BLZZRD™

## Portable Refrigerated Sampler

### Section 5 Extended Programming

The BLZZRD sampler has two sets of programming screens. The first set, called standard programming screens, lets you set up typical sampling programs easily and efficiently. The second set, extended programming screens, lets you create sophisticated programs for complex sampling applications.

All of the sampling features available with standard programming screens, discussed in Section 4, are available with extended programming screens. However, the extended screens provide several additional features and some screen display differences which are discussed in this section. The menu flowcharts for both sets of programming screens are in Appendix A, *Menu Flowcharts*.

#### **Note**

This section assumes that you have read, and are familiar with, the instructions in Sections 3 and 4.

#### 5.1 Switching Between Standard and Extended Modes

As a factory default, the first time you turn the sampler on it will begin in standard programming mode. The start-up screen tells you which programming mode the sampler is currently using. You can switch between programming modes by entering a numerical command at the main menu:

<p style="text-align: center;"><b>RUN</b> PROGRAM VIEW REPORT OTHER FUNCTIONS</p>
---

- Type **6712.1** to enter standard programming mode.
- Type **6712.2** to enter extended programming mode.

#### 5.2 One-Part and Two-Part Programs

Extended programming lets you set up a “one-part” program or a “two-part” program. One-part programs let you fill all bottles of the tub with one set of pacing, distribution, and enable settings. Two-part programs add an additional set of extended pacing, distribution, and enable screens to the sampling program. Each set of screens is called a program part. The program parts are simply called part ‘A’ and part ‘B’. Both parts share the program settings for suction-line length, suction head, and rinses and retries. They also share one start-time setting.

Each part has its own group of bottles. Because settings for each part are independent of each other, the sampler, when running a two-part program, fills each bottle group as if they were being filled by two different programs. Two-part programs finish when both parts are done.

Two-part programming is ideal for sampling storm events. Many storm-water run-off monitoring permits require a first-flush grab sample within the first 30 minutes of the storm event and flow-paced samples during the remainder of the event. With a two-part program, you can set up part 'A' for the first-flush sample and part 'B' for the remainder. The example "*Storm Water Runoff Sampling Program*" shows such a program.

### 5.3 Storage for Extended Programs

The sampler stores five sampling programs: one standard and four extended. Program storage eliminates the need to reprogram the sampler with frequently used settings. This feature is especially useful for complex extended programs such as two-part programs or programs with many nonuniform-time settings.

Teledyne ISCO ships the sampler to you with factory programs that you can modify to suit your own needs. All four extended programs are identical to each other.

The stored programs will also save your 700 Series Module program settings and any sampler enable conditions using rain gauge data or SDI-12 sonde parameters. These program settings remain as long as you do not change the type of module, or change the rain gauge or SDI-12 Hardware Setup.

#### 5.3.1 Selecting a Stored Program

When the sampler is in extended programming mode, it always has one current and three noncurrent programs, each occupying a permanent location in memory. The first time you access the extended mode, the current program is "EXTENDED 1," one of the factory extended programs. The name of the current extended program appears in quotations on the first line of the main menu.

The factory programs are named "EXTENDED 1," "EXTENDED 2," "EXTENDED 3," and "EXTENDED 4." You can rename them with more descriptive names, making them easier to identify. The program name is one of the extended program's settings.

After selecting a stored program, review the settings with the quick view screens. Each sampling program stores all the settings required by the sampler to run a program. These settings include the bottle-kit information (number of bottles and the bottle volume), the suction-line length, the site description, start times, and so on.

If you use the same program at two sites—each requiring different bottle configurations, suction line lengths, or other details—always check these settings before running a program. You may need to make minor changes to the program to make sure the settings match your equipment.



### Example: Selecting a Stored Extended Program

1. Press the On/Off button to turn the sampler on. Press ↵ to clear the start-up message. If the words "STANDARD PROGRAMMING" appear on the second line of the start-up message, the sampler is using the standard programming screens. To see the extended screens, type "6712.2" at the main menu shown in step 2.

```
BLZZRD
EXTENDED PROGRAMMING
For HELP at any
screen, press ? key.
```

2. Select PROGRAM from the main menu.

```
RUN "EXTENDED 1"
PROGRAM
VIEW REPORT
OTHER FUNCTIONS
```

3. Select PROGRAM NAME: "EXTENDED 1" from the quick view menu. You can go through QUICK VIEW screens quickly by pressing ↵ at each screen. You can also change settings in quick view screens; see section 3.3, *Quick View Screens*.

```
↑ PROGRAM NAME:
"EXTENDED 1"
SITE DESCRIPTION:
" " ↓
```

4. Select SELECT NEW PROGRAM to change the program.

```
SELECT NEW PROGRAM
CHANGE PROGRAM NAME
```

5. Select the name of the extended program you want to use. Each program's name appears on the same line. If you inadvertently give two programs identical names, you can still identify them by their location on the screen.

```
"EXTENDED 1"
"EXTENDED 2"
"EXTENDED 3"
"EXTENDED 4"
```

6. Step through the remaining programming screens. To run the program immediately, select YES. To run the program later, select NO.

```
PROGRAMMING COMPLETE
RUN THIS PROGRAM
NOW?
YES NO
```

Alternatively, you can press the Stop button after selecting the new program. The display returns to the Main screen with "EXTENDED 2" as the selected program.

```
RUN "EXTENDED 2"
PROGRAM
VIEW REPORT
OTHER FUNCTIONS
```

The current program is the program you see when you select PROGRAM from the main menu and page through the programming screens. It is the program the sampler uses when you select RUN from the main menu.

You cannot delete programs, but you can modify them as often as necessary. To run a program with different settings without losing the settings in your current program, select a stored program and modify its settings.

For example, assume you have been running the current program, named "EXTENDED 1," at site A for several weeks. This program contains all the settings you need for site A, and you don't want to lose those settings.

Nevertheless, you need another program because you want to move the sampler to site B, so you select “EXTENDED 2” and modify it for site B. The settings in “EXTENDED 1” are preserved automatically. When you return the sampler to site A, you won’t need to reprogram the sampler. Merely select “EXTENDED 1” again, double-check the program settings, and run the program. To select an extended program, follow the steps in the example “*Selecting a Stored Extended Program.*”

## 5.4 Programming for 700 Series Modules

The bay on the controller’s side accepts any of Teledyne ISCO’s 700 Series Modules. These optional accessories are not required to operate the sampler. However, the modules offer an economical way to combine flow-rate or parameter monitoring with sampling.

The programming of these modules is integrated with the programming of the sampler, so attachment of a module will cause different screens to be displayed, depending on the type of module and the selections you make while programming.

Two modules that are commonly used are the 730 Bubbler Module, which monitors a flow stream’s level and flow rate, and the 750 Area Velocity Module for monitoring a flow stream’s level, velocity, and flow rate. Section 4 contains two examples of programming for modules. Module programming is the same in Extended as it was in Standard programming.

## 5.5 Programming Examples

In addition to the features available with standard programming which are discussed in Section 4, extended programming offers:

- Nonuniform time pacing
- Random interval pacing
- Event pacing
- Multiple bottle compositing
- Time switched bottles or bottle sets
- Flow proportional sample volumes
- Programmable sampler enable
- Pauses and resumes
- Two-part programming
- Automatic or user-defined suction head
- Suction line rinses
- Sampling retries

This manual provides several programming examples. Some, demonstrating simple tasks, appear in the margins. Others, demonstrating more complex tasks, appear as examples. The example “*Storm Water Runoff Sampling Program*” steps you through a storm water runoff sampling using extended programming.

### 5.5.1 Storm Water Runoff Sampling

Storm water runoff sampling is an ideal application for two-part programming. Typical monitoring permits require that a series of timed samples be taken during the initial portion of the storm event, followed by flow paced samples during the remainder of the event.

This example assumes that a Rain Gauge and a Bubbler Flow Module are connected to the sampler.

 **Note**

To enable the Bubbler Module programming screens, insert the module before turning on the sampler.

The example uses a 14-bottle kit and divides the bottles into two groups. Six bottles assigned to part 'A' and 8 to part 'B.' The sampler will begin taking samples after being enabled according to programmable enable controls; that is, when the rain gauge detects 0.15 inches of rainfall in 30 minutes, and the module detects a flow stream level of more than 0.25 feet. Because both program parts use the same enable settings, the sampler will enable both parts at the same time. A list of settings for the program follow:

- Program type: Extended
- Program name: STORM
- Site description: SITE 54
- Units:
  - Length: ft
  - Flow Rate: gpm
  - Flow Volume: gal
- Mode of Operation: Flowmeter, Flow-Insert, Round, 12"
- Bottle kit: 14, 950-milliliter bottles
- Suction-line length: 5 feet
- Suction head: Auto suction head
- Line rinses: None
- Sampling retries: None
- Part 'A' and 'B' bottle assignments:
  - Part 'A': Bottles 1–6
  - Part 'B': Bottles 7–14
- Part 'A'
  - Pacing: Time pacing, sampling every 5 minutes
  - Distribution: Sequential, 1 sample per bottle
  - Sample volume: 800 milliliters
  - Enable: At least 0.15 inches of rainfall in 30 minutes and a flow stream level of more than 0.25 feet; once enabled, stay enabled, sample taken when the sampler is enabled.

- Pause and resumes: None
- Part 'B'
  - Pacing: Flow pacing, sampling every 2,500 gallons
  - Distribution: Sequential, 1 sample per bottle
  - Sample volume: 500 milliliters
  - Enable: At least 0.15 inches of rainfall in 30 minutes and a flow stream level of more than 0.25 feet; once enabled, stay enabled, no sample taken when the sampler is enabled; first sample taken 20 minutes after the sampler is enabled.
  - Pause and resumes: None
  - Start time: Delayed start of 1 minute.

The example “Storm Water Runoff Sampling Program” presents several screens that are not available in standard programming. Descriptions of features available in extended programming follow the example.

### Example: Storm Water Runoff Sampling Program

1. This screen will appear briefly. Note that controller is in the EXTENDED PROGRAMMING mode.

BLZZRD  
 EXTENDED PROGRAMMING  
 For HELP at any  
 screen, press ? key.

2. Select PROGRAM from the main menu.

RUN “EXTENDED 1”  
**PROGRAM**  
 VIEW REPORT  
 OTHER FUNCTIONS

3. Select YES. *Note:* If the quick view screen shows up, press STOP then select OTHER FUNCTIONS>SOFTWARE OPTIONS>QUICKVIEW/CHANGE>NORMAL. Then press STOP and go back to step 2.

PROGRAM NAME:  
 “EXTENDED 1”  
 CHANGE?  
**YES** NO

4. Select CHANGE PROGRAM NAME.

SELECT NEW PROGRAM  
**CHANGE PROGRAM NAME**

5. Enter the word “STORM” for the program name. For instructions on entering text you can refer to the example *Entering a Site Description* in Section 3.

NAME: “STORM”  
 ABCDEFGHIJKLMNOPQRST  
 UVWXYZ-& 0123456789  
 BACK-UP DONE

6. Select YES.

SITE: DESCRIPTION  
 “FACTORY051”  
 CHANGE?  
**YES** NO

*Continued...*

Selecting a Stored Extended Program, continued...

7. Enter "SITE 54" for the site description.

```
SITE: "SITE 54"
ABCDEFGHIJKLMNQRST
UVWXYZ-& 0123456789
BACK-UP  DONE
```

8. Select ft.

```
SELECT UNITS FOR
LENGTH:
ft      m
```

9. Select °F.

```
SELECT UNITS FOR
LENGTH:
°F     °C
```

10. Select gpm.

```
SELECT UNITS FOR
FLOW RATE:
cfs  gps  gpm  Mgd
lps  m3s  m3h  m3d
```

11. Select gal.

```
SELECT UNITS FOR
FLOW VOLUME:
cf  gal  Mgal
m3  lit
```

12. Select YES.

```
PROGRAM MODULE?
YES             NO
```

13. Select FLOWMETER.

```
MODE OF OPERATION
FLOWMETER
LEVEL ONLY
```

14. Select FLOW-INSERT.

```
FLOW CONVERSION
WEIR  FLUME
DATA POINTS
MANNING  FLOW-INSERT
```

15. Select ROUND.

```
WEIR/ORIFICE TYPE
V-NOTCH
ROUND
```

16. Select 12.

```
FLOW INSERT SIZE:
6"  8"  10"  12"
```

17. Select DONE. *Note:* Partition memory status message will be displayed briefly.

```
NEW MODULE SETUP--
DOWNLOAD DATA NOW
OR LOSE ALL DATA!
DONE
```

18. Press Enter.

```
CURRENT LEVEL IS
___ . ___ ft.
ADJUST LEVEL TO
___ . ___ ft.
```

19. Select 5. *Note:* If the data storage interval is different from the current program, you will be informed that the storage interval has changed, and that the stored data will be lost. Download stored data, if desired and select DONE.

```
DATA STORAGE
INTERVAL IN MINUTES
1  2  5
10 15 30
```

20. Select 14.

```
NUMBER OF BOTTLES:
1  4  14
```

21. Enter 950.

```
BOTTLE VOLUME IS
950 ml (300-30000)
```

Continued...

Selecting a Stored Extended Program, continued...

22. Enter 5. The sampler calculates new pump tables.

SUCTION LINE LENGTH  
 IS **5** ft  
 (3-99)

23. Select AUTO SUCTION HEAD.

**AUTO SUCTION HEAD**  
 ENTER HEAD

24. Enter 0.

**0** RINSE CYCLES  
 (0-3)

25. Enter 0.

RETRY UP TO **0** TIMES  
 WHEN SAMPLING  
 (0-3)

26. Select TWO-PART PROGRAM.

ONE-PART PROGRAM  
**TWO-PART PROGRAM**

27. Enter 6. The screen will say "Beginning Part A" before proceeding with step 28.

14 BOTTLES AVAILABLE  
 ASSIGN BOTTLES  
 1 THRU **6** TO  
 PART 'A' (1-13)

28. Select UNIFORM TIME PACED.

**UNIFORM TIME PACED**  
 FLOW PACED  
 EVENT PACED  
 NONUNIFORM TIME

29. Enter 0 for hours, 5 for minutes.

TIME BETWEEN  
 SAMPLE EVENTS:  
**0** HOURS, **5** MINUTES

30. Enter 1.

**1** BOTTLES PER  
 SAMPLE EVENT (1- 6)

31. Select NUMBER OF SAMPLES.

SWITCH BOTTLES ON:  
**NUMBER OF SAMPLES**  
 TIME

32. Enter 1.

SWITCH BOTTLES EVERY  
**1** SAMPLES (1-47)

33. Select NO.

RUN CONTINUOUSLY?  
 YES **NO**

34. Select NO.

DO YOU WANT  
 SAMPLE VOLUMES  
 DEPENDENT ON FLOW?  
 YES **NO**

35. Enter 800.

SAMPLE VOLUME  
**800** ml (10-950)

36. Select RAIN. *Note: All valid enable options will appear here, so your screen may differ.*

ENABLE:  
**RAIN** LEVEL FLOW  
 FR-TEMP NONE

37. Select AND.

ENABLE: RAIN  
**AND** OR DONE

Continued...

Selecting a Stored Extended Program, continued...

38. Select LEVEL.

ENABLE: RAIN AND  
**LEVEL** FLOW FR-TEMP  
DONE

39. Enter 0.15 for inches, then enter 30 minutes.

"RAIN" SET POINT:  
**0.15** INCHES PER  
\_\_HOURS, \_\_MINUTES

40. Select ABOVE SET POINT.

ENABLED WHEN:  
**ABOVE SET POINT**  
BELOW SET POINT

41. Select YES.

RESET RAIN HISTORY?  
**YES** NO

42. Select SET POINT.

"LEVEL" CONDITION:  
**SET POINT**  
RANGE  
RATE OF CHANGE

43. Enter 0.25.

"LEVEL" SET POINT  
**0.25** ft  
(0.001-30.000)

44. Select ABOVE SET POINT.

ENABLED WHEN:  
**ABOVE SET POINT**  
BELOW SET POINT

45. Select YES.

ONCE ENABLED,  
STAY ENABLED?  
**YES** NO

46. Select YES.

SAMPLE AT ENABLE?  
**YES** NO

47. Select DONE.

PAUSE RESUME  
1. HH:MM DD HH:MM DD  
2. HH:MM DD HH:MM DD  
CLEAR **DONE**

The screen will display "End of Part A" then "Beginning Part B."

48. Select FLOW PACED.

UNIFORM TIME PACED  
**FLOW PACED**  
EVENT PACED  
NONUNIFORM TIME

49. Select FLOW MODULE VOLUME.

PACED BY:  
FLOW PULSES  
**FLOW MODULE VOLUME**

50. Enter 2500.

FLOW BETWEEN  
SAMPLE EVENTS:  
**2500** gal  
(0.001-99999)

51. Select YES. This will cause the first sample to be taken 20 minutes (the delay to start of sampling specified in Step 68) after the enable for Part B.

SAMPLE AT START?  
**YES** NO

52. Enter 1.

**1** BOTTLES PER  
SAMPLE EVENT (1 - 8)

Continued...

Selecting a Stored Extended Program, continued...

53. Select NUMBER OF SAMPLES.

SWITCH BOTTLES ON:  
**NUMBER OF SAMPLES**  
TIME

54. Enter 1.

SWITCH BOTTLES EVERY  
**1** SAMPLES (1- 47)

55. Select NO.

RUN CONTINUOUSLY?  
YES **NO**

56. Enter 500.

SAMPLE VOLUME:  
**500** ml (10-950)

57. Select RAIN.

ENABLE:  
**RAIN** LEVEL FLOW  
FR-TEMP 'A' DONE  
NONE

58. Select AND.

ENABLE: RAIN  
**AND** OR DONE

59. Select LEVEL.

ENABLE: RAIN AND  
**LEVEL** FLOW 'A' DONE  
DONE

60. Enter 0.15 for inches, then enter 30 minutes.

"RAIN" SET POINT:  
**0.15** INCHES PER  
\_\_HOURS, \_\_MINUTES

61. Select ABOVE SET POINT.

ENABLED WHEN:  
**ABOVE SET POINT**  
BELOW SET POINT

62. Select YES.

RESET RAIN HISTORY?  
**YES** NO

63. Select SET POINT.

"LEVEL" CONDITION:  
**SET POINT**  
RANGE  
RATE OF CHANGE

64. Enter 0.25.

"LEVEL" SET POINT  
**0.25** ft  
(0.100-30.000)

65. Select ABOVE SET POINT.

ENABLED WHEN:  
**ABOVE SET POINT**  
BELOW SET POINT

66. Select YES.

ONCE ENABLED,  
STAY ENABLED?  
**YES** NO

67. Select NO.

SAMPLE AT ENABLE?  
YES **NO**

68. Enter 20.

START OF SAMPLING  
AFTER A  
**20** MINUTE DELAY  
(1-999)

Continued...



*Selecting a Stored Extended Program, continued...*

69. Select DONE.

PAUSE	RESUME
1. HH:MM DD	HH:MM DD
2. HH:MM DD	HH:MM DD
CLEAR	<b>DONE</b>

The screen will display "End of Part B."

70. Select DELAYED START.

NO DELAY TO START  
**DELAYED START**  
CLOCK TIME

71. Select NO. (Option for samplers using a 0.01 inch tip rain gauge. See section 5.11.1, *Sampler Enable Responses, "Dry Period" Option*).

START "STORM"  
AFTER A 72 HOUR  
"DRY" PERIOD?  
YES **NO**

72. Enter 1.

START "STORM"  
AFTER A  
**1** MINUTE DELAY  
(1-999)

73. Enter 0.

MAXIMUM RUN TIME:  
**0** HOURS

74. Select NO.

PROGRAMMING COMPLETE  
RUN THIS PROGRAM  
NOW?  
YES **NO**

## 5.6 Suction Head

Suction head is the vertical distance between the surface of the liquid source and the sampler's pump. At each sample event, the sampler determines the suction head automatically.

In extended programming, you can manually enter the suction head if you want. Select ENTER HEAD only when the head remains stable and you can measure it accurately. You will then be prompted to enter the suction head height.

AUTO SUCTION HEAD  
**ENTER HEAD**

SUCTION HEAD OF  
\_\_\_\_\_ ft  
(min - max)

## 5.7 Rinses and Retries

Rinses and retries are extended programming features. You can program the sampler to rinse the suction line automatically. During a line rinse cycle, the sampler draws liquid through the line to the liquid detector. As soon as it detects liquid, the sampler reverses the pump to purge the line. You can program the controller to perform as many as three rinse cycles before each sampling cycle.

The sampling retries feature lets you set the number of times, from 0 to 3, that the sampler tries to detect liquid in the line before skipping the sample.

__ RINSE CYCLES (0-3)
RETRY UP TO __ TIMES WHEN SAMPLING (0-3)

## 5.8 Pacing

Sample pacing is the rate at which the sampler takes samples. Depending on the type of pacing you select, the rate is controlled by the sampler's internal clock or by inputs received from connected instruments.

Standard programming provides time pacing and flow pacing, which are described in Section 4. Extended programming provides additional pacing types, so its pacing screen has four items listed instead of two.

“Uniform time” pacing is identical to “time” pacing in standard programming with one exception. Uniform time pacing adds the ability to collect flow-proportional sample volumes. “Flow paced” is the same for both types of programming. Event paced and non-uniform time pacing are described below.

### 5.8.1 Event Pacing

Up to fourteen event conditions can be defined (for two-part programs, up to fourteen event conditions can be defined for each part). An event occurs when any of the programmed event conditions becomes true. By default, event paced sampling uses a sequential distribution scheme, meaning that each time an event occurs the sampler takes one sample and places it in one bottle. However, any distribution setting can be selected with event paced sampling.

UNIFORM TIME PACE FLOW PACED <b>EVENT PACED</b> NONUNIFORM TIME
--

To program the sampler for event pacing, select event paced. Refer to the example “*Event Paced Sampling*” for a program that uses three event conditions. This example uses a sampler with an attached 730 Bubbler Flow Module, Rain Gauge, and an SDI-12 sonde at address 5 with four parameters.

If you have programmed multiple conditions, and more than one condition comes true at the same time (within your measurement interval), one sample will be taken and all conditions that caused the event will be noted on reports.

The “*Event Paced Sampling*” example uses *rate of change* for an event condition (see step 7). This feature enables the event condition under specific circumstances.

## Example: Event Paced Sampling

1. Select EVENT PACED.

```
UNIFORM TIME PACED
FLOW PACED
EVENT PACED
NONUNIFORM TIME
```

2. Use the arrow keys to select the parameters that will trigger the event. To define the first event for this example, select LEVEL.

```
EVENT01:
RAIN LEVEL FLOW
FR-TEMP pH5 SP_CO5
DO5
```

3. Choose SET POINT.

```
"LEVEL" CONDITION:
SET POINT
RANGE
RATE OF CHANGE
```

4. Enter a set point of 1 foot.

```
"LEVEL" SET POINT:
1.000 ft
( 0.001 - 30.00)
```

5. Select ABOVE SET POINT.

```
ENABLED WHEN:
ABOVE SET POINT
BELOW SET POINT
```

6. Select LEVEL again.

```
EVENT02:
RAIN LEVEL FLOW
FR-TEMP pH5 SP_CO5
DO5 DONE
```

7. Choose RATE OF CHANGE.

```
"LEVEL" CONDITION:
SET POINT
RANGE
RATE OF CHANGE
```

8. Choose RISES.

```
CONDITION IS TRUE
WHEN "LEVEL"
RISES FALLS
```

Selecting RISES will enable the event condition if a rise of the amount and duration specified in step 9 occurs.

9. Enter 1 ft and press ↵. Then enter 2 hours.

```
"LEVEL" RISES
1.000 ft
__HOURS, __MINUTES
```

10. Select pH.

```
EVENT03:
RAIN LEVEL FLOW
FR-TEMP pH5 SP_CO5
DO5 DONE
```

11. Select RANGE.

```
"pH5" CONDITION:
SET POINT
RANGE
RATE OF CHANGE
```

12. Enter 4 for the lower range and 10 for the upper range.

```
"pH5" RANGE:
LOWER: 4.00
UPPER: 10.00
( 0.00 - 14.00)
```

13. Choose INSIDE RANGE.

```
ENABLED WHEN:
INSIDE RANGE
OUTSIDE RANGE
```

14. To stop entering events select DONE.

```
EVENT04:
RAIN LEVEL FLOW
FR-TEMP pH5 SP_CO5
DO5 DONE
```

Proceed with the rest of the programming steps.

```
↑ PACING:
EVENT PACED
3 EVENT CONDITIONS ↓
```

In the *Event Paced Sampling* example, a rate of change is set so that if the level rises more than one foot in a two hour time period, the event condition would be enabled and a sample taken.

In determining if the rise occurs in this example, the sampler looks back at the previous two hour time period to see if there is a rise of more than one foot between any valley (low point) or peak (high point) within that time period. Once the condition becomes true (a rise of more than one foot is noted), the event condition is enabled. It will not be enabled again unless the condition becomes false, and then true again, within the two hour time window.

In event pacing, up to three different rate of change event conditions can be programmed (for two-part programs, a total of six rate of change event conditions can be defined, three for each part).

#### 5.8.2 Nonuniform Time Pacing

The extended time pacing options are uniform time and non-uniform time. “Uniform time” pacing is identical to “time” pacing in standard programming. Nonuniform time pacing uses an irregular interval between sample events; each interval is individually programmable. There are three types of nonuniform time pacing:

- Nonuniform clock time pacing.
- Nonuniform interval pacing.
- Random interval pacing.

#### 5.8.3 Nonuniform Clock Time Pacing

For clock time pacing, enter specific times for each sample event (see screen 2 in the example). You can enter as many as 99 clock times, but the interval between times cannot be greater than 24 hours. (Remember to use a 24-hour clock when entering times.) The sampler takes a sample at the start time.

#### 5.8.4 Nonuniform Interval Pacing

For nonuniform interval pacing, enter the number of sample events spaced at intervals of minutes: 12 samples at 5 minute intervals, 6 samples at 10 minute intervals, and so on. The sampler accepts as many as 99 nonuniform interval entries. The sampler takes a sample at start time.

#### 5.8.5 Random Interval Pacing

To program the sampler for random interval pacing, enter the length of time you want to sample. The run time is the only random programming setting you’ll need to enter because the sampler derives the number of sample events required for the program from the distribution settings. (You can combine random pacing with any distribution.) Each time you run the program, the sampler generates a new set of random intervals. This makes each sample event unpredictable from run to run.

In both nonuniform clock time pacing and nonuniform interval pacing, the sampler takes a sample at the start time. For random pacing, however, it takes the first sample at the end of the first interval, not at the start time.

## Example: Nonuniform Time Pacing

**Nonuniform Clock Time Pacing** allows you specify the exact time at which each sample will be taken.

1. Select CLOCK TIMES.

```
NONUNIFORM TIME:
CLOCK TIMES
INTERVALS IN MINUTES
RANDOM INTERVALS
```

2. Enter as many as 99 clock times using a 24-hour format. *Note:* Intervals between samples cannot exceed 24 hours.

```
TAKE SAMPLES AT:
1. START TIME
2. HH:MM
3. HH:MM
```

**Random Time Pacing** collects samples at random time intervals.

1. Select RANDOM INTERVALS.

```
NONUNIFORM TIME:
CLOCK TIMES
INTERVALS IN MINUTES
RANDOM INTERVALS
```

2. Enter the program run time. The controller derives the number of sample events required for the program from the distribution settings.

```
PROGRAM RUN TIME:
__ HOURS, __ MINUTES
```

**Nonuniform Time Interval Pacing** allows you specify the time interval between each sample.

1. Select INTERVALS IN MINUTES.

```
NONUNIFORM TIME:
CLOCK TIMES
INTERVALS IN MINUTES
RANDOM INTERVALS
```

2. The controller briefly displays...

```
FIRST SAMPLE AT
START TIME,
THEN . . .
```

...then displays the screen to enter as many as 99 nonuniform time intervals. Enter the number of samples to be taken at each interval, then specify the interval in minutes.

```
QUANTITY AT INTERVAL
1. __ AT __ MIN
2. __ AT __ MIN
3. __ AT __ MIN
```

## 5.9 Distribution

Distribution describes how the sampler is to deposit samples. A sample is the volume of liquid deposited in a bottle. A sample event includes the full sampling cycle and may deposit a sample into more than one bottle. You can program the sampler for several distribution methods:

- Sequential
- Bottles per sample
- Samples per bottle
- Composite
- Multiple Bottle Compositing
- Time Switching

### 5.9.1 Sequential

In sequential distribution, the sampler deposits one sample in each bottle. A sequential sample represents a “snapshot” of the flow stream at a point in time.

### 5.9.2 Bottles Per Sample

In bottles per sample distribution, the sampler deposits a sample in each of a set of bottles. A bottle set includes at least two bottles but may include all bottles. Use bottles per sample when the volume to be collected is larger than the amount one bottle can hold or when you need identical samples.

### 5.9.3 Samples Per Bottle

In samples per bottle distribution, the sampler deposits samples from several sample events in a single bottle before moving to the next bottle. Use samples per bottle distribution to collect a series of small composite samples.

### 5.9.4 Composite

For single bottle configurations, samples per bottle distribution is known as composite sampling. A composite sample represents an average of the flow stream’s characteristics during the sampling period.

### 5.9.5 Multiple Bottle Compositing

Multiple bottle compositing is a combination of bottles-per-sample and samples-per-bottle. At each sample event, the sampler deposits a sample into a set of bottles. It moves to a new bottle set only after each bottle of the first set contains the programmed number of samples. Use multiple-bottle compositing to collect identical sets of composite samples or a composite sample that is larger than the capacity of a single bottle. Multiple bottle compositing is available only in extended programming.

### 5.9.6 Time Switching

In extended programming, the sampler offers an additional distribution feature, time switching. This feature determines when the sampler fills the next bottle or bottle set, and may be used with time-paced or flow-paced sample programs. Time switching distribution occurs at regular intervals (see the *Time Switching Distribution* example, step 3). The switch times are relative to the programmed “FIRST SWITCH TIME” (step 4).

**Example: Time Switching Distribution**

---

1. Enter the number of bottles that should receive samples at each sample event.

2 BOTTLES PER  
SAMPLE EVENT (1-max)

2. Select TIME.

SWITCH BOTTLES ON:  
NUMBER OF SAMPLES  
TIME

3. Enter the time between bottle switches.

SWITCH BOTTLES EVERY  
\_\_ HOURS, \_\_ MINUTES

4. Enter the first bottle switch time.

FIRST SWITCH TIME  
AT: \_\_:\_\_

Switch times occur regardless of the state of the sampler. Be aware that the sampler may leave empty bottles if it becomes disabled. Pause and resumes are an exception; the sampler will not switch bottles during a pause if the current bottle is empty.

If the program settings instruct the sampler to take a sample at the start time, the sampler also draws a sample at the switch time. When this occurs, the pacing will also be reset at the switch time.

Time switching is normally used with flow-paced programs that do not take a sample at start. These programs will place a sample in the current bottle if it is empty at the switch time, and then move to the next bottle/set. Pacing intervals are reset at every switch time.

When you use the time switching feature, the volume in each bottle or bottle set may vary. Sample volumes that would exceed the bottle capacity are not taken and a “PROBABLE OVERFILL” message is logged.

### 5.9.7 Continuous Sampling

Sample programs can be run indefinitely by selecting “YES” at the “RUN CONTINUOUSLY?” screen. Continuous sampling resets the distribution when the distribution sequence is complete. That is, when the last bottle of the set is reached, the next bottle is the first bottle of the set. All pacing modes except RANDOM are supported.


When running a program continuously, the bottles must be serviced at regular intervals to prevent overfilling the bottles. The sampler assumes that the next bottle/set is empty and ready to receive samples. If you are using this feature with two-part programming, keep in mind that the parts will most likely reset at different intervals.

**Note**


Sampling and Combined Results reports are cleared when the distribution sequence resets. If reports are needed, they must be collected before sampler advances to the next bottle set.

*BLZZRD™ Portable Refrigerated Sampler*  
 Section 5 *Extended Programming*


**Sequential**



**Bottles-per-Sample**



**Samples-per-Bottle**



**Multiple Bottle Compositing**



**Time Switched (Flow-paced, No sample at start)**




Figure 5-1 Sample Distribution



## 5.10 Flow Proportional Sample Volumes

The BLZZRD can collect flow proportional sample volumes. For some sampling protocols, these flow-weighted, variable-volume samples can be more representative of the stream. This is because stream characteristics often change as the flow fluctuates.

Variable-volume sampling is possible when the sampler is programmed for Uniform Time pacing. To enable this feature, select YES at the display asking “DO YOU WANT SAMPLE VOLUMES DEPENDENT ON FLOW?”

```
DO YOU WANT
SAMPLE VOLUMES
DEPENDENT ON FLOW?
YES      NO
```

After answering YES, the sampler displays screens to program the flow proportional volumes. If the sampler has a 700 Series Flow Module inserted, you will be asked which flow measurement should be used to determine the volumes.

```
FLOW PULSES
FLOW MODULE VOLUME
```

To base the volumes on an external flow measurement device attached to the Flow Meter connector, select FLOW PULSES. To base the volumes on the 700 Series Flow Module readings, select FLOW MODULE VOLUME.

Next, the sampler asks for the SAMPLE VOLUME to collect at each time interval. This volume is expressed as 10 ml for every X number of pulses, or if using a flow module, for every volume unit.

```
SAMPLE VOLUME:
10 ml FOR EVERY
___ PULSES
```

Refer to section 5.10.1, *Calculating Variable Sample Volume Settings*, for a more detailed discussion on this setting.

The final screen used to program variable-volume sampling is the MINIMUM SAMPLE VOLUME. If the number of flow pulses or flow module volume is not sufficient enough to collect this minimum volume, the sample is skipped and the flow count carries over to the next sample time. This setting ranges from 10 ml to a maximum determined by the bottle volume.

```
MINIMUM
SAMPLE VOLUME
ml
(10 - maximum)
```

### 5.10.1 Calculating Variable Sample Volume Settings

The entry for the SAMPLE VOLUME display can be calculated based on expected flow rates and collection volume requirements. The number of pulses or flow module volume that you enter should ensure that the minimum volume of liquid required for analysis is collected without exceeding the bottle volume capacity of the sampler. You may find it helpful to determine a range of values by calculating both possibilities. The number you enter would be selected from this range. This range is illustrated in Figure 5-2.




Figure 5-2 Factors affecting flow-weighted variable-volume samples

In other words, at the greatest expected flow, this SAMPLE VOLUME value should be set to prevent filling the bottle before the end of the program. An early full-bottle condition would result in a composite sample that is not representative of the flow stream.

At the lowest expected flow, the SAMPLE VOLUME value should be set to ensure that enough sample is collected. If too little sample liquid is collected, there may not be enough for analysis although it is a representative composite sample.

For example, consider an NPDES sampling program used to collect a flow-proportional composite sample for BOD, TSS, and NH<sub>3</sub>-N analysis. The laboratory requires at least 1,000 ml for BOD, 100 ml for TSS, and 400 ml for NH<sub>3</sub>-N tests. The minimum composite volume needed for analysis is then 1,500 ml, but to allow for duplicates and spillage, this volume is doubled to 3,000 ml. The samples are to be collected over an eight hour period at 15 minute intervals into a 9,400 ml composite bottle. Flow rates at the site range from 1.5 to 2.0 mgd, the equivalent volume of 500,000 and 666,667 gallons respectively over the eight hour period. This flow is measured by an external flow meter that sends a flow pulse every 100 gallons. Figure 5-3 depicts this scenario.




Figure 5-3 Variable-volume scenario

**5.10.2 Calculations when Using an External Flow Meter**

SAMPLE VOLUME:  
10 ml FOR EVERY  
\_\_\_ PULSES

The highest possible SAMPLE VOLUME value is determined by the minimum volume required for analysis and the least expected flow rate over the sampling duration. To calculate the number of pulses at this high end of the range, you should know:

- the least expected flow volume that will pass the flow measurement point
- the flow volume interval between each flow pulse from the external flow meter
- the minimum composite sample volume required for analysis. Allow for duplicates and spillage.

These factors would be applied to the following equation.

$$10 \text{ ml for every } \_ \text{ pulses} = \frac{(\text{Least Total Flow Vol}) / (\text{Flow Pulse Vol})}{(\text{Minimum Required Vol}) / (10 \text{ ml per unit})}$$

From the scenario we can apply the factors for the equation.

$$10 \text{ ml for every } 16.7 \text{ pulses} = \frac{(500,000 \text{ gallons}) / (100 \text{ gallons per pulse})}{(3,000 \text{ ml} / 10 \text{ ml})}$$

Because the SAMPLE VOLUME entry must be in whole numbers, 16.7 is rounded down to 16. It is important to note that this number is inversely proportional to the amount of volume collected. Rounding up would result in less volume collected.

The lowest possible SAMPLE VOLUME value is dependent upon the highest expected flow and the available bottle volume. To calculate the number of pulses or volume at this low end, you should know:

- the greatest expected flow volume that will pass the flow measurement point
- the flow volume interval between each flow pulse from an external flow meter
- the available bottle volume for sample collection.

These factors would be applied to the following equation.

$$10 \text{ ml for every } \underline{\hspace{1cm}} \text{ pulses} = \frac{(\text{Greatest Total Flow Vol}) / (\text{Flow Pulse Vol})}{(\text{Total Bottle Vol}) / (10 \text{ ml per unit})}$$

From the scenario we can apply the factors to the equation.

$$10 \text{ ml for every } 7.0 \text{ pulses} = \frac{(666,667 \text{ gallons} / 100 \text{ gallons})}{(9,400 \text{ ml} / 10 \text{ ml})}$$

At high flow rates, a number below 7 would overflow the bottle. At low flow rates, a number greater than 16 would collect too little sample. Therefore, a value should be selected between 7 and 16 pulses.

**5.10.3 Calculations when Using a 700 Series Flow Module**

The same principles discussed in *Calculations when Using an External Flow Meter* apply. However, the equations differ slightly because the flow module directly measures the stream's flow volumes. The equation for the high end of the range is:

$$10 \text{ ml for every } \underline{\hspace{1cm}} \text{ units} = \frac{(\text{Least Total Flow Vol})}{(\text{Minimum Required Vol}) / (10 \text{ ml per unit})}$$

Using the same scenario the calculations would be:

$$10 \text{ ml for every } 1667 \text{ gallons} = \frac{500,000 \text{ gallons}}{(3,000 \text{ ml}) / (10 \text{ ml per gallon})}$$

SAMPLE VOLUME:  
 10 ml FOR EVERY  
 \_\_\_\_\_ Gallons

The equation for the low end of the range is:

$$10 \text{ ml for every } \underline{\hspace{1cm}} \text{ units} = \frac{(\text{Greatest Total Flow Vol})}{(\text{Total Bottle Vol}) / (10 \text{ ml per unit})}$$

The scenario would be applied as:

$$10 \text{ ml for every } 709 \text{ gallons} = \frac{666,667 \text{ gallons}}{(9400 \text{ ml}) / (10 \text{ ml per unit})}$$

Therefore, a value should be selected between 709 and 1667 gallons for the given scenario.

## 5.11 Sampler Enable

Certain external instruments can enable (start) or disable (stop) a sampler by sending a signal to pin F of the sampler's flow meter connector. Teledyne ISCO flow meters, flow loggers, and the Model 1640 Liquid Level Actuator have a programmable sampler-enable feature that lets them enable or disable the sampler. You can define the condition (such as level, flow rate, pH, temperature, percent, rainfall, and I/O) or combination of conditions that need to be true in order to enable the sampler.

The programming example "*Sampler Enable Programming*" show a sample enable condition that uses a combination of conditions. In this example the sampler has an attached 730 Bubbler Flow Module, Rain Gauge and an SDI-12 sonde at address 5 with four parameters. Refer to Figure A-12 in Appendix A for details on programming sampler enables.

When the sampler detects that it is disabled at start time, it suspends the program until it is enabled. Once enabled, the sampler collects a sample only when programmed to take a sample at the enable time. The diagram in Figure 5-4 shows the sampler's response when enabled or disabled at the start time.




Figure 5-4 Start Time Diagram for Sampler Enable Settings

### Example: Sampler Enable Programming

In this example, you will be enabling the sampler if it rains 0.5" in 2 hours and the level goes above 1 foot.

1. Select NONE PROGRAMMED to begin defining the sampler enable.

```

    ↑  ENABLE:
      NONE PROGRAMMED  ↓
    
```

2. Select RAIN.

```

    ENABLE:
    RAIN LEVEL FLOW
    FR_TEMP pH5 SP_CO5
    D05 NONE
    
```

3. Select AND.

```

    ENABLE:  RAIN
    AND OR DONE
    
```

4. Select LEVEL.

```

    ENABLE: RAIN AND
    LEVEL FLOW FR-TEMP
    pH5 SP_CO5
    D05 DONE
    
```

5. Enter 0.5 inches of rain and press ↵. Then enter 2 hours.

```

    "RAIN" SET POINT:
    .5 INCHES PER
    __HOURS, __MINUTES
    
```

6. Select ABOVE SET POINT.

```

    ENABLED WHEN:
    ABOVE SET POINT
    BELOW SET POINT
    
```

7. Select YES.

```

    RESET RAIN HISTORY?
    YES NO
    
```

8. Select SET POINT.

```

    "LEVEL" CONDITION
    SET POINT
    RANGE
    RATE OF CHANGE
    
```

9. Enter 1 foot.

```

    "LEVEL" SET POINT:
    1 ft
    ( 0.001 - 30.00)
    
```

10. Select ABOVE SET POINT.

```

    ENABLED WHEN:
    ABOVE SET POINT
    BELOW SET POINT
    
```

11. Your enable conditions will be summarized on the display. If correct, press ↵ to continue. If necessary, make changes to the enable conditions.

```

    ↑  ENABLE:
      RAIN >0.50" / 2:00
      AND
      LEVEL >1.000 ft  ↓
    
```

### 5.11.1 Sampler Enable Responses

In addition to programming enable conditions, extended programming lets you control the sampler's response to its enable state. You can:

- Stay enabled after the first enable.
- Set up a repeatable enable.
- Take a sample at enable or disable.
- Delay the start of sampling after the enable.
- Reset the sampling interval countdown each time the sampler is enabled.
- Control the sampling interval countdown while disabled.
- Enable Part B of a two-part program when Part A is done.

**Once Enabled Stay Enabled** – For certain monitoring programs, you may want the sampler to continue to sample, even though the conditions that enabled the sampler no longer exist. When you use the ONCE ENABLED, STAY ENABLED feature, after becoming enabled the sampler takes samples until it reaches the end of its program.

ONCE ENABLED, STAY ENABLED? YES NO
--

#### **Note**

The STAY ENABLED feature works only when the sampler becomes enabled following program start. If you need this feature with enable *prior* to program start, contact Teledyne ISCO for details.

**Repeatable Enable** – After enabling the sampler, the enabling instrument (or instruments) continues to monitor the conditions and disables the sampler when the conditions are no longer satisfied. If re-enabled, the sampler resumes the sampling program.

#### **Note**

Repeatable enable is called “sampler latching” in flow meters, flow loggers, and FLOWLINK. Latching a sampler produces the same results as selecting YES; the difference is that the external instrument controls the latching.

**Sample at Enable or Disable** – You have the option of directing the sampler to take a sample as soon as it receives the enable and/or disable signal. If you program the sampler to stay enabled and choose to take a sample at enable, the sample

interval will be reset at the enable time. Samples at disable do not affect the sample intervals. Standard programs do not take a sample at the enable time.

SAMPLE AT ENABLE?  
YES NO

**Delay To Start Of Sampling** – If you have selected ONCE ENABLED STAY ENABLED and have chosen to not sample at the enable, you have the opportunity to suspend the start of sampling. This choice is useful for storm programs that require a flow-paced sample taken after the initial grab sample.

START SAMPLING  
AFTER A  
\_\_\_ MINUTE DELAY  
(1-999)

**“Dry Period” Option** – For samplers used with a 0.01 inch tip rain gauge, you can cause the sampler to wait for a “dry” period of under 0.10 inches of rain over a 72-hour period before running the program. At RUN time, the sampler asks if you want to restart the 72-hour period. Select this choice if you want the sampler to disregard any “dry” period that took place prior to the program start. The sampler will wait for up to one year for the “dry” period to occur.

START “EXTENDED 1”  
AFTER A 72 HOUR  
“DRY” PERIOD?  
YES NO

**Resetting the Sample Interval at Enable** – You may want the pacing countdown to start over each time the sampler becomes enabled. To accomplish this, select “YES” when asked if you want to reset the sample interval at enable. The next sample event will occur at the end of a full interval. Furthermore, if the sampler is enabled several times during the program, it resets the countdown each time.

Use the reset-interval feature carefully. For example, avoid resetting the interval when you sample with uniform-time pacing and want to take samples at specific times.

The effects of resetting the interval can be seen in this example: A sampler has been programmed to sample every 15 minutes and to reset the interval when enabled. It receives an enable signal at 10:03 and takes a sample. After resetting the interval, the sampler takes subsequent samples at 10:18 (not 10:15), 10:33 (not 10:30), and so on.

RESET SAMPLE  
INTERVAL AT ENABLE?  
YES NO



```
CONTINUE COUNTDOWN  
WHILE DISABLED?  
YES NO
```

If you select “NO” when asked if you want to reset the sample interval at enable, you will be asked if you want to continue the countdown while disabled. You can have the sampler continue the countdown, repeating it as often as necessary while the sampler is disabled, or you may suspend the countdown while disabled, resuming the countdown as the sampler is enabled. The standard program allows the countdown to continue while disabled.

### 5.12 Pauses/Resumes

Pause and resume settings, available in extended programming, create an intermittent sampling schedule. A program with pauses and resumes begins sampling at its programmed start time, continuing until the first pause time and day of the week. It then suspends sampling until the first resume time, when it begins sampling again. If the start time falls within a pause period, the sampler will not begin sampling until the next resume time. Pause and resumes repeat weekly.

You can enter up to nine pairs of pause and resume times. The pause and resume times shown as examples below suspend the program between 5:00 PM Monday and 8:00 AM on Tuesday and between 5:00 PM Tuesday and 8:00 AM on Wednesday. If the start time is Monday at 8:00 AM, the sampler takes samples between 8:00 AM and 5:00 PM Monday; between 8:00 AM and 5:00 PM Tuesday; and, continues after 8:00 AM Wednesday. To remove pause and resume settings from a program, select CLEAR.

```
PAUSE RESUME  
1. 17:00 MO 08:00 TU  
2. 17:00 TU 08:00 WE  
CLEAR DONE
```

```
SAMPLE AT RESUME?  
YES NO
```

A sampler running a pause/resume program resets the pacing interval at each resume time. You can program the sampler to take a sample at resume times.

### 5.13 Running Programs

To run an extended program, select run from the main menu. You may also select YES at the RUN THIS PROGRAM NOW? screen at the end of the programming screens.

Extended programs scheduled to start at a clock time will not start until the programmed clock time on or after the first valid date.

Note that if you have recently calibrated a parameter probe, the sampler may delay the start and display “PLEASE WAIT” until valid readings are expected from the probe. The delay accounts for the time typically required to install the probe, and lasts for five minutes after you complete the calibration.

#### 5.13.1 Run Time Screens

Run time screens, which report the program’s status while running a program, are explained in Section 6.

#### 5.13.2 Interrupting a Running Program

Interrupting a running program is also explained in Section 6. The scrolling menu that appears during the manual paused state can have two extra options when using extended programming:

**Adjust Enable** – This option is available for extended sampling programs that have programmed enable conditions. Selecting this option allows you to change the set-points or ranges for the enable conditions. The sampler displays all of the programming screens that define the thresholds of the programmed enables. You cannot change the type of enables used or the way they are combined.

If rain is an enabling condition, you also have the option of resetting the rain history.

**Adjust Suction Head** – If you have chosen to manually enter the suction head, and are in the manual paused state, the scrolling menu will have an option to adjust the suction head. If you make changes to the suction head value, the sampler logs a changed suction head event.

### 5.14 Other Functions

In extended programming, the “Other Functions” menu screen includes options for:

- Maintenance (see Section 9)
- Manual Functions (see Section 6)
- Software Options (see Section 5.15)
- Hardware (see Section 5.16)

### 5.15 Software Options

The software option screens allow you to set up a number of different software options:

- Liquid Detector Enable
- Full Bottle Detection
- Measurement Interval
- Dual Sampler Mode
- Normal/Quick View Screens
- Display Backlighting
- Event Marks
- Pump Purge Counts
- Serial Output
- Interrogator Connect Power

#### 5.15.1 Liquid Detector Enable/Disable

Under some conditions (pressurized lines, sampling foam), a more repeatable sample volume may be delivered by disabling the liquid detector. Selecting NO will turn the Liquid Detector off.

USE LIQUID DETECTOR?  
YES NO

If the Liquid Detector is off:

- You must enter the SUCTION HEAD.
- No RINSES or RETRIES will be available.
- The COUNT TO LIQUID column of the RESULTS report will read “0.”
- A complete purge will occur between samples when using multiple bottles-per-sample distribution.

### 5.15.2 Programming Style

The sampler has two different styles of programming screens: normal and quick view. These are both explained in Section 3.

When you are in extended programming, the default style is Quick View, which allows you to move more quickly through the screens. If you prefer a step-by-step display for the screens, switch to the Normal setting.

### 5.15.3 Measurement Interval

MEASUREMENT INTERVAL  
5 15 30 SECONDS  
1 2 5 MINUTES  
STORAGE INTERVAL

If using a 700 Series module or SDI-12 Sonde parameter for enabling, the sampler will request a Measurement Interval.

The Measurement Interval allows the sampler to compare the parameter to the enabling conditions at a rate faster than the Data Storage interval. If the Measurement Interval is equal to or greater than the Data Storage interval, the parameter will be checked at the rate of Data Storage interval.

Normally, there are two considerations when selecting the Measurement Interval: power consumption and duration of the enabling event. This setting for battery powered samplers will require some degree of balance between the need to conserve power yet detect the enabling event before it has passed.

### 5.15.4 Dual Sampler Mode

DUAL SAMPLER MODE?  
YES NO

Dual Sampler Mode lets you operate two samplers as a pair. The primary sampler disables the secondary one until the primary sampler completes its sampling program. The samplers then reverse roles with the second sampler becoming the primary one.

#### Setting Up Dual Samplers –


1. Turn on the dual sampler mode in both samplers.
2. Install the samplers.
3. Connect the samplers with the Dual Sampler Mode Interconnect Cable by attaching the cable to the sampler’s flow-meter connectors. If your installation includes a flow meter or flow logger, use the Dual Sampler Mode to Flow Meter Cable instead.
4. Decide which sampler is to be the primary sampler; the primary sampler should be the sampler with the program you want to run first.
5. Turn the primary sampler on and run its sampling program by selecting RUN. After at least 5 seconds, turn the secondary sampler on and select RUN from its main menu.

**How Does It Work?** – The primary sampler sends a disable signal out its flow meter connector. The secondary sampler receiving the signal is held disabled until the primary sampler completes its program.

When using a flow meter or flow logger a Dual Sampler Mode to Flow Meter cable must be used. This cable is constructed so that the pacing instrument receives only the event marks. Therefore, a connected flow meter will not receive a bottle number signal, nor will it be able to control the sampler inhibit.

### 5.15.5 Bottle Full Detection

The sampler can detect a full bottle condition when collecting samples in a single bottle (composite sampling). When the sampler detects a full bottle, it stops the sampling routine. A BOTTLE FULL condition will also stop a sampling routine that is set to RUN CONTINUOUSLY.




The Bottle Full Detect screen allows you to enable or disable this feature. Selecting “YES” enables the sampler to detect a full bottle and stop the sampling routine. Selecting “NO” disables this feature. The Bottle Full Detect setting is part of the sampling program. Because it is a program setting, full bottle detection can be enabled or disabled for each of the four extended programs. If the Bottle Full Detect setting is critical to your sampling program, always check this Software setting after you select a different Extended Program.

This feature relies on a post-purge after each sample. When using the Bottle Full Detect, make sure that post-sample purges greater than 100 are used.

Bottle Full Detect is always disabled in the Standard Programming mode.

### 5.15.6 Display Backlighting

The display has a backlight that is adjustable so that the backlight is always off, timed or always on.




In the timed backlight mode the sampler turns the backlight off after detecting no key strokes within 60 seconds. Any keystroke (except ON-OFF) turns the backlighting on again. In this mode the backlight must be on before the sampler accepts a menu choice or number entry. If you press a key and a screen does not respond, it is because the keystroke turned the backlight on.

It is recommended that you use either BACKLIGHT ALWAYS OFF or TIMED BACKLIGHT when using a battery.


### 5.15.7 Event Marks

The sampler generates an event mark at each sample event, sending it through the flow meter connector to a connected instrument. The event mark may also control an external device; a connected solenoid relay, for example. When using the event mark for a data logger or to control external devices, you may need to adjust the mark by changing the factory settings.

The event mark is adjustable through the software setup screens. You can set up the sampler to generate an event mark for EVERY SAMPLE or for COMPLETE SAMPLES.



If you select COMPLETE SAMPLES, the sampler sends a three-second pulse only after it successfully collects a sample. The sampler will not send an event mark, such as “No Liquid Detected.”



This diagram shows the pulses sent for a bottles-per-sample distribution scheme

Figure 5-5 Event Mark Timing Diagram

If you select EVERY SAMPLE, the sampler will send an event mark every time a sample is initiated. The event mark signal can be a:

- 3-second pulse at the beginning of the pre-sample purge.
- 3-second pulse at the beginning of forward pumping only, marking the time the sampler deposited the sample.
- Variable duration pulse generated during the pump cycle, from the beginning of the pre-sample purge to the end of the postsample purge.
- Variable duration pulse generated only during forward pumping.

Figure 5-5 shows event mark timing diagrams.

**Bottle Number** – Concurrent with the event mark, the sampler also sends a bottle number signal. The bottle number signal is a series of pulses. The number of pulses sent represents the bottle currently accepting a sample (see Figure 5-5).

### 5.15.8 Pre-sample and Post-sample Purge Counts

The pre-sample purge is when the pump runs in reverse clearing the strainer of any debris before taking a sample. The pre-sample purge is set at 200 pump counts. After the sample is taken a post-sample purge occurs to clear the liquid from the suction line.

The duration of the post-sample purge is determined by the length of the suction line to maximize the battery life. The pre- and post-sample purge can also be manually set.

```
PRE-SAMPLE PURGE:
  _____
    COUNTS
  (10-9999)
```

```
POST-SAMPLE PURGE:
DEPENDENT ON HEAD
FIXED COUNT
```

```
POST-SAMPLE PURGE:
  _____
    COUNTS
  (10-9999)
```

### 5.15.9 Serial Output

Serial ASCII data can be output through the controller's interrogator port. The output frequency is user-selectable: every 15 seconds, every minute, every 5 minutes, every 15 minutes, or by command.

To receive serial data as a *periodic output*:

```
PERIODIC
SERIAL OUTPUT?
YES  NO
```

```
PORT SETTINGS:
9600N81  4800N81
2400N81  1200N81
```

```
SERIAL OUTPUT EVERY:
15 SEC  1 MIN
5 MIN   15 MIN
```

1. Select YES at screen asking if you want periodic serial output.
2. Select the appropriate baud rate. Note that at all baud rates the data is sent with no parity, eight data bits, and one stop bit. If you have difficulty with the connection, try a slower baud rate.
3. Select the desired output interval in screen.
4. Connect to the sampler controller's interrogator connector using a standard interrogator cable.

```
PERIODIC
SERIAL OUTPUT?
YES  NO
```

To receive serial data by *command*:

1. Select NO at the screen asking for periodic serial output.
2. Connect your computer to the sampler controller's interrogator connector with a standard interrogator cable, P/N 60-2544-040 (25 pin) or 60-2544-044 (9 pin).
3. Configure your communications software for a baud rate no higher than 19,200, no parity, 8 data bits, and 1 stop bit.
4. From the computer's keyboard, type "?" repeatedly until the sampler controller determines the baud rate. Once the baud rate is known, the sampler returns an identification banner. This banner indicates that the connection is working properly.
5. Type DATA<CR> each time the data string is desired.

The serial data is in comma-separated-value format. Only the values that the sampler is set up to measure will be output in the data string. The string is terminated with a <CR><LF>. To read

the data in the string, refer to Table 5-1. This serial data can also be accessed from a remote location using the optional dial-out modem (see *Section 8, Remote Operation*).

**☑ Note**

Connecting either the serial output or the standard interrogator cable to the sampler disables the optional internal dial-out modem, if installed. The interrogator cable must be disconnected before you can use the modem.

**Table 5-1 Serial Data Codes**

Identifier	Parameter	Units
B? <sup>1</sup>	Bottle number and time	days since 1-Jan-1900
CR	Current day's rain	tips <sup>2</sup>
CS <sup>3</sup>	Checksum	unsigned long
DE	Description	BLZZRD Sampler
FL	Flow	cubic meters per second
ID	Unit specific identifier	10 digit unsigned long
LE	Level	meters
MO	Model	BLZZRD
PE	Percent	percent of full scale
PH	pH	none
PR	Previous day's rain	tips <sup>2</sup>
RA	Rain	tips <sup>2</sup>
RTE	Refrigerator Temperature	degrees celsius
SS	Sampler enable state	logical
TE	Temperature	degrees celsius
TI	Current time	days since 1-Jan-1900
VE	Velocity	meters per second
VO	Volume	cubic meters
VSI	Velocity signal strength	percent
VSP	Velocity spectrum strength	percent
CO_ <sup>4</sup>	Conductivity	milliseimens per centimeter
DO_ <sup>4</sup>	Dissolved oxygen	milligrams per liter
PH_ <sup>4</sup>	pH	none
SA_ <sup>4</sup>	Salinity	parts per thousand
SP_ <sup>4</sup>	Specific Conductance	milliseimens per centimeter
TD_ <sup>4</sup>	Total Dissolved Solids	grams per liter
TE_ <sup>4</sup>	Temperature	degrees Celsius
OR_ <sup>4</sup>	Oxidation reduction potential	millivolts

**Table 5-1 Serial Data Codes (Continued)**

Identifier	Parameter	Units
LE_ <sup>4</sup>	Level	meters
MM_ <sup>4</sup>	Ammonium-nitrogen	milligrams nitrogen per liter
MA_ <sup>4</sup>	Ammonia-nitrogen	milligrams nitrogen per liter
NI_ <sup>4</sup>	Nitrate-Nitrogen	milligrams nitrogen per liter
TB_ <sup>4</sup>	Turbidity	nephelometric turbidity units
CL_ <sup>4</sup>	Chloride	milligrams per liter
CP_ <sup>4</sup>	Chlorophyll	micrograms per liter
FL_ <sup>4</sup>	Flow rate	cubic meters per second
VO_ <sup>4</sup>	Flow volume	cubic meters
xxD_ <sup>4, 5</sup>	Other data not listed above	

<sup>1</sup> The three most recent samples will have entries as B?. The “?” will be the first bottle number in the set. If the sample is a grab sample, it will be shown as B0.

<sup>2</sup> Rainfall tips roll over every 256 counts.

<sup>3</sup> The checksum does not include the checksum, carriage return, and linefeed. The string is terminated with <CR><LF>.

<sup>4</sup> A “\_” in the identifier indicates an SDI-12 address.

<sup>5</sup> “xx” matches the data label for the SDI-12 parameter.

**5.15.10 Interrogator  
 Connector Power**

POWER ALWAYS ON  
 POWER ON/OFF TIMES  
 ALARM DIALOUTS ONLY

Through the software options, the sampler can switch the 12 VDC power at pin A of the interrogator connector. This power-saving feature will power this port only when communications with the sampler are likely. This feature also may be used to provide timed control of a connected device.

**Power Always On** – Select this option to always provide power at the interrogator connector.

**Power On/Off Times** – Select this option to switch on power only during user-specified times of the day. These times would coincide with your data collection schedules. Up to four pairs of Power On/Off times may be entered.

Power ON/Off times are useful when powering a user-supplied external modem or cellular phone for data collection. If the controller is equipped with the optional internal dialout modem, pin A also will be powered if a programmed dialout condition is true.

Because this pin can be switched under program control, it is possible to use this output to open or close relay contacts which operate an external pump or other device.

**Alarm Dialouts Only** – Select this option to power this port only when a programmed dialout condition is true. This option is intended for applications that use the optional internal dialout modem with a cellular phone. To control the operation of other external devices based on alarm conditions, use *I/O Pin Programming* instead.



## 5.16 Hardware Setup

RUN "PROGRAM NAME"  
PROGRAM  
VIEW REPORT  
OTHER FUNCTIONS

- Select Other Functions

MAINTENANCE  
MANUAL FUNCTIONS  
PROGRAMMING STYLE  
HARDWARE

- Select Hardware

The hardware setup screens make adjustments to the sampler's hardware-based features:

- SDI-12 Sonde Setup
- Rain Gauge Setup
- I/O Pin Programming
- Dialout Alarms
- Pager Numbers
- Text Messaging
- Refrigerator Temperature
- Analog Outputs

### 5.16.1 SDI-12 Sonde Setup

The sampler supports SDI-12 sondes. For more information on setting up sondes, refer to Section 7.

### 5.16.2 Rain Gauge Setup

With standard programming, the sampler does not record rain gauge readings. However, with extended programming, recording the rain gauge readings and the rain enable option can be turned on or off. Enter the hardware setup and follow the screens to the left in the margin.

DO YOU HAVE A  
RAIN GAUGE CONNECTED?  
YES NO

- Select YES.

RAIN GAUGE:  
0.01 inch TIP  
0.1 mm TIP

- The rain gauges are set up to tip on either .01 inch or .1 mm. Select the desired tip.

When asked if you have a rain gauge connected, select YES to record the rain gauge readings and add RAIN as a sampler enable option. Select the appropriate tip for the rain gauge. The sampler will then create a rain memory partition and update the available sampler enable conditions.

Select NO to disable the readings and the sampler enable option. When you select NO, the sampler removes the rain memory partition, leaving the memory space available for other reading types.

### 5.16.3 I/O Pin Programming

Pins C, H, and I on the Rain Gauge Connector are user-configured as I/O1, I/O2, and I/O3 (Input/Output pins) respectively. Each pin can be either an input for sampler enabling, or an output for alarms.

 <b>Note</b>
---

The state of the I/O lines is not defined when the sampler is OFF.

#### *Sampler enable*

The enable input requires a pulse signal of 5 to 15 volts. Teledyne ISCO offers optional interface modules for converting from a 4-20 mA analog or pulse duration signal.

#### *Alarms*

The alarm output can be set for any RUNNING PROGRAM or RUN ERROR state. In addition, the CONDITIONS output can report alarm conditions using the parameters the 6712 is configured to measure. You can set thresholds for one or two conditions and the CONDITIONS output can then trigger alarms when the measured parameters exceed the threshold.

Alarm function requires an alarm interface module, factory-installed talking modem, or external digital cellular modem (see Section 5.16.4 *Dialout Alarms*).

Teledyne ISCO offers a module designed to provide one, two, or three contact closures. The alarm module can interface the sampler with many types of non-ISCO devices. For more information about the alarm module, contact your Teledyne ISCO sales representative.

## Example: I/O Pin Programming

Although the screens displayed may vary depending on your site setup, the programming sequence will be the same or similar to the following example.

1. Select OTHER FUNCTIONS to enable the dialout alarm.

```
RUN "EXTENDED 1"  
PROGRAM  
VIEW REPORT  
OTHER FUNCTIONS
```

2. Select hardware and use the down arrow to scroll to the I/O screen.

```
MAINTENANCE  
MANUAL FUNCTIONS  
SOFTWARE OPTIONS  
HARDWARE
```

```
↑ I/O1 = NONE  
I/O2 = NONE  
I/O3 = NONE ↓
```

3. Select the first pin to program.

```
SET I/O1:
```

From this screen, your choices are viewed by scrolling with the arrow buttons:

```
PGM RUNNING  
PGM ENABLED  
'A' ENABLED  
'B' ENABLED  
PROGRAM DONE  
'A' DONE  
'B' DONE  
RUN ERROR  
FLOW-THRU  
CONDITIONS  
NONE  
I/O ENABLE
```

Note: All choices listed are alarm output settings, except for I/O ENABLE. This is the input function to enable the sampler via an external device.

4. If you select CONDITIONS, the sampler will display:

```
I/O1:  
RAIN NONE
```

Depending on what equipment is connected to your sampler, you may see other parameters in addition to RAIN, such as LEVEL, FLOW, VELOCITY, etc. Under CONDITIONS, you can program limits for a condition to trigger the alarm when the measured parameter exceeds its limit. You can use one or two conditions to trigger one alarm, using AND or OR operators.

5. When you select a parameter with the arrow keys and Enter, you can then program the limits for alarm triggering. For example, if you select rain, the sampler will display:

```
"RAIN" SET POINT:  
1.00 INCHES PER  
1 HOURS, 0 MINUTES
```

```
I/O IS HIGH WHEN:  
ABOVE SET POINT  
BELOW SET POINT
```

"I/O IS HIGH" means "Alarm is triggered." Select ABOVE or BELOW, then select whether or not to reset rainfall history when the alarm is triggered:

```
RESET RAIN HISTORY?  
YES NO
```

Continue with programming I/O2 and I/O3 in the same manner, or if additional alarms are not needed, select NONE.

### 5.16.4 Dialout Alarms

Dialout alarms require the sampler to be equipped with a 6712 LTE modem. The dialout alarm allows the sampler to notify a contact list when an “alarm” condition exists (is true).

At the transition to the true state, any programmed I/O output can cause the sampler to warn of the alarm condition. As many as three telephone numbers can be entered. The sampler tries each telephone number in succession until the alarm is acknowledged. For each alarm condition, the sampler will attempt each telephone number entered a maximum of five times, at user specified intervals.

Alarms are set up through the I/O programming menu, shown in Section 5.16.3. “I/O1”, “I/O2”, and “I/O3” represent the three alarms.

### Example: Dialout Alarms

If the controller is equipped with a dialout modem, it can be programmed to notify a telephone or pager contact list when an alarm condition is true.

Before proceeding the I/O conditions must be defined. Refer to section 5.16.4. These steps program a dialout alarm when the condition defined for I/O1 is true.

1. Select YES to enable the dialout alarm.

```
DIALOUT WHEN IO1
BECOMES TRUE?
YES  NO
```

The controller asks the same question for IO2 and IO3. Select NO.

```
DIALOUT WHEN IO2
BECOMES TRUE?
YES  NO
```

```
DIALOUT WHEN IO3
BECOMES TRUE?
YES  NO
```

2. Using the number entry keys, enter as many as 3 telephone numbers to the contact list.

```
DIALOUT NUMBERS:
1. 9,5555550001
2. 9,5555550002
3. P5550003
```

If the phone number requires extended characters, or if you need to designate a number as a pager, press the decimal key. The controller displays:

```
SELECT CODE
, # * PAGER
```

Use an arrow key to select a code and press Enter to insert it in the number on the DIALOUT NUMBERS screen.

3. Enter the delay time between each attempt. A delay will allow time between attempts so you can return a call to the controller to acknowledge the alarm condition.

```
__ MINUTE
DELAY BETWEEN CALLS
(0-99)
```

*Note:* Numbers may be removed from the list by advancing the cursor to the very end of the number with the right arrow key, then pressing the left arrow key to erase the number.

### 5.16.5 Digital Modem Text Messaging

The sampler can send alarms in the form of text messages containing the same information as that found in voice messages.

To use this option, you must have:

- a cellular phone with text messaging capabilities
- an external digital modem  
recommended: Teledyne ISCO's 6712 LTE modem,  
part #60-9004-499.

#### **5.16.6 Refrigerator Temperature**

The sampler monitors the refrigerated compartment temperature with a sensor. The sensor transmits its temperature readings to the controller.

Temperature readings are updated approximately every 2.5 seconds. The readings are stored as FR-TEMP readings at the data storage interval. You can retrieve a history of temperature readings through the sampler's reports. The temperature data can also be used as a sampler enable condition. On the sampler, this hardware feature is always enabled to operate the refrigeration system.



# BLZZRD™

## Portable Refrigerated Sampler

---

### *Section 6 Running Programs*

This section provides information on:

- Starting a program
- Reading display screens while a program is running
- Pausing a program
- Performing manual operations
- Viewing and retrieving sampler data from a running program.

#### **6.1 Starting a Program**

To start a program, select RUN from the main menu. You may also select YES at the RUN THIS PROGRAM NOW? screen at the end of the programming screens.

#### **6.2 Run Time Screens**

While running a sampling program the sampler displays a variety of messages that report the program's status. For example, if the sampler has not yet reached the programmed start time, it displays the scheduled start time as well as the current time.

Once the program reaches the start time, you can determine the time of the next sample, the next bottle to receive a sample, sample distribution, and other information. Other messages appear while the sampler runs through a sampling cycle and takes a sample. When the sampler needs to report multiple messages, it alternates them, displaying each for one to three seconds.




Figure 6-1 Running a Program

### 6.2.1 Module Readings

Samplers with an attached module display the module's readings: pH and temperature for the pH module; level, percentage, or flow rate for the 4–20 mA module; level, flow rate, and flow volume readings for the flow modules; level, velocity, flow rate, and positive flow volume readings for the area velocity module. Some module screens will alternate with the sampler screens.

BOTTLE 2	
IN 00:14:39	
150 gpm	0.82 ft
00001438200 gal	

The module readings are updated every second when displaying the reading during programming. Otherwise, readings are updated every minute.

An \* (asterisk) appears next to the reading if the module was unable to take a reading. If an asterisk appears, the reading displayed is the last available reading. Flow rates will be reported as zero flow if the error persists for more than five minutes (for a level or velocity reading that flow is calculated from).



If an entire reading is filled with asterisks, the value exceeds the number of characters that can be displayed. The sampler stores the actual reading in memory, so it may be collected later.

```
BOTTLE 2
IN 00:12:00
***** gpm 5.73 ft
00001457600 gal
```

For those programs that have delayed or scheduled start times, the readings will be displayed while waiting for the start time. Keep in mind that any totalizer values will be reset at the start time. Module and sonde readings are also displayed after a program is complete, until a key is pressed.

It is not necessary to run a program to view module or sonde readings. Real time data can be viewed on the display by selecting “VIEW REPORT” from the main menu.

### 6.3 How Does the Sampler Work?

When the sampler takes a sample, it draws liquid through the strainer and suction line to the pump tube (refer to Figure 6-2). The liquid flows through the pump tube and past the liquid detector, which senses the liquid. From the detector, the liquid follows the pump tube through the pump to the bulkhead fitting and then through the discharge tube to the sample bottle.




Figure 6-2 Identifying sampler components

### 6.3.1 A Typical Sampling Cycle

1. The sampler moves the distributor arm over the bottle that is to receive the sample.
2. The pump reverses for the pre-sample purge.
3. The pump direction changes, filling the suction line.
4. When the detector senses liquid, the sampler begins measuring the sample.
5. After depositing the sample, the pump again reverses for the post-sample purge.

 **CAUTION**

Prolonged exposure to noise while the pump is operating could result in hearing loss. Tests indicate that the pump produces sound levels in excess of 85db at one meter. If you are manually operating the pump for long durations, always wear hearing protection.

Sampling cycles vary somewhat according to program settings for distribution. The sampler can move the distributor arm clockwise and counterclockwise, making a number of distribution methods possible. (In this manual, the words “sample event” refer to a full sampling cycle for any distribution.)

Other variations include extended-program settings for suction-line rinses and sampling retries. A sampler running a program with line rinses completes the pre-sample purge and then rinses the line. The suction line is purged after each rinse. When programmed for sampling retries, the sampler will attempt to pump a sample again if it fails on previous attempts.

### 6.3.2 Measuring Sample Volume

“Sample volume” refers to the amount of liquid delivered into a bottle. The volume is determined by the programmed value. The volume is dependent on the volume per revolution of the pump, which is dependent on the suction head. The BLZZRD delivers the sample by counting the pump revolutions and automatically compensating for the suction head.

## 6.4 How Does the Refrigerator Work?

The operation of the refrigeration system is under control of the sampler controller. As long as the controller is powered, the refrigeration system is active. This is true for all states (including OFF), except for the time between entering RUN and the completion of the first sample, and when the pump is running. To conserve power, the sampler assumes that during this time there is no sample liquid to cool.

The refrigeration system has two modes of operation. Both modes are under control of the sampler, which senses the air temperature inside the refrigerated compartment.

After the first sample until completion of the running program, the sampler cools the refrigerated compartment to 1°C, ±1. This low temperature and narrow range allows the refrigeration system to quickly cool liquid as it is deposited in the bottles.

One hour after the last sample of a program is taken, the sampler adjusts its control of the refrigerator compressor to maintain the samples at 3°C, ±1. At this time, the samples have been cooled and no more liquid at ambient temperature will be added. The 3°C target temperature maintains the samples within recognized standards while conserving power.

The measured temperature is reported as FR\_TEMP on the run state displays and in the stored data.

## 6.5 Interrupting a Running Program

You can interrupt a sampling program by pressing the Stop key while the sampler is waiting for the next sample event. Pressing Stop places the sampler into *Manual Paused* operation and records a manual pause in the sample event log.

In this state, the sampler allows you to access several functions. You can modify the program, reset alarms, take a grab sample, etc., without adversely affecting the running program. Modifying the program is limited to adjusting the pacing intervals, enable conditions and sample volume. The basic program structure, such as the type of pacing, can not be changed.

While in the manual paused state, the sampling program continues to operate as normal, with the exception of taking samples. If a sample was to be taken, it is skipped. The sampler records “sample skipped” in the sample event log and continues to operate as normal.

The manual paused state displays a scrolling menu with several options (see example below). Use the Arrow keys to scroll through the manual paused options and the ↵ key to make a selection. Return to the running program by selecting resume program or pressing the Stop key. If you do not make a selection, the sampler automatically resumes the original sampling program five minutes after the last key was pressed.

←	STOP PROGRAM	
	RESUME PROGRAM	
	VIEW DATA	
	GRAB SAMPLE	→

- 6.5.1 Stop Program** This option terminates the running program and records program stopped in the sample event log. The program cannot be resumed.
- 6.5.2 Resume Program** Selecting this option will cause the sampler to exit the manual paused state and return to normal program operation. A manual resume event will be logged.
- 6.5.3 View Data** This option allows you to view the data recorded by the sampler. See Section 6.6.1 for more information.
- 6.5.4 Grab Sample** When selected, you will be asked to enter the volume to be pumped. A sample will then be taken as if it is to be *placed into a container outside the sampler base*. Do this by removing the pump tubing from the bulkhead fitting. See Section 6.6.1. A grab sample event will be logged.
- 6.5.5 Pump Tube Alarm** If the pump tube is changed, this option should be selected to reset the pump counts on the counter. This will remove the “WARNING: REPLACE PUMP TUBING” message that may appear while the program is running. A pump tube replaced message is logged. See Section 9, *Maintenance*.
- 6.5.6 Calibrate Volume** When selected, you will be asked to enter the volume to be pumped. A sample will then be taken as if it is to be *placed into a container outside the sampler base*. Do this by removing the pump tubing from the bulkhead fitting.  
After the sample has been pumped, you will then be asked to enter the amount actually delivered. Once you have entered the delivered volume and accepted the calibration, a volume delivered event will be logged. See Section 6.6.2 for more information.
- 6.5.7 Cal/Adj Parameters** When the sampler is configured for operation with a module, this option becomes available. Selecting this option will display the appropriate level adjustment screens and/or the calibration screens. If the level is adjusted, an adjusted level event is logged. If a parameter is calibrated, an appropriate event is logged.  
Calibrating a parameter probe will temporarily “turn off” the partition data storage and the sample enable/disable functions. These functions are disabled during the calibration and for five minutes after the program is resumed. During this time, parameter data normally collected at the data storage interval will be logged as a “252” error message.
- 6.5.8 Adjust Pacing** This option is available when the running program is paced by Time, Flow Pulses or Flow Volume. If you select this option, you are asked to enter a new pacing interval. The sampler logs an interval changed event.  
The original pacing interval continues to count down until you return the sampler to the running program (select resume program). If the new pacing interval is less than the original count remaining, the new interval is used.

If the original count remaining is less than the new pacing interval, the original count will continue to count down to the next sample event. Subsequent samples are then paced by the new interval.

#### 6.5.9 Adjust Volume

Select this option to change the sample volume within the limits of the currently programmed sample distribution. If you change a sample volume, a volume changed event is logged.

#### 6.5.10 Power Used

The BLZZRD sampler has a “fuel gauge” that gives an indication of power usage. The controller keeps track of how much power has been consumed since the last time it lost power. If the controller is powered by an external battery, this fuel gauge can help you estimate the condition of the battery.

Keep in mind that the sampler has no idea as to how much charge was originally stored in the battery (or even if a battery is being used), and therefore has no idea as to the *remaining capacity* of its power source.

### 6.6 Manual Functions

The MANUAL FUNCTIONS programming screens let you:

- Take grab samples.
- Calibrate sample volumes.
- Operate the pump manually.
- Move the distributor arm.

To access these Manual Functions, select OTHER FUNCTIONS from the Main Menu. Then, select OTHER FUNCTIONS.

RUN  
PROGRAM  
VIEW REPORT  
**OTHER FUNCTIONS**

MAINTENANCE  
**MANUAL FUNCTIONS**  
PROGRAMMING STYLE

GRAB SAMPLE  
**CALIBRATE VOLUME**  
OPERATE PUMP  
MOVE DISTRIBUTOR

### 6.6.1 Grab Samples

Grab samples let you take a single sample on demand, collecting the sample in an external container (Figure 6-3). Keep in mind:

- When the sampler delivers a grab sample, it runs through a complete sampling cycle, using the current settings for volume and for line rinses or retries. If it is a two-part program, the sampler uses the settings for part A.
- To take a grab sample while the sampler is running a program, you must interrupt the program by pressing the Stop key. Restart the program by running it again.

To take a grab sample:

1. Disconnect the pump tube from the bulkhead fitting.
2. Place the end of the tube over a collection container.
3. Initiate the grab sample. Refer to the *Taking a Grab Sample* example.
4. Reconnect the pump tube to the bulkhead fitting.



*Figure 6-3 Taking a grab sample*

### Example: Taking a Grab Sample

1. After preparing to collect the grab sample in an external container, select GRAB SAMPLE from the MANUAL FUNCTIONS or the PAUSED screen.

**GRAB SAMPLE**  
CALIBRATE VOLUME  
OPERATE PUMP  
MOVE DISTRIBUTOR

2. Enter the sample volume.

SAMPLE VOLUME:  
\_\_\_ ml (10-9990)

3. Place the end of the pump tube over the container and press ↵.

GRAB SAMPLE  
PRESS ↵ WHEN READY!

4. The controller will complete a sample collection cycle.

PURGING  
SUCTION LINE

PUMPING 200 ml

PURGING  
SUCTION LINE

5. The collection cycle is complete when the display returns the MANUAL FUNCTIONS screen. Press the Stop key to return to the main or paused menu.

**GRAB SAMPLE**  
CALIBRATE VOLUME  
OPERATE PUMP  
MOVE DISTRIBUTOR

### 6.6.2 Calibrate Volume

The sampler delivers accurate sample volumes without calibration. If you find that sample volumes vary significantly from the programmed values, first check the suction line for proper installation. Be sure it slopes continuously downhill to the liquid source and drains completely after each sampling cycle. Then, compare the actual length of the suction line to the suction line length settings in the program to see that they match. Also check the pump tube for excessive wear and replace it if necessary.

You may want to calibrate when:

- A new pump tube is installed. Run the pump for five minutes before calibrating.
- The sample source is above the sampler.
- Sampling from pressurized lines (15 psi maximum).
- The controller has been reinitialized. Reinitializing the controller clears the calibration data.

Volumes cannot be calibrated while a program is running. The program must be paused or stopped first. For best results:

- Calibrate after the sampler has been installed on site.
- Use a graduated cylinder for volume measurement.

 <b>Note</b>
---

When the sampler delivers the sample volume, it runs through a complete sampling cycle, using the current settings for volume line rinses and retries. If it is a two part program, the sampler uses the volume setting for part A.

To calibrate sample volumes:

1. Disconnect the pump tube from the bulkhead fitting.
2. Place the end of the tube over a collection container.
3. Follow the steps in the *Calibrating Sample Volume* example.
4. Reconnect the pump tube to the bulkhead fitting.



## Example: Calibrating Sample Volume

1. After preparing to collect the sample volume in a graduated cylinder or similar container, select **CALIBRATE VOLUME** from the **MANUAL FUNCTIONS** or the **PAUSED** screen.

GRAB SAMPLE  
**CALIBRATE VOLUME**  
OPERATE PUMP  
MOVE DISTRIBUTOR

2. Enter the sample volume.

SAMPLE VOLUME:  
\_\_\_ ml (10-9990)

3. Place the end of the pump tube over the container and press ↵.

CALIBRATE VOLUME  
PRESS ↵ WHEN READY!

4. The controller will complete a sample collection cycle to deliver the programmed volume using its generated pump tables.

PURGING  
SUCTION LINE

PUMPING 200 ml

PURGING  
SUCTION LINE

5. The collection cycle is complete when the controller displays the programmed volume. Measure the actual volume delivered to the collection container and enter the amount here.

VOLUME DELIVERED:  
\_\_\_ ml

6. If you enter a volume more than twice or less than half the programmed volume, the message below is displayed.

200 ml  
ARE YOU SURE?  
**YES** NO

Select YES when the volume delivered matches the volume displayed on this screen. Select NO to re-enter the volume delivered.

7. The collection cycle is complete when the display returns the **MANUAL FUNCTIONS** screen. Press the Stop key to return to the main or paused menu.

**GRAB SAMPLE**  
CALIBRATE VOLUME  
OPERATE PUMP  
MOVE DISTRIBUTOR

### Example: Operating the Pump Manually

1. Select OPERATE PUMP from the MANUAL FUNCTIONS screen.
3. While pumping, the screen reports PURGING or PUMPING.

```
GRAB SAMPLE
CALIBRATE VOLUME
OPERATE PUMP
MOVE DISTRIBUTOR
```

```
PURGING . . .
PRESS THE RED 'STOP'
KEY WHEN DONE
```

2. Select the pump direction and press ↵. For this example, reverse is selected.
4. After pressing the Stop key, the display reports the number of pump counts.

```
SELECT DIRECTION
PUMP FORWARD
PUMP REVERSE
```

```
PURGED ____COUNTS
```

#### 6.6.3 Operating the Pump

You can manually operate the sampler's pump in the forward or reverse direction. You can do this to test the pump, draw a large sample, or purge the suction line.

When you run the pump manually, the sampler reports the number of pump counts during the pump's operation. Use this feature to obtain an estimate of the pump counts required to purge a nonstandard suction line.

You can also manually run the pump from the Main Menu. Press the "1" key and then ↵ to run the pump in reverse. Press the "3" key and then ↵ to run the pump forward.

#### CAUTION

Prolonged exposure to noise while the pump is operating could result in hearing loss. Tests indicate that the pump produces sound levels in excess of 85db at one meter. If you are manually operating the pump for long durations, always wear hearing protection.

### Example: Moving the Distributor Arm

1. Select MOVE DISTRIBUTOR from the MANUAL FUNCTIONS screen.
2. Enter the bottle number and press ↵.
3. The distributor arm moves to the selected bottle position then displays:

GRAB SAMPLE  
CALIBRATE VOLUME  
OPERATE PUMP  
**MOVE DISTRIBUTOR**

NOW AT BOTTLE \_\_\_\_

2. Enter the bottle number and press ↵.

GO TO BOTTLE \_\_\_\_  
(1-MAX)

Press the Stop key to return to the Manual Functions screen.

#### 6.6.4 Moving the Distributor Arm


The sampler lets you reposition the distributor arm. Use this feature to verify a bottle location when installing a bottle kit. You cannot move the distributor arm while the sampler is running a program.

Always use the steps shown in the *Moving the Distributor Arm* example.

#### CAUTION

Distributor drive gears may be permanently damaged if you attempt to rotate the distributor arm by hand or with any tool. Only move the distributor arm through the Manual Functions.

## 6.7 Reports



Interrogator

The BLZZRD sampler records a variety of data while running a sampling program. It uses the data to produce four reports:

- The Program Settings report, listing the program settings.
- The Sampling Results report, listing the program settings, time of samples and other program events.
- The Combined Results report, combining the sample event times with rainfall or module data.
- The Summary report, listing daily summaries of data collected. The Summary report may be rainfall and/or module data.

The sampler stores the data for each report in memory where they remain until you select RUN. Selecting RUN clears the memory so that it can store the data from the next program.

### 6.7.1 Collecting Reports



581 RTD

There are a number of ways to collect the reports:

- View reports (program settings, sampling results, and abbreviated summary reports) on the sampler's display.
- Collect the reports (except for summary reports) with a Teledyne ISCO 581 RTD and use a computer running Flowlink or Samplink to transfer the reports from the RTD to a file on the computer.

#### **Note**

The RTD is not compatible with samplers running software version 2.50. Refer to Section 1.4.2 for complete information.

- Collect the reports (except for summary reports) with an IBM PC or compatible computer running Flowlink.
- Collect the reports with an IBM PC or compatible computer running Samplink.

### 6.7.2 Viewing the Data

Because the display area is small, the data on the reports will be displayed a little differently from the data that is transferred to and viewed on your computer.

- For the Sampling Report, the sampler displays each program event, one at a time.
- For the Module and Rainfall Data reports, it displays daily summaries, instead of the full reports available with the RTD, Flowlink, or Samplink.

If you select sampling report or rainfall (step 3 in the *Viewing Reports* example), the sampler begins displaying the report data. The sampler advances automatically through the report items, displaying each item briefly. While the sampler advances automatically through the displays:

- Stop the automatic displays by pressing Stop once. Then, use the arrow keys to move manually through the report.
- Return to the main menu by pressing Stop twice.

**Example: Viewing Reports**

---

1. Select View Report from the Main Menu or View Data while in the Paused state.

RUN  
PROGRAM  
**VIEW REPORT**  
OTHER FUCTIONS

2. Select VIEW DATA and press ↵.

**VIEW DATA**  
SYSTEM IDS  
CONFIGURE REPORTS

3. Select a report or data type.

SELECT DATA TO VIEW:  
SAMPLING REPORT  
**FR-TEMP**  
RAINFALL DATA

4. Select DAILY SUMMARY to view the stored data, or CURRENT READINGS to view real-time data.

VIEW:  
**DAILY SUMMARY**  
CURRENT READINGS

Press the Stop key twice to exit the reports.

At the end of the report, the sampler leaves the last item displayed until you press:

- the arrow keys to move forward or backward manually through the report.
- Stop to return to the main menu.
- ↵ to start the automatic displays again.

If you select Module Data in step 3, the sampler displays the screen shown in step 4. Select the Daily Summary option and the sampler displays the stored data just as the Sampling Report was described above. Selecting the Current Readings option causes the sampler to display real-time readings. When there are multiple screens, they will be displayed alternately. You can quickly advance through the screens by pressing an arrow key or ↵. Pressing Stop exits the Current Readings mode.

### 6.7.3 Configuring Reports

The Configure Reports option lets you specify which of the Sampling Reports will be included in any output request. Each sampler is shipped with the option Flowlink Report selected.

The Samplink Report and Flowlink Report options allow you to tailor the output to the Teledyne ISCO software used to collect the reports. The Samplink Report option provides the Sampling Reports, but disables the output of the detailed partition data.

Flowlink report provides the Sampling Reports and the detailed partition data. Selecting the Flowlink report option enables the output of the detailed partition data so Flowlink can add the data to the site's database file.

When you are familiar with the available Sampling Reports, you may want to specify an output containing only the specific data you require. The custom report option will allow selection of each report and of detailed data. Data collection time, text file length, and storage space may be optimized using this option.

### Example: Configuring Reports

1. Select View Report from the Main Menu or View Data while in the Paused state.
3. Select the option that matches your data collection software, or tailor the output according to the reports needed.

```
RUN
PROGRAM
VIEW REPORT
OTHER FUCTIONS
```

```
SAMPLINK REPORT
FLOWLINK REPORT
CUSTOM REPORT
ALL REPORTS
```

2. Select VIEW DATA and press ↵.

```
VIEW DATA
SYSTEM IDS
CONFIGURE REPORTS
```

#### 6.7.4 Sampling Results Report

As the sampler runs a program, it records the program's events. Program events include such items as sample events, program enables and disables, power losses, and so on. Table 6-1 lists each event with a short description. You can view the report by following the steps in Section 6.7.2 as the Sampling Results report.

In the text file version of the report, sample-event entries include the time of the event as well as the sample and bottle numbers.

The Sample column contains the sample number deposited into the bottle. This information reflects the distribution selected for the program.

The letters in the Source columns are codes for the cause of the event. The letters in the Error column are codes for the causes of missed samples. The source and error codes appear with explanations at the end of the report. A list of codes appears in Table 6-1.

The last column in the Sampling Results report records the pump counts for the sample event. Large variations in pump counts from event to event indicate fluctuating heads, a relocated suction line, or a temporarily clogged strainer.

#### 6.7.5 Combined Results Report

At the time of each sample event, the sampler records the readings from each connected module, rain gauge, or SDI-12 sonde. The Combined Results Report presents the event time, sample and bottle number, and the respective readings from the device.

For flow modules, the readings are level and flow rate. For the pH module, the readings are temperature and pH. For the rain gauge and SDI-12 Sonde, the readings vary according to the selections enabled during the hardware setup in extended programming. When recording four or more sonde parameters, the Combined Results report will exceed the standard 40 character width. To read the additional columns, the data must be collected with Flowlink 4 for Windows software.

### 6.7.6 Summary Report

This report summarizes the rain gauge or module data. The types of readings are the same as those available to the Combined reports. The sampler compiles the data to present a daily total, the day's average, maximum, and minimum reading, and a listing of hourly averages. The sampler then graphs the day's readings.

Figures 6-4, through 6-8 show examples of reports configured to show program settings, sampling results, combined results, and a summary report.

**Table 6-1 Reports: Program Events, Source Codes, and Error Codes**

* Indicates events and codes that occur in Standard Programming	
Program Events (listed in the order that they most likely would occur)	
* Program Run	The sampler ran the sampling program.
* Program Started	The sampler started the sampling program.
* Program Enabled (Disabled)	The sampler was enabled or disabled according to program settings or an external instrument.
'A'/'B' Enabled (Disabled)	Program part 'A' or 'B' was enabled or disabled according to the program's settings.
Sample	A sample was taken as programmed. (See Source Codes in this table.)
* Sample Switch Time	A sample was taken at programmed switch time.
* Program Done	The sampler finished the sampling program.
'A'/'B' Done	The sampler finished program part 'A' or 'B'.
* Manual Pause	Stop was pressed during the program to enter the Manual Paused state.
* Program Stopped	You selected Stop from the manual paused options, terminating the program.
* Grab Sample	A grab sample was taken while in the Manual Paused state.
Pump Tube Replaced	The pump tube alarm was reset while in the Manual Paused state.
Calibrate Volume	The volume was calibrated.
Calibrate SDI Parameter	An SDI-12 sonde or module parameter was calibrated.
* Adjusted Pacing	The pacing interval was changed while in the Manual Paused state.
* Volume Changed	Sample volume was changed while in the Manual Paused state.
Adjust Enable	Sampler enable thresholds were changed while in the Manual Paused state.
Reset Rain History	The rain history was reset.
Adjust Suction Head	The suction head was changed while in the Manual Paused state.
* Manual Resume	The resume program option was selected from the Manual Paused state.
Program Paused (Resumed)	The program reached a programmed pause or resume time.
'A'/'B' Paused (Resumed)	Program part 'A' or 'B' reached a programmed pause or resume time.
* Power Failed	The power source was disconnected.
* Power Restored	The power source was reconnected.
Alarm Occurred	An alarm occurred.
Alarm Acknowledged	An alarm was acknowledged.
* Missed Password	The password was missed.

**Table 6-1 Reports: Program Events, Source Codes, and Error Codes (Continued)**

* Indicates events and codes that occur in Standard Programming		
<b>Source Codes</b> – The Source Codes listed below refer to why a sample was taken.		
* C	Calibrate sample	Sample volume delivered during calibration.
* D	Disable	Sample event at disable time.
* E	Enable	Sample event at enable time.
Ev	Event	Event-paced sample event.
* F	Flow	Flow-paced sample event.
* G	Grab sample	Sample volume delivered as a grab sample.
M	Command driven sample	A command driven sample.
PH	Phone command	Phone command sample.
R	Resume	Sample event at resume time.
* S	Start	Sample event at start time.
Sw	Switch time	Sample event at switch time.
* T	Time	Time-paced sample event.
TO	Time Override	Sample event at bottle-switch time.
<b>Error Codes</b> – The Error Codes listed below refer to what happened while a sample was being taken.		
* DJ	Distributor Jammed	The distributor jammed.
* IP	Sample in progress	Sample in progress when report was printing.
* L	Pump latch open	Unable to take sample because the pump band was opened.
LF	Not enough flow	Not enough flow for the minimum sample volume entered for samples with volumes proportional to flow.
M	Bottle Full	The composite bottle is full.
* ND	No distributor arm	There is no distributor arm or the stop is damaged.
* NL	No Liquid Detected	The sampler was unable to detect liquid.
* NM	No More Liquid	After the sampler detected liquid and while the sample was being taken, the liquid detector stopped detecting liquid.
* O	Sampler Shut Off	The sampler was turned off with the On-Off key during the sample event.
Ov	Probable Overflow	The sampler was directed to take a sample that would not fit into the bottle. Occurs for flow paced, time override programs only.
* P	Power Failed	The power supply failed during the sample event. Note: Power failures during five consecutive sample events results in a LOW BATTERY shut-down.
* PJ	Pump Jammed	The pump jammed.
* Sk	Sample Skipped	Sample not taken.
* US	User Stopped	You pressed the Stop key during the sample event.



SAMPLER ID# 3687447734 08:32 19-SEP-03  
Hardware: A1 Software: 2.10  
\*\*\*\*\* PROGRAM SETTINGS \*\*\*\*\*

-----  
SITE DESCRIPTION:  
"FACTORY051"

-----  
UNITS SELECTED:  
TEMPERATURE: F  
FLOW RATE: cfs  
FLOW VOLUME: Mgal

-----  
BUBBLER MODULE:  
WEIR  
90  
V-NOTCH

-----  
14, 950 ml BTLS

10 ft SUCTION LINE

-----  
PACING:  
TIME, EVERY  
2 HOURS, 0 MINUTES

-----  
DISTRIBUTION:  
SEQUENTIAL

-----  
200 ml SAMPLES

-----  
FIRST SAMPLE: 08:00  
MO TU WE TH FR

-----  
*Figure 6-4 Report: Program Settings*

*BLZZRD™ Portable Refrigerated Sampler*  
*Section 6 Running Programs*

---

```

SAMPLER ID# 3687447734 08:32 19-SEP-03
Hardware: A1          Software: 2.10
***** SAMPLING RESULTS *****
SITE: FACTORY051
Program Started at 07:53 TH 18-SEP-03
Nominal Sample Volume = 200 ml

                                COUNT
                                TO
SAMPLE  BOTTLE TIME  SOURCE ERROR LIQUID
-----  - - - - -  - - - - -
          08:00  PGM ENABLED
1,1     1    08:00     S           250
1,1     2    10:00     T           247
1,1     3    12:00     T           247
1,1     4    14:00     T           249
1,1     5    16:00     T           247
1,1     6    18:00     T           247
1,1     7    20:00     T           247
1,1     8    22:00     T           248
-----FR 19-SEP-03-----
1,1     9    00:00     T           237
1,1    10    02:00     T           236
1,1    11    04:00     T           237
1,1    12    06:00     T           241
1,1    13    08:00     T           241
          08:06  PGM STOPPED 19-SEP

SOURCE S ==> START
SOURCE T ==> TIME
-----

```

*Figure 6-5 Report: Sampling Results*

SAMPLER ID# 3687447734 08:32 19-SEP-03

Hardware: A1 Software: 2.10

BUBBLER MODULE: 2215419583

Hardware: A00000 Software: 1.07

\*\*\*\*\* COMBINED RESULTS \*\*\*\*\*

SITE: FACTORY051

Program Started at 07:53 TH 18-SEP-03

Nominal Sample Volume = 200 ml

SAMPLE	BOTTLE	TIME	LEVEL ft	FLOW RATE cfs	TOTAL FLOW Mgal
1,1	1	08:00	0.009	2.495	000000.000
1,1	2	10:00	0.012	2.576	000000.102
1,1	3	12:00	0.012	2.666	000000.262
1,1	4	14:00	0.012	3.162	000000.396
1,1	5	16:00	0.012	3.341	000000.556
1,1	6	18:00	0.012	3.713	000000.764
1,1	7	20:00	0.012	5.428	000000.928
1,1	8	22:00	0.012	6.864	000001.128
-----FR 19-SEP-03-----					
1,1	9	00:00	0.012	3.339	000001.306
1,1	10	02:00	0.012	2.773	000001.444
1,1	11	04:00	0.009	2.165	000001.560
1,1	12	06:00	0.012	2.066	000001.668
1,1	13	08:00	0.012	2.731	000001.799

*Figure 6-6 Report: Combined Results*

*BLZZRD™ Portable Refrigerated Sampler*  
*Section 6 Running Programs*

---

SAMPLER ID# 3687447734 08:32 19-SEP-03  
Hardware: A1 Software: 2.10  
\*\*\*\*\* COMBINED RESULTS \*\*\*\*\*  
SITE: FACTORY051  
Program Started at 07:53 TH 18-SEP-03  
Nominal Sample Volume = 200 ml  
FR-TEMP


SAMPLE	BOTTLE	TIME	F
1,1	1	08:00	41.4
1,1	2	10:00	34.2
1,1	3	12:00	34.4
1,1	4	14:00	34.5
1,1	5	16:00	34.5
1,1	6	18:00	34.3
1,1	7	20:00	34.3
1,1	8	22:00	34.1
-----FR 19-SEP-03-----			
1,1	9	00:00	34.1
1,1	10	02:00	34.2
1,1	11	04:00	34.3
1,1	12	06:00	34.4
1,1	13	08:00	34.2

SAMPLER ID# 3687447734 08:32 19-SEP-03  
Hardware: A1 Software: 2.10  
\*\*\*\*\* COMBINED RESULTS \*\*\*\*\*  
SITE: FACTORY051  
Program Started at 07:53 TH 18-SEP-03  
Nominal Sample Volume = 200 ml  
SAMPLE BOTTLE TIME  
-----  
NO RAIN GAUGE  
-----

*Figure 6-7 Report: Combined Results (continued)*

SAMPLER ID# 3687447734 08:32 19-SEP-03  
 Hardware: A1 Software: 2.10  
 BUBBLER MODULE: 2215419583  
 Hardware: A00000 Software: 1.07  
 SITE: FACTORY051






Summary Report for 18-SEP-03 (TH)  
 Day's Flow: 000001.487 Mgal  
 Average Flow Rate: 2.3061 cfs  
 09:20 Minimum Flow Rate: 2.4237 cfs  
 21:45 Maximum Flow Rate: 7.0207 cfs  
 Hourly Average Flow Rate:  
 00:00-01:00: NO DATA  
 01:00-02:00: NO DATA  
 02:00-03:00: NO DATA  
 03:00-04:00: NO DATA  
 04:00-05:00: NO DATA  
 05:00-06:00: NO DATA  
 06:00-07:00: NO DATA  
 07:00-08:00: NO DATA  
 08:00-09:00: 2.519 cfs  
 09:00-10:00: 2.447 cfs  
 10:00-11:00: 2.591 cfs  
 11:00-12:00: 2.552 cfs  
 12:00-13:00: 2.721 cfs  
 13:00-14:00: 2.978 cfs  
 14:00-15:00: 3.096 cfs  
 15:00-16:00: 3.223 cfs  
 16:00-17:00: 3.402 cfs  
 17:00-18:00: 3.668 cfs  
 18:00-19:00: 3.811 cfs  
 19:00-20:00: 3.828 cfs  
 20:00-21:00: 5.577 cfs  
 21:00-22:00: 6.264 cfs  
 22:00-23:00: 3.652 cfs  
 23:00-24:00: 3.023 cfs



Hour Ending: 08: 16: 24:  
 Units are 'cfs'

Figure 6-8 Report: Summary

**Example: Viewing System IDs**

1. Select View Report from the Main Menu or View Data while in the Paused state.  
  

2. Select SYSTEM IDs and press ↵.  
  
  

3. The controller will briefly display the controller ID and if installed, the Module ID, followed by any installed options.  
Press the Stop key to return to the Main Menu at any time.

## 6.8 System IDs

In addition to viewing report data and configuring reports, the View Report option on the main menu lets you view system identification information.

The sampler reports the sampler model number, ID (identification number), and software version number as system identification notes. The sampler also reports the same information for an attached module.

The model and ID numbers remain unchanged through the life of the sampler. The ID numbers appear on the sampler's reports and in files created by flowlink and samplink when they interrogate the sampler. The software packages use the ID number to make sure that data from different samplers is not combined in the same file.

The software version number changes when your sampler's software is updated.

The system will also list any software options that you have activated on your unit. These specialized software options are available for purchase from Teledyne ISCO's Special Product Applications Department.

# BLZZRD™

## Portable Refrigerated Sampler

### Section 7 SDI-12 Sondes

In order to provide you with the flexibility to use the water parameter sensors that best meet your needs, Teledyne ISCO has implemented support for SDI-12 protocol. Support of this standard allows you to use SDI-12 compatible devices in conjunction with your BLZZRD in Extended Programming mode.

#### 7.1 Teledyne ISCO AQ700 Sonde

The AQ700 Multi-Parameter Sonde can be used in any natural water up to 50 degrees C and 49 feet (15 meters) depth. Its ease-of-use design includes direct USB connection to your computer, real-time data viewing, and LED indicators for operating status.

Using SDI-12 protocol, the AQ700 can connect to the following Teledyne ISCO instruments:

- 6712 or BLZZRD Waste-water Sampler
- Signature® Flow Meter
- 2105 Interface Module

There are two versions of the AQ700 to choose from, depending on your monitoring requirements:

##### AQ702

- pH
- ORP (Oxidation Reduction Potential)
- Specific Conductance
- Optical DO
- Temperature
- Depth

##### AQ703

- Everything the AQ702 has, plus:
- Turbidity (w/ wiper)



Figure 7-1 Teledyne ISCO AQ700 Multi-Parameter Sonde

## 7.2 SDI-12 Data Parameters

The BLZZRD Sampler can accept up to 16 parameters from up to ten sensors with SDI-12 addresses from 0 - 9 (refer to Table 7-1). This data can be used in different ways. Data can be logged at user selectable time intervals for later download and processing by Teledyne ISCO's Flowlink® program. You can set up sampling programs that will be enabled to run only if conditions based on parameters monitored by the connected SDI-12 devices are met. The sampler can also be programmed to take samples each time certain parameter limits are exceeded.

The data parameters used by the sonde and sampler must match. Otherwise, the equipment could record the data values for one parameter as another's, such as reporting a pH value as being temperature.

Some sondes, such as the AQ700, have "plug and play" setup. Known as "ISCO Ready," these sondes can tell the sampler what values they have, their order, and units of measure. Other SDI-12 compatible sondes can still be used, but require manual setup to identify proper data types for each data value reported.

This section shows you how to set up both ISCO Ready and other SDI-12 sondes for use with the sampler.

 <b>WARNING</b>
---

**SDI-12 Sondes have not been approved for use in hazardous locations as defined by the National Electrical Code.**

 <b>CAUTION</b>
--

Before you install any sonde, you must take the proper safety precautions. Refer to Appendix C in the back of this manual. Additional safety considerations may be required. Check applicable guidelines, codes, and regulations of federal, state, city, and county agencies.



**Table 7-1 SDI-12 Sonde Parameters**


Parameter	Units	Range	Resolution	Bytes
Temperature	°C	-40–100	0.1	2
Conductivity	mS/cm	0–100	0.01	
Specific Conductance				
Total Dissolved Solids	g/L	0–90		
Salinity	ppt	0–70	0.1	1
Dissolved Oxygen	mg/L	0–20		
pH	—	0–14		
ORP	mV	-999–999	0.0001	4
Level	m	0–6.5279		
Ammonium-Nitrogen	mgN/L	0–200		
Ammonia-Nitrogen				
Nitrate-Nitrogen				
Turbidity	NTU	0–5000		
Chloride	mg/L	0–1000		
Chlorophyll	ug/L	0–400		
Other	—99999.0–99999.0			

### 7.3 Scanning for Sondes

The sampler controller is shipped with the automatic sonde scan disabled (OFF). To use a sonde with the a sampling program, the controller must be configured to scan for sondes.

To scan for a sonde, connect the sonde to the sampler’s Rain Gauge connector and:

1. From the EXTENDED PROGRAMMING mode’s Main Menu, select OTHER FUNCTIONS, then HARDWARE.
2. Step through the HARDWARE screens until you see the following:



This screen indicates that the sampler has not detected an SDI-12 device and the automatic scan at startup is turned off. These settings are the factory defaults. If you have made changes to the settings or have previously configured a sonde, it may appear different than above.

3. Use the arrow keys to select NO SDI-12 SONDE and press Enter. The next display indicates that it is scanning for sondes.

```
PLEASE WAIT!  
  
LOOKING FOR  
SDI-12 SONDES
```

4. If a sonde is found during the scan, the controller will briefly report the number of sondes, then advance to a screen to set up the sondes. (refer to the *Sonde Setup screens* example.)

Also, if a sonde is found during the scan, the controller changes AUTO SDI-12 SCAN to ON. This enables the controller to scan all SDI-12 devices each time the sampler is turned on to ensure that the data will be available for the program.

If a sonde is not found, it reports 0 SONDE(S) FOUND. The screen reverts to the HARDWARE screens without changing the AUTO SDI-12 SCAN setting. Verify that the connections are correct and that the SDI-12 device is powered and configured for use.

After the sonde has been found and configured, the parameter data is available for use in the sampling programs. Section 5 describes the Extended Programming features.

Note that after a sonde has been found and configured, the SDI-12 HARDWARE screen will no longer appear as it did in step 2. Instead, it will report the available SDI-12 data, similar to the screen below.

```
↑ SDI-12 DATA:  
  TEMP0 pH0 SD_CO0  
  SAL0 05DATA0 DO0  
  ORP0 LEVEL0 09DATA0 ↓
```

Selecting SDI-12 DATA will give you the following options:

```
NO SDI-12 SONDE  
SETUP KNOWN SONDES  
SCAN SDI-12 SONDES
```

- NO SDI-12 SONDE – select this option to “remove” the sonde from the sampler’s configuration. The sampler will rebuild memory partitions without the sonde parameters. (See section 7.8.)
- SETUP KNOWN SONDES – select this option to:
  - configure a sonde (see *Sonde Setup screens* example).
  - select which parameter data to store in memory (see section 7.6).
  - calibrate parameters (section 7.7.1).
  - validate parameter data (7.7.2).
- SCAN SDI-12 SONDES – select this to rescan the SDI-12 data bus to detect any new or removed sondes.

### Example: Sonde Setup screens

1. Turn off the sampler controller and connect the SDI-12 sonde to the rain gauge connector.
2. Turn on the sampler controller.
3. From the EXTENDED PROGRAMMING mode's Main Menu, select OTHER FUNCTIONS, then HARDWARE. Select NO SDI-12 SONDE.

```

↑ NO SDI-12 SONDE
  AUTO SDI-12 SCAN OFF ↓
  
```

4. The controller searches sondes.

```

PLEASE WAIT!
  LOOKING FOR
  SDI-12 SONDES
  
```

5. When it finds a sonde, it will briefly report the number of sondes found. You can wait for the screen to advance or press ↵.

```

1 SONDE(S) FOUND
  
```

6. Select the address of the sonde you want to set up. If more than one sonde is connected, multiple addresses will be displayed. Use the arrow keys to

select the sonde and press ↵.

```

SELECT ADDRESS OF
SONDE TO SETUP:
  7 DONE
  
```

7. The controller will briefly display the screen below. Wait a few seconds for it to advance, or press ↵.

```

Company Model
Optional Information
SONDE AT ADDRESS 7
HAS 5 DATA VALUES
  
```

8. The next screen displays all available parameter data types. If there are more than three parameter data types, the display scrolls. Use the arrow keys to scroll. The example screens below show five parameters from a Teledyne ISCO-ready sonde.

```

DATA FROM SONDE 7
1=TEMPERATURE (°C)
2=COND. (mS/cm)
3=SALINITY (ppt) →
  
```

```

DATA FROM SONDE 7
←4=D.O. (mg/L)
5=pH
  DONE →
  
```

Continued...

#### Note

As many as ten SDI-12 devices may be connected to the sampler controller. Each device must have a unique address, numbered 0 through 9. Conflicting SDI-12 device addresses must be reassigned. Refer to the manufacturer's instructions.

## 7.4 ISCO-Ready Sondes

If the BLZZRD sampler scans for sondes and finds one that is ISCO-ready, it will automatically recognize the sonde's data types and units. You can then use Hardware options to setup the definitions for data storage and calibrate the sonde.

Even with an ISCO-ready sonde, the sampler may find a parameter that is not on its list of supported parameters (Table 7-1). It can still work with such a parameter. Such a parameter will be identified as something like O5DATA1 and you will have to mentally associate the appropriate data type with that label. However, once such data is brought into Flowlink, data sets can be manually changed so that reports show the appropriate labels.

*Sonde Setup screens, continued...*

Teledyne ISCO-ready sondes automatically report the correct parameter labels and units for each data type in the sonde. Sondes that are not Teledyne ISCO-ready will only report that a data type is available. These undefined parameters will appear as "OTHER."

```
DATA FROM SONDE 7
1=OTHER
2=OTHER
3=OTHER
```

```
DATA FROM SONDE 7
←4=OTHER
5=OTHER
DONE
```

You can label the parameters so the BLZZRD controller will display the parameters and measurement units correctly. To do so, you must know what parameters are reported by the sonde. This is typically documented by the manufacturer of your SDI-12 device, or available by connecting to the device using a terminal communications program from a personal computer. Refer to your SDI-12 device's manual for more information.

Use the arrow keys to select each channel and press ↵. Then use the arrow keys to scroll through the list of recognized data types (see Table 7-1). Highlight the correct type and press ↵. After reviewing or modifying the data types, select

DONE and press ↵.

- The controller then asks which parameter data to store. Stored data will be available for reports and for retrieval with Flowlink. To store all parameters in memory select YES. To choose only a few parameters, select NO.

```
SONDE 7, STORE
ALL PARAMETER DATA?
YES NO
```

If NO, the screens will ask about each parameter individually. Repeat for each parameter.

```
SONDE 7, STORE
TEMPERATURE DATA?
YES NO
LABEL IS TEMP7
```

- The sonde and controller are set up. You can repeat steps 7 through 8 for additional sondes, or select DONE. When you select DONE, the sampler configures the memory partitions according to step 8.

```
SELECT ADDRESS OF
SONDE TO SETUP:
7 DONE
```

## 7.5 Non-ISCO-Ready SDI-12 Sondes

If you have connected an SDI-12 sonde that is not ISCO-ready, you will need to define the sonde's data values.

When you select the address of the sonde you want to set up, the company model and other information will be displayed along with the number of data values the sonde has.

When you advance to the next screen (see step 8 in the example), the data types will be listed as OTHER because you have not defined them yet. Move between the data types using the arrow keys. Select one to change by pressing ↵.

A list of possible parameters will be displayed. Use the arrow keys to move through the options. When the one you want is blinking, press ↵. Repeat the process until you have defined all of the data parameters for your sonde.

Use the arrow key to move to DONE and press ↵. You will then be asked if you want to store the parameter data.

## 7.6 Sonde Setup - Storing Parameter Data

Whether you are using an ISCO-ready sonde or other SDI-12 compatible sonde, you will need to tell the sampler controller if you want to store all of the parameter data or data from selected parameters.

Data is stored in memory partitions. A fixed amount of memory is available to store data. If you are using some parameters to simply trigger an event or enable condition, it is not necessary to store the data.

If *all* of the parameters' data will fit in memory, you will be asked if you want to store all of the parameter data. Answering NO will still allow you to selectively store parameter data; otherwise, all data will be stored.

You will be asked to store parameter data during the setup process. You can also manually change the setup using options in the Hardware menu.

To access the setup option in the Hardware menu, select Other Function from the main menu, then choose Hardware. Advance through the screens until you see one that displays SDI-12 DATA. Select this screen to display the following:

```
NO SDI-12 SONDE
SETUP KNOWN SONDES
SCAN SDI-12 SONDES
```

Select SETUP KNOWN SONDES and then indicate the address of the sonde you want to setup. A screen containing the sonde model and number of data values will appear briefly. Then you will be asked to verify, parameter by parameter, what data you want to store. The screens will look similar to the one below:

```
SONDE 7, STORE
TEMPERATURE DATA?
YES NO
LABEL IS: TEMP7
```

Indicate if you want to store the data for that particular parameter and then advance through the rest of the parameter screens. Once you have defined your data storage options, select DONE.

You will be prompted that you have entered a new hardware setup and asked if you want to erase the stored data. Indicate YES or NO. A partition memory status message will be displayed and (if a parameter exists that can be calibrated) you will then advance to the calibration screen (refer to section 7.7). If you do not want to calibrate or validate, select DONE to return to the Hardware Options screens.

## 7.7 Sonde Calibration and Validation

**You should calibrate your sonde in the lab, according to manufacturer's recommendations.**

### **Note**

The AQ700 sonde's PC software program provides calibration for its sensors. Refer to the AQ700 user manual for complete instructions for calibration of each parameter.

The BLZZRD sampler allows for field calibration of level, pH, Dissolved Oxygen (DO), and Oxidation Reduction Potential (ORP); however, field calibrations are stored in the sampler, **not in the sonde**. They are used to modify readings from the sonde until a new sonde is connected. When a sonde is moved to another sampler, it will retain internal calibration constants set in the lab, but will not retain any that were set in the field.

Although field calibration is limited to four parameters, you can *validate* any of the available parameters. Validation allows you to immerse the sonde in a known solution and verify the sonde's readings. The validation process does not modify any readings like calibration does; it merely alerts you to inaccurate ones.

No data will be stored, nor will enable conditions be updated, while calibration or validation is being performed.

### 7.7.1 Calibration Screens

To access the calibration screens from the main menu, select Other Functions > Hardware. Advance to SDI-12 DATA:

```
NO SDI-12 SONDE
SETUP KNOWN SONDES
SCAN SDI-12 SONDES
```

Select SETUP KNOWN SONDES and DONE. The screen will list the parameters you can calibrate, similar to the example below. The sonde's address will appear after the parameter (such as pH0 and pH5).

```
SELECT PARAMETER
TO CALIBRATE:
pH5 DO5 ORP5 DONE
```

The parameter selected will determine the content of the calibration screens:

- Level - Enter the current measured level. The sampler will use this value to offset the level reading of the sonde.
- pH - You will be given the choice of three different pH solutions: 4.00, 7.00, and 10.00. After making your selection, rinse the probe and place in the pH buffer. If the reading is off by more than 0.5 pH units, you will be asked to verify if the reading is correct.

- DO - For dissolved oxygen, you can choose between two methods: DO Standard and Saturation. The saturation method requires a temperature reading between 0.0° and 47.0° C. You will also need to enter your altitude and barometric pressure.
- ORP - You will asked to enter the value of your calibration standard.

When calibration is complete, select DONE to proceed to the validation screen. If you do not want to validate any parameters, select DONE again to return to the Hardware option screens.

### 7.7.2 Validation Screen

The validation screen appears after you select DONE on the calibration screen. It will be similar to the screen shown below.

```
SELECT PARAMETER  
TO VALIDATE:  
← SAL5  O5DATA5  DO5  
  ORP5  LEVEL5   DONE →
```

Select the desired parameter to validate. At the prompt, place the probe in the standard. When the reading is stable, press ↵. The reading will appear at the bottom of the screen.

Select DONE to return to the Hardware option screens.

### 7.8 Disconnecting Sondes

When you have disconnected all sondes, and have interrogated the sampler for all its data, you can indicate that there is no SDI-12 sonde.

From the main menu, select Other Functions, Hardware. Advance through the screens until you see one that displays SDI-12 DATA. Select this screen to display the following:

```
NO SDI-12 SONDE  
SETUP KNOWN SONDES  
SCAN SDI-12 SONDES
```

Select NO SDI-12 SONDE. You will advised that you have a new hardware setup:

```
NEW HARDWARE SETUP--  
DOWNLOAD DATA NOW  
OR LOSE ALL DATA!  
DONE
```

If you have collected all the data, select DONE. A partition memory message will be displayed briefly, and you will be returned to the Hardware options screens.

If you do not have a sonde attached, you may want to turn off automatic scanning for sondes. This will eliminate the time taken to scan for sondes whenever the sampler is turned on.





# BLZZRD™

## Portable Refrigerated Sampler

---

### *Section 8 Remote Operation*

The BLZZRD controller can be operated remotely from external devices. External controlling devices may be a computer via RS-232 serial communications. Remote telephone operation requires the 6721 LTE modem.

#### **8.1 Computer Operation**

The sampler controller can be operated from a remote location using a computer. There are two different levels of computer control:

**Menu control** is used to start and stop the current sampler program (or programs if in Extended programming mode) and to access other menu commands. This level of remote operation relies on the sampler program for much of its capabilities.

**External program control** ignores the sampler's program and current state. Using external control, you can turn on the sampler, take a sample, and request sampler status data. The pump, distributor, and serial data output simply respond to the external commands.

To operate the sampler remotely on either level, you must first establish RS-232 serial communication between the computer and the controller. To do so:

1. Connect your computer to the BLZZRD controller's interrogator port with a standard interrogator cable, P/N 60-2544-044 (for 9-pin serial ports).
2. On your computer, open a terminal emulator program such as HyperTerminal, and configure it for:
  - baud rate of 19,200 or less
  - no parity
  - 8 data bits
  - 1 stop bit
  - VT100 terminal emulation
3. Using your communications software, connect to the sampler. Press and hold the question mark (?) on the computer keyboard until the sampler returns its banner string. A caret (>) prompt will appear on the screen when the sampler is ready to accept commands.




Figure 8-1 Press and hold ?? to connect to the sampler

### 8.1.1 Menu Control

Once you have remotely connected to the sampler, menu commands can be entered from your computer. These menu commands are listed in Table 8-1.

To access the menu commands, type MENU at the “>” prompt and press ENTER. The screen will return a list of the available options. Note that the current state of the controller will determine the available menu options. Then, type the desired command(s), confirming with ENTER.




Figure 8-2 Example list of remote menu commands

**Table 8-1 BLZZRD Remote Menu Commands**

Menu Command	Description
0 or START	This command will immediately start any program set to WAIT FOR PHONE CALL and currently waiting for the remote start command (see Section 4.6).
1 or RUN1	This command will load Program 1 as the sampler's current Extended program, reset partition if the storage interval has been changed, and run the program. RUN1 is valid when the sampler is in standby, or when it is waiting for a start command. RUN1 will also run the current program while in Standard programming.
2 or RUN2 3 or RUN3 4 or RUN4	These commands will load the selected program number as the sampler's current Extended program, reset partition data if the storage interval has changed, and run the program. They are valid when the sampler is in standby, or when it is waiting for a start command. The commands are available for Extended programming only.
5 or DISABLE	This command is similar in functionality to a pin F low signal (a signal typically used to enable or disable the sampler from a flow meter). At run time, the sampler starts with the sampler in the enabled state. DISABLE is valid while a program is running. If successful, a display of REMOTE ENABLE IS FALSE will be returned.  A Remote Disable is the highest priority disable. If the sampler is disabled with this command, no condition other than menu command 6, ENABLE, will allow the program to become enabled again.
6 or ENABLE	This command is used to re-enable a disabled sampler. ENABLE is valid while a program is running. If successful, a display of REMOTE ENABLE IS TRUE will be returned.
7 or TAKE_SAMPLE	This command causes the sampler currently running a program to take a sample. The sample will count as one of the program's samples and is placed in the current bottle (Part A's current bottle when using two-part programming). TAKE_SAMPLE is valid while the program is running, after the start time, but not while in the MANUAL PAUSE screens. If successful, a response similar to PUMPING XXML SAMPLE INTO BOTTLE ## will be returned. If not, a response similar to NO SAMPLE, PROGRAM WAITING START will be returned.
ST or STATUS	This command causes the sampler to return its status information and current parameter data.
SD or SCREEN_DUMP	This command causes the sampler to return the information currently displayed on the sampler's screen. The remote screen follows the sampler's display as the information is updated. This mode can be used to observe from a remote location while the sampler is programmed on-site.
P or PAUSE	This command causes a sampler currently running a program to enter the MANUAL PAUSE menu. While you are connected in this mode, the sampler keypad becomes disabled, and your computer's keypad will be redirected to simulate the sampler keypad. The following keys will be activated:  <Enter>, decimal point, and number keys.  <Esc>, S, or s = STOP.  <Backspace>, L or l, and U or u = Left/Up arrows.  R or r, and D or d = Right/Down arrows.

*Sampling reports*

You can retrieve the sampling reports at any time using the same serial connection. Type "Q" to exit the menu. Then at the > prompt, type "REPORT" and press Enter. The sampler controller will send reports as configured.

### 8.1.2 External Program Control

Under external program control, the sampler operates according to received commands. An external program can be as simple as manually entered commands to collect a sample when needed, or as complex as a user-developed process control program (for SCADA or other systems) that directs sample collection according to a system's needs.

The BLZZRD controller will respond to four commands:

- Turn on the sampler
- Take a sample
- Send status
- Send data

Note that commands and responses are comma-separated pairs. The first set of alpha-characters is an identifier; the second set of numeric characters indicates the value. The BLZZRD controller will not receive and process the command unless it is followed by a carriage return.

**Turn on the sampler** – If the sampler is OFF, the sampler responds by placing itself in the ON state and returning a status string to report the new status. (See *Send Status* for a description of this string.)

The command to turn the sampler ON and return status is:

STS,2<CR>

**Take a Sample** – This directs the sampler to take a sample. The command specifies how much volume to take, and which bottle to place the sample in. The sampler responds by taking a sample and returning a status string that reports the new status.

An example of a command to take a sample is shown below:

BTL,2,SVO,100<CR>

This example instructs the sampler to place the distributor arm over bottle 2 (BTL,2) and take a 100 ml sample (SVO,100).

While most settings of the sampler program (start times, enables, distribution schemes, etc.) are ignored when responding to external commands, the sampler does rely on two program settings that “configure” the sampler. These settings are:

- Number of Bottles—is used to determine the distributor arm position and the acceptable range for the BTL values.
- Suction Line Length—Used with liquid detection to deliver correct sample volumes according to the SVO value.

**Send Status** – This command causes the sampler to return a data string that identifies the controller and lists its current status.

The command to send current status is:

STS,1<CR>

The status is received as pairs of comma-separated identifiers and values. A complete description of each pair is listed below.

MO,BLZZRD, is the model number of the sampler.

ID,2424741493, is the sampler's unique ID number.

TI,35523.50000, is the sampler's current time, shown in a date-time format based on the number of days since 00:00:00 1-Jan-1900, and the time shown as a fraction.

The possible states returned by the STS,1 command are:

- 1 = WAITING TO SAMPLE.
- 4 = POWER FAILED (for short time after power is restored).
- 5 = PUMP JAMMED (must be resolved before continuing).
- 6 = DISTRIBUTOR JAMMED (must be resolved before continuing).
- 9 = SAMPLER OFF.
- 12 = SAMPLE IN PROGRESS.
- 20 = INVALID COMMAND. Possible causes may be:
  - identifier code is not supported.
  - bottle requested is not in current configuration
  - sample volume requested is outside its range (10 - 9990 ml)
  - day (Set\_Time) must be 5 digits and more recent than 1977
- 21 = CHECKSUM MISMATCH. (see "Optional checksum" on page 78-8)
- 22 = INVALID BOTTLE. (bottle requested is not in the current configuration)
- 23 = VOLUME OUT OF RANGE. (the sample volume requested is outside its range (10-9990 ml)

STI,35523.41875, is the most recent sample time.

BTL,2, is the bottle that received the most recent sample.

SVO,100, is the most recent sample's volume.

SOR,0, is the results of attempting the most recent sample.

The possible results are:

- 0 = SAMPLE OK
- 1 = NO LIQUID FOUND
- 2 = LIQUID LOST (not enough liquid)
- 3 = USER STOPPED (using the Stop Key)
- 4 = POWER FAILED
- 5 = PUMP JAMMED

- 6 = DISTRIBUTOR JAMMED
- 8 = PUMP LATCH OPEN
- 9 = SAMPLER SHUT OFF (while sampling)
- 11 = NO DISTRIBUTOR
- 12 = SAMPLE IN PROGRESS

CS,xxxx is the byte-by-byte sum of the entire string "MO,...,CS," (see "Optional checksum" on page 8-8)

**Note**

For commands that are not understood (INVALID COMMAND or CHECKSUM MISMATCH) or when no samples have been taken, the most recent bottle information is left off the response string.

**Send Data** - This command causes the sampler to return the Send Status appended with data from any module, rain gauge, or SDI-12 device. Refer to Table 8-2 to interpret the returned string.

**Note**

This returns the same data as the programmable software option Serial Output, as described in Section 5.15.9 of this manual.

The command to send data is:

DATA<CR>

**Table 8-2 Serial Data Codes**

Identifier	Parameter	Units
B? <sup>1</sup>	Bottle number and time	days since 1-Jan-1900
CR	Current day's rain	tips <sup>2</sup>
CS <sup>3</sup>	Checksum	unsigned long
DE	Description	BLZZRD Sampler
FL	Flow	cubic meters per second
ID	Unit specific identifier	10 digit unsigned long
LE	Level	meters
MO	Model	BLZZRD
PE	Percent	percent of full scale
PH	pH	none
PR	Previous day's rain	tips <sup>2</sup>
RA	Rain	tips <sup>2</sup>
RTE	Refrigerator Temperature	degrees celsius
SS	Sampler enable state	logical
TE	Temperature	degrees celsius

**Table 8-2 Serial Data Codes (Continued)**

<b>Identifier</b>	<b>Parameter</b>	<b>Units</b>
TI	Current time	days since 1-Jan-1900
VE	Velocity	meters per second
VO	Volume	cubic meters
VSI	Velocity signal strength	percent
VSP	Velocity spectrum strength	percent
CO_ 4	Conductivity	milliseimens per centimeter
DO_ 4	Dissolved oxygen	milligrams per liter
PH_ 4	pH	none
SA_ 4	Salinity	parts per thousand
SP_ 4	Specific Conductance	milliseimens per centimeter
TD_ 4	Total Dissolved Solids	grams per liter
TE_ 4	Temperature	degrees Celsius
OR_ 4	Oxidation reduction potential	millivolts
LE_ 4	Level	meters
MM_ 4	Ammonium-nitrogen	milligrams nitrogen per liter
MA_ 4	Ammonia-nitrogen	milligrams nitrogen per liter
NI_ 4	Nitrate-Nitrogen	milligrams nitrogen per liter
TB_ 4	Turbidity	nephelometric turbidity units
CL_ 4	Chloride	milligrams per liter
CP_ 4	Chlorophyll	micrograms per liter
FL_ 4	Flow rate	cubic meters per second
VO_ 4	Flow volume	cubic meters
xxD_ 4, 5	Other data not listed above	

<sup>1</sup> The three most recent samples will have entries as B?. The “?” will be the first bottle number in the set. If the sample is a grab sample, it will be shown as B0.

<sup>2</sup> Rainfall tips roll over every 256 counts.

<sup>3</sup> The checksum does not include the checksum, carriage return, and linefeed. The string is terminated with <CR><LF>.

<sup>4</sup> A “\_” in the identifier indicates an SDI-12 address.


<sup>5</sup> “xx” matches the data label for the SDI-12 parameter.

*Optional checksum*

Optionally, commands to the sampler can be validated using a checksum. Checksum validation can avoid problems caused by faulty program control or communication lines. When using the checksum, add the CS identifier and a comma at the end of the command, and the checksum value of the string up to the last comma. For example, the command with checksum validation to turn the sampler on would appear as:

STS,2,CS,582<CR>

The numeric value of each character is its ASCII equivalent expressed in decimal format. The checksum of “582” was calculated as shown below.



**8.1.3 Remote Control of Sampler Keypad**

To control the sampler keypad remotely, type CONTROL at the “>” prompt and press ENTER. The sampler display appears on your computer monitor as you step through the programming screens. While in this mode, the computer keys will be redirected to simulate the sampler’s keypad, and the sampler keypad itself will be disabled to avoid any conflict. The active keys and their corresponding functions are given in Table 8-3.

Table 8-3 Remote Control of Sampler Keypad	
Computer	Sampler
<Esc>, S, s	STOP
L, l, U, u, <Backspace>	Left / Up
R, r, D, d	Right / Down
O, o	ON
<Enter>, arrows, decimal, numbers	Same as sampler
NOTE: You will not be able to turn the sampler off remotely.	



## 8.2 SMS Commands

BLZZRD controllers equipped with an 6712 LTE modem can be operated using SMS commands that you send to the sampler.

Commands to start the program, load a selected program, enable/disable the sampler, take a sample, and report sampler status can be sent to the sampler by cell phone.

When you send a command to the sampler, it will carry out the command and also respond with an acknowledging message to all programmed destination phone numbers.

With a cell phone, simply send the command in a text message to the 6712 LTE modem's phone number.

The command must contain the string "6712," along with one of the commands listed in the table below. The middle and right columns are interchangeable, with the right column containing abbreviated commands. Commands are not case-sensitive and do not require spacing.

<b>Prefix</b>	<b>Command</b>	<b>Short Version</b>
6712	START	C0
	RUN1	C1
	RUN2	C2
	RUN3	C3
	RUN4	C4
	DISABLE	C5
	ENABLE	C6
	TAKE SAMPLE	C7
	STATUS	C9

Example command: 6712status

Shortened Version: 6712C9



# BLZZRD™

## Portable Refrigerated Sampler

---

### *Section 9 Maintenance*

This section describes how to maintain and troubleshoot your BLZZRD. If you think your unit requires repair, or if you have questions concerning its operation or maintenance, contact Teledyne ISCO's Technical Service Department:

Phone: (866) 298-6174  
(402) 464-0231  
FAX: (402) 465-3001  
E-mail: [iscowatersupport@teledyne.com](mailto:iscowatersupport@teledyne.com)

#### 9.1 Maintenance Checklist

Before each use:

- Inspect the pump tube for wear. Replace it if necessary.
- Clean the pump tubing housing.
- Change the suction line if necessary.
- Clean the bottles, suction line, strainer and pump tube.
- Check the humidity indicator.
- Defrost the refrigerator compartment if necessary.

Every five years:

- Replace the controller's internal battery.

#### 9.2 Cleaning Guidelines

Keeping the sampler clean and protected from harsh elements may extend the usable life of the sampler. The following cleaning instructions can be performed as needed.

##### 9.2.1 Cleaning the BLZZRD Exterior

The BLZZRD controller and refrigerator exterior may be cleaned with warm soapy water and a rag. Never use acids or solvents to clean the sampler. If there is excessive dirt and debris, the sampler may be hosed off with water. Do not use a pressure washer; this may force water past the protective seals.

 <b>CAUTION</b>
--


Do not tip the sampler on its side or completely upside-down. Positions other than the sampler's normal upright position may cause oil to run into the compressor inlet, which can permanently damage the cooling system. If the sampler is tipped or turned over, the sampler's refrigeration system must not be operated for at least one hour after returning the sampler to its upright position.

When cleaning the controller and top cover, cap the connectors at the back of the controller tightly. Make sure the pump drain hole (located on the bottom right-hand side of the pump, beneath the pump band) is open and free of debris or buildup.

### **9.2.2 Cleaning the Refrigerated Compartment**

The refrigerator interior may be cleaned with warm soapy water and a rag. Never use acids or solvents to clean the sampler. Water that has collected in the refrigerated compartment can be drained by removing the plug at the bottom of the compartment.

To clean the bottom of the compartment, the bottom plate can be removed from the refrigerator by lifting it out by the arrow handle. Do this by lifting the front edge and pulling up. The plate will lift past the evaporator and out from the refrigerator assembly, allowing access to the bottom of the compartment. Be sure to reinstall the plate in the same orientation. Use the arrows to properly orient the bottom plate.



*Figure 9-1 Align arrows when installing the bottom plate*

### **9.2.3 Defrosting the Refrigerated Compartment**

Defrosting the refrigerator compartment may be necessary, depending on the humidity of its operating environment. If ice forms on the walls of the interior compartment, never use sharp objects to remove it. Instead, simply remove power and allow the ice to melt. Drain the ice melt, or mop it up with a towel.

 **CAUTION**

Do not tip the sampler on its side or completely upside-down. Positions other than the sampler's normal upright position may cause oil to run into the compressor inlet, which can permanently damage the cooling system. If the sampler is tipped or turned over, the sampler's refrigeration system must not be operated for at least one hour after returning the sampler to its upright position.

### **9.2.4 Sample Bottles**

The sample bottles have a wide mouth to facilitate cleaning. Wash them with a brush and soapy water, or use a dishwasher. You can autoclave the glass bottles.

### 9.2.5 Cleaning or Replacing Wetted Parts

For general cleaning, you can wash the strainer and sample bottles with a brush and soapy water, then rinse with clean water. You can clean the liquid path through the wetted parts by placing the strainer in a cleaning solution and pumping it through the delivery system. Then place the strainer in clean water and pump it through the delivery system to rinse it. If these items are severely contaminated, replace them.


For application-specific requirements, consult with your laboratory to establish cleaning or replacement protocols.

 **Note**

For critical sampling applications, consider replacing the suction line, pump tube, and discharge tube. Replacement eliminates the possibility of any cross contamination from previous sampling sites.

## 9.3 Maintenance Screens

When you select OTHER FUNCTIONS from the main menu, and then select MAINTENANCE, the following screen will be displayed:



Using these options, you can set the sampler's time and date, set the number of pump counts for the pump tube warning, reset the internal battery count-down timer, and run a set of diagnostic tests for troubleshooting purposes.

### 9.3.1 Set Clock

Teledyne ISCO ships samplers from the factory with the clocks set to Central Standard Time. Reset the clock when installing the sampler in a different time zone, for daylight savings time, or when needed.

When setting the clock and calendar, use a 24-hour clock for times and the day-month-year format for dates. Note that the example "Setting the Clock" displays some positions with blanks. The blank positions are called fields. When viewing the actual screen on your sampler, the fields will contain the sampler's current settings for the time and date.

### Example: Setting the Clock

1. From the Main Menu select OTHER FUNCTIONS.

```

RUN
PROGRAM
VIEW REPORT
OTHER FUNCTIONS
    
```

2. Select MAINTENANCE.

```

MAINTENANCE
MANUAL FUNCTIONS
PROGRAMMING STYLE
    
```

3. Select SET CLOCK.

```

SET CLOCK
PUMP TUBE ALARM
INTERNAL BATTERY
DIAGNOSTICS
    
```

4. Enter the time and date using the arrow and number keys.

```

ENTER TIME AND DATE:
HH:MM DD-MON-YY
_: _ - _ - _ - _
    
```

To move from one field to another without changing the setting, press the arrow keys. Use this technique when you want to change only one or two settings.

Change the setting by typing a new number. Press  $\downarrow$  to accept the new setting.

For example, to enter 2:00 pm (2:00 pm is 14:00 on a 24-hour clock), type 14. Press  $\downarrow$ . Next, type 0 (zero) for the minutes, and press  $\downarrow$ . To enter a date, such as January 22, 2003, type: 22  $\downarrow$  01  $\downarrow$  03  $\downarrow$ .

5. Press Stop to return to the Main Menu.

### Example: Resetting the Pump Tube Alarm

1. From the Main Menu select OTHER FUNCTIONS.

```

RUN
PROGRAM
VIEW REPORT
OTHER FUNCTIONS
    
```

2. Select MAINTENANCE.

```

MAINTENANCE
MANUAL FUNCTIONS
PROGRAMMING STYLE
    
```

3. Select PUMP TUBE ALARM.

```

SET CLOCK
PUMP TUBE ALARM
INTERNAL BATTERY
DIAGNOSTICS
    
```

4. The controller briefly displays the current pump count information. Line one lists the pump counts since the last reset. Line two lists the current alarm setting. The screen will advance automatically.

```

1024000 PUMP COUNTS,
WARNING AT 1000000
    
```

5. To reset the counter to zero, select YES. Always reset the counter after replacing a pump tube. Select NO when merely checking the current count.

```

1024000 PUMP COUNTS,
RESET PUMP COUNTER?
YES NO
    
```

6. You can modify the pump tube alarm setting to tailor it to your needs. The factory default setting is 1,000,000 pump counts. However you may experience tube wear more or less frequently. Change the pump-count alarm setting by typing the first two digits of the new setting. The sampler accepts entries between 1 and 99. For example, to increase the count to 1,500,000, enter 15.

```

WARNING AT 1000000
PUMP COUNTS
(1 - 99)00000
    
```

A pump tube alarm should be set to notify you when the pump tube should be replaced. A pump tube should be replaced when it is beginning to show signs of wear, which is long before the tube wall fails. For more information, refer to section 9.4.






### 9.3.2 Pump Tube Alarm

The sampler tracks the pump counts in both the forward and reverse cycles with a resettable counter. When the counter reaches the default count of 1,000,000 or the count you have entered, the sampler displays a message, “WARNING! REPLACE PUMP TUBE.” The message appears each time you turn the sampler on or run a program until you reset the counter.

After replacing the pump tube (see section 9.4), reset the count to zero so the sampler can begin tallying the pump counts for the new tube. Replacing the pump tube does not reset the counter.

Experience may suggest a significantly different pump tube life. You can change the alarm count to represent the pump tube life for your application. The example “Resetting the Pump Tube Alarm” shows how to set the alarm count and reset the pump counts.

**Example: Resetting the Internal Battery Countdown Timer**

- From the Main Menu select OTHER FUNCTIONS.  

- Select MAINTENANCE.  

- Select INTERNAL BATTERY.  

- The controller briefly displays the battery expiration date. The screen will advance automatically.  

- The controller asks if you replaced the battery.  


If you are just checking the expiration date, select NO.  
If in fact you changed the internal battery (see section 9.7), select YES to reset the countdown.

Resetting the countdown timer directs the sampler to calculate a new expiration date, five years in the future.

If you reset the countdown without replacing the battery, the sampler will not be able to monitor the battery's impending expiration.







### 9.3.3 Internal Battery

A lithium battery housed inside the controller maintains power to the sampler's memory when the controller is disconnected from a power source. If it discharges completely, the sampler will lose all program settings and all data stored in memory when the external power source is disconnected. The battery requires replacement every five years.


The sampler keeps track of this five-year replacement date, which is calculated from the date the battery was installed.

The example *Resetting the Internal Battery Countdown Timer* shows how to view the expiration date and how to set a new date after replacing the internal battery. Refer to section 9.7 for battery replacement instructions.

### Example: Self-diagnostics

- From the Main Menu select OTHER FUNCTIONS.  

- Select MAINTENANCE.  

- Select DIAGNOSTICS  

- The controller starts the diagnostics routine by first testing the RAM (Random Access Memory). If the controller passes a memory test, it will advance to the next test. Otherwise it will report a failure and advise you to contact Teledyne ISCO.  

- The controller continues by testing the ROM (Read-only Memory).  

- The controller then fills the display, first with all pixels "on," then with characters.  


If the display is faulty (for example some pixels always on or always off), the display or its driver circuitry may need repair. Call Teledyne ISCO. The diagnostic routine automatically advances to the next test.

- Select YES to test the pump or NO to skip this test.  


*Continued...*

### 9.3.4 Diagnostics

The sampler has several self-diagnostic tests that check the memory (both the RAM and ROM), pump, and the distributor arm position. It also contains screens that let you reset the sampler's program settings and memory to factory settings.

You do not need to run the diagnostic tests on a routine basis. The diagnostic tests are a troubleshooting tool; Teledyne ISCO's Technical Service department staff may ask you to run these tests when they work with you in diagnosing problems with your sampler.

Access the diagnostic screens by selecting Other Functions from the main menu, then Maintenance, and Diagnostics. The unit will automatically test the memory and display; you can choose whether it tests the pump and distributor, and whether it reinitializes the unit to its original factory settings.



**Memory Tests** – Test the unit’s memory. If either screen shows you a message saying that the memory failed the test, contact Teledyne ISCO’s Technical Service.

*Self-diagnostics, continued...*

If you select YES, the controller tests the forward and reverse pumping operation.

PUMPING . . .

PUMPING . . .  
ON/OFF RATIO = 1.06

PURGING . . .

PURGING . . .  
ON/OFF RATIO = 1.08

The ratio should fall within the 0.80 to 1.25 range. Values outside this range indicate a pump problem; contact Teledyne ISCO Technical Service.

8. Select YES to test the distributor, NO to skip this test.

TEST DISTRIBUTOR?  
YES NO

If you select YES, the controller tests the distributor operation. *Note:* the controller must be mounted on the sampler assembly with a distributor arm attached.

FINDING ARM FLEXURE

After finding the arm flexure, the arm is positioned over 24 bottle positions.

NOW AT BOTTLE \_\_

9. The final screen of the diagnostics routine asks if you want to reinitialize the sampler controller. Select YES only if you want to restore all programs to the default settings and to clear all memory partitions. Select NO to retain the programs and data and to exit the diagnostics.

RE-INITIALIZE?  
YES NO

**Display Test** – Displays a test pattern on the screen. If the pattern does appear correctly, the display or its circuitry require repair. Some indications of a faulty display during this test would be pixels always on or off, or unreadable characters.

**Pump Test** – Runs a test of the pumping system. If selected, the controller first runs the pump forward for a short period. Next, it displays an ON/OFF ratio number. After displaying the forward pumping ratio, the sampler reverses the pump to obtain a purge ratio. The acceptable range is between 0.80 and 1.25. If either ratio is outside the acceptable range, the pump requires repair. You do not need to pump liquid while running the test, and you can run the test with or without a tube in the pump.

 **CAUTION**

Prolonged exposure to noise while the pump is operating could result in hearing loss. Tests indicate that the pump produces sound levels in excess of 85db at one meter. If you are operating the pump for long durations, always wear hearing protection.

**Distributor Test** – The distributor test is provided for factory personnel to verify the distributor’s position as it rotates through the 24 positions. The distributor arm flexes slightly when it contacts the stop, and the sampler measures this flexure. The sampler uses this measurement to position the arm accurately over each bottle.

**Reinitialize Controller** – The sampler allows you to reset all the program settings to the settings shipped with the sampler. Reinitializing also clears the memory that stores the sampler’s event log and module data. **Use this option cautiously.**

## 9.4 Replacing the Pump Tube

Replace the pump tube only with Teledyne ISCO’s BLZZRD/6712 pump tubing. **Other pump tubes will not work.** The BLZZRD/6712 pump tubing is easily recognized by the blue alignment collars.

Improper pump tubes include those made for Teledyne ISCO’s other samplers (3700, Glacier, etc.) and tubing from non-Teledyne ISCO vendors. Also note that the discharge tube is not the same as the pump tube. You could experience several problems if you install the wrong pump tubing:

- The sampler will not pump the liquid.
- Pump jams
- Inaccurate sample volumes
- Faulty liquid detection

Inspect the pump tube periodically. Replace the tube when it cracks or appears worn. Inspect the tube frequently when the sample liquid contains a high percentage of suspended or abrasive solids.

 **Note**

The importance of regular tubing replacement cannot be overstated. The key is to replace the tube before failure, not after. When a pump tube ruptures, grit and other abrasive debris can be driven into the pump shaft seal. Over time, this abrasive material will degrade the pump seal, jeopardizing the NEMA 4x 6 rating of the controller.

Failure to maintain the pump tube may result in permanent damage to the sampler. Check the condition of the pump tube regularly and if the tube shows signs of fatigue or wear, replace it immediately. A properly maintained sampler will provide years of the reliable service that is expected of a Teledyne ISCO Sampler.

### Checklist For Replacing Pump Tube:

1. Disconnect power from the sampler.
2. Loosen the liquid detector's cover by unscrewing the black knob on top of the detector. Unlatch the pump band. (The band is the rounded metal cover that holds the tube in the pump.)
3. Pull the tube away from the bulkhead fitting. Pull it from the pump and detector.
4. Clean the pump rollers, the inside of the pump band, and the two drain holes at the bottom-right side of the pump housing.
5. Thread the new tubing through the pump so that the tube follows its natural curve around the pump rollers. You may need to move the pump rollers to make the installation easier.
6. The blue collars on the tube help align the tube in the detector and pump. Align the tube by placing the collars in the grooves inside the liquid detector.
7. Close the detector's cover and tighten the black knob. Close the pump strap and latch it.
8. Reset the pump-tube counter. (See the "Resetting the Pump Tube Alarm" example.)
9. Take a "dry" grab sample to test the tube installation.

 <b>WARNING</b>
--

**The pump's safety lock prevents the sampler from running the pump when the pump band is open. DO NOT tamper with the safety mechanism. The pump is extremely powerful. The pump rollers can injure you severely if the sampler activates the pump while you are working on it or inside it. Disconnect power from the sampler before replacing the pump tube.**


#### 9.4.1 Pump Tube Life

Several factors shorten the pump tube life. Among them are:


- Improper installation.
- Abrasive materials suspended in sample liquid.
- Frequent line rinses.
- Long purge cycles, such as those used with long suction lines.

To extend the life of the pump tubes:

- Always use Teledyne ISCO pump tubes.
- Install the tube properly, aligning the blue collars correctly in the liquid detector's grooves.
- Follow the natural curve of the tube when wrapping the tube around the pump rollers.
- Minimize the line rinses and sampling retries in the sampling programs
- Use the shortest possible suction line.



*Figure 9-2 Removing and replacing the pump tube*



*Figure 9-3 Placing the pump tube in the liquid detector*

## 9.5 Opening the Controller Case

You will need to open the controller case to gain access to:

- The desiccant bag that dries the controller's interior
- The battery that provides backup power to the controller.

### CAUTION

The circuit boards can be damaged from a discharge of static electricity. To protect the circuit boards, only open the controller at an antistatic workstation.

#### To open the case:

1. Pull the discharge tube and pump tube away from the bulkhead fitting.
2. If configured for multiple bottles, remove the distributor arm by unscrewing the nut that attaches the arm to the distributor shaft. If configured for 2.5 gallon composite bottles, remove the composite tube guide.

### CAUTION

Distributor drive gears may be permanently damaged if you attempt to rotate the distributor arm by hand or with any tool. Only move the distributor arm through the Manual Functions.

3. Remove the controller from refrigerator assembly by removing four screws. See Figure 9-4.




Figure 9-4 Controller mounting screw (1 of 4)

4. Open the case by removing the 11 screws that attach the case bottom to the case top.

### Note

The case top and bottom fit together with a sealed tongue-and-groove joint. Any damage—nicks or cuts—to the tongue, groove, or sealing gasket prevent the case from sealing completely when you close it. Use extreme care when opening the case to avoid damaging the joining surfaces.

## 9.6 Replacing the Desiccant

A humidity indicator, labeled “INTERNAL CASE HUMIDITY,” is located to the right of the keypad. It indicates the amount of moisture present inside the control box. The paper indicator is **blue** when the control box is dry.

The control box is a completely sealed unit. You do not need to open the control box during normal operation. The 6712 Samplers use a desiccant bag inside the controller case to prevent moisture damage to its electronic components. Thus, the humidity indicator should remain blue under normal conditions. If moisture does accumulate, the numbered areas on the indicator will turn light **pink** or **white**, starting with the area numbered “20.” This indicates that the relative humidity inside the control box exceeds 20%. As more moisture accumulates, the areas numbered “30” and “40” will turn light pink or white, indicating relative humidities of 30% and 40% respectively.

If the 30% area of the humidity indicator turns light pink or white, open the control unit, inspect it for leaks, and replace the desiccant bag.

### CAUTION

Do not attempt to reactivate the used desiccant. Reactivation may cause desiccant to leak out of the bag, causing mechanical damage inside the controller.

Follow the instructions in Section 9.5 to open the controller. Remove the bag from the cardboard box at the front of the case.




Figure 9-5 Desiccant location




Figure 9-6 Removing the desiccant bag

### 9.7 Replacing the Internal Battery

A lithium battery housed inside the controller maintains power to the sampler's memory when the controller is disconnected from a power source. If it discharges completely, the sampler will lose all program settings and all data stored in memory when the external power source is disconnected. The battery requires replacement every five years. You can check this replacement date by viewing the sampler's maintenance screens (see the example *Resetting the Internal Battery Countdown Timer*).

While replacing the battery, it is likely that the stored readings will be lost and your program will revert to the default program. Before proceeding, retrieve the stored data and record your program settings.

#### Note

Teledyne ISCO recommends you replace the battery every five years. To ensure that the replacement meets Teledyne ISCO's specifications, use only Teledyne ISCO replacement batteries. (See Appendix D, *Replacement Parts*.)

To replace the internal battery:

1. Open the controller case by following the instructions in Section 9.5. Separate the case bottom and top by disconnecting the four connectors.
2. Lift the desiccant box from the case.
3. Remove the grounding strap by loosening the  $\frac{1}{4}$ -inch stop nut (Figure 9-7).
4. Disconnect the remaining connectors from the main circuit board (Figure 9-7).




Figure 9-7 Main Circuit Board Connections

5. Remove the four screws attaching the main circuit board to the case.
6. Detach the main circuit board from the keypad connector by pulling gently upward on the edges of the board next to the connector. Pull the main board up until the keypad connector is almost even with the edge of the case. Next, detach the main circuit board from the connector board.
7. Turn the board over and locate the lithium battery on the component side (Figure 9-8). Remove the lithium battery, using care to prevent damage to the circuit board.
8. De-solder the cylindrical battery or the coin-style battery. Insert the replacement battery and solder it in place.

**⚠ WARNING**

**Danger of explosion if lithium battery is incorrectly replaced. Replace with the same type, ensuring correct polarity. Do not dispose of used lithium battery in fire. Dispose of battery in accordance with local laws or authority.**






Figure 9-8 Lithium battery

9. Reconnect the main circuit board to the connector board before reconnecting it to the keypad. Replace the four mounting screws, and reconnect the connectors.
10. Replace the grounding strap and attach with the  $\frac{1}{4}$  inch stop nut. Reconnect the four connectors that connect the case bottom assembly.
11. Ensure that all connections and hardware are secure. Replace the desiccant and close the case.
12. Mount the case on the refrigerator assembly.
13. Apply power and turn the BLZZRD on.
14. Reset the Battery Countdown. See the example *Resetting the Internal Battery Countdown Timer*.

## 9.8 Error Messages

If the sampler detects an error that prevents it from taking a sample or continuing the program, it displays an error message.

There are messages for the following conditions:

- Pump Jammed
- Distributor Jammed
- Probable Overflow
- Power Failed
- Low Battery
- User Stopped
- No Distributor Arm
- No Liquid Detected
- No More Liquid
- Pump Latch Open
- Sampler Shut Off
- Bottle Full
- Sample In Progress

Since the sampler may encounter more than one error during a program, it does not display the messages continuously. Instead, it records each error and the time it occurred in the Sampling Results and Combined Results report. It alerts you to the recorded errors by displaying the message, “ERRORS HAVE OCCURRED DURING PROGRAM.”



ERRORS HAVE OCCURED  
DURING PROGRAM

If the sampler detects a low battery condition, it stops the running program and displays the “LOW BATTERY” error message. A low battery condition exists when five consecutive sample events have resulted in a POWER FAILED error message.

### 9.8.1 Fatal Error Message

The Fatal Error message will only show up if you are using a multiple bottle configuration and the distributor system fails. Failing the test can indicate any of the following conditions:

- The distributor arm is not attached to the sampler.
- The distributor drive mechanism inside the controller is malfunctioning.
- The controller is not securely mounted to the refrigeration compartment.

### 9.8.2 Pump Warning Message

Occasionally, the sampler displays a pump warning message.



WARNING: REPLACE  
PUMP TUBE

This warning reminds you to check the condition of the pump tube and to replace it if necessary. Refer to section 9.4 for replacement instructions.

# BLZZRD™ Portable Refrigerated Sampler

## Appendix A Menu Flowcharts




Figure A-1 Menu Tree for Standard Programming





Figure A-2 Standard Programming: Programming Screens



**Notes**

B. Appears for standard flow-paced programs.	START FLOW COUNT AFTER A --- MINUTE DELAY (1-999)
C. Appears for extended programs.	START "EXTENDED 1" AFTER A --- MINUTE DELAY (1-999)
D. Appears for standard flow-paced programs.	START FLOW COUNT AT HH:MM
E. Main menu for extended programming.	RUN "EXTENDED 1" PROGRAM VIEW REPORT OTHER FUNCTIONS

Figure A-3 Standard and Extended Programming: Start Times

Units selected screens are for a 730 Bubbler Flow Module. Screens for other modules may vary.




Figure A-4 Standard Programming: Quick View; Programming Screens




Figure A-5 Standard Programming: Quick View; Start Times




Figure A-6 View Report






Figure A-7 Maintenance Screens




Figure A-8 Manual Functions Screens




Figure A-9 Extended Programming: Programming Screens




Figure A-10 Extended Programming: Equipment Setup




Figure A-11 Extended Programming: Pacing and Distribution




Figure A-12 Extended Programming: Event Conditions




Figure A-13 Extended Programming: Programmed Sampler Enable




Figure A-14 Extended Programming: Quick View; Programming Screens






Figure A-15 Extended Programming: Quick View; Equipment Set-Up




Figure A-16 Extended Programming: Quick View; Pacing and Distribution




Figure A-17 Extended Programming: Quick View; Programmed Sampler Enable




Figure A-18 Extended Programming: Normal View; Software Options





Figure A-19 Extended Programming: Quick View; Software Options



**Notes**

A. <- denotes input signal.  
 ->denotes output signal.

Figure A-20 Extended Programming: Normal View; Hardware





Figure A-21 Extended Programming: Quick View; Hardware



Proceed to calibration and validation screens  
 and remaining Hardware screens.

Figure A-22 SDI-12 Sonde Screens



# BLZZRD™

## Portable Refrigerated Sampler

---

### *Appendix B Material Safety Data Sheets*

This appendix provides Material Safety Data Sheets for the desiccant used by the BLZZRD Sampler.

Teledyne ISCO cannot guarantee the accuracy of the data. Specific questions regarding the use and handling of the products should be directed to the manufacturer listed on the MSDS.

# Material Safety Data Sheet

Indicating Silica Gel

Identity (Trade Name as Used on Label)

Manufacturer : Address: Phone Number (For Information): Emergency Phone Number:	MULTISORB TECHNOLOGIES, INC. (formerly Multiform Desiccants, Inc.) 325 Harlem Road Buffalo, NY 14224 716/824-8900 716/824-8900	MSDS Number* : M75 CAS Number* : Date Prepared: July 6, 2000 Prepared By* : G.E. McKedy
--	---	--

## Section 1 - Material Identification and Information

Components - Chemical Name & Common Names (Hazardous Components 1% or greater; Carcinogens 0.1% or greater)	%*	OSHA PEL	ACGIH TLV	OTHER LIMITS RECOMMENDED
Silica Gel SiO <sub>2</sub>	98.0	6mg/m <sup>3</sup> (total dust)	10mg/m <sup>3</sup> (total dust)	
Cobalt Chloride	>2.0	0.05mg/m <sup>3</sup> (TWA cobalt metal dust & fume)	.05mg/m <sup>3</sup> (Cobalt, TWA)	
Non-Hazardous Ingredients				
<b>TOTAL</b>	<b>100</b>			

## Section 2 - Physical/Chemical Characteristics

Boiling Point	N/A	Specific Gravity (H <sub>2</sub> O = 1)	2.1
Vapor Pressure (mm Hg and Temperature)	N/A	Melting Point	N/A
Vapor Density (Air =1)	N/A	Evaporation Rate (=1)	N/A
Solubility in Water	Insoluble, but will adsorb moisture.	Water Reactive	Not reactive, but will adsorb moisture.
Appearance and Odor	Purple crystals, no odor.		

## Section 3 - Fire and Explosion Hazard Data

Flash Point and Methods Used	N/A	Auto-Ignition Temperature	N/A	Flammability Limits in Air % by Volume	N/A	LEL	UEL
Extinguisher Media	Dry chemical, carbon dioxide and foam can be used.						
Special Fire Fighting Procedures	Water will generate heat due to the silica gel which will adsorb water and liberate heat.						
Unusual Fire and Explosion Hazards	When exposed to water, the silica gel can get hot enough to reach the boiling point of water. Flooding with water will reduce the temperature to safe limits.						

## Section 4 - Reactivity Hazard Data

<b>STABILITY</b> <input type="checkbox"/> Stable <input type="checkbox"/> Unstable	Conditions To Avoid	Moisture and high humidity environments.
Incompatibility (Materials to Avoid)	Water.	
Hazardous Decomposition Products	Carbon dioxide, carbon monoxide, water	
<b>HAZARDOUS POLYMERIZATION</b> <input type="checkbox"/> May Occur	Conditions To Avoid	None.

\*Optional

Indicating Silica Gel

Will Not Occur

Page 2

**Section 5 - Health Hazard Data**

<b>PRIMARY ROUTES OF ENTRY</b>	<input type="checkbox"/> Inhalation <input type="checkbox"/> Ingestion <input type="checkbox"/> Skin Absorption <input type="checkbox"/> Not Hazardous	<b>CARCINOGEN LISTED IN</b>	<input type="checkbox"/> NTP <input type="checkbox"/> IARC Monograph	<input type="checkbox"/> OSHA <input type="checkbox"/> Not Listed
<b>HEALTH HAZARDS</b>	Acute    May cause eye, skin and mucous membrane irritation. Chronic    Prolonged inhalation may cause lung damage.			
Signs and Symptoms of Exposure	Drying and irritation.			
Medical Conditions	Asthma.			
Generally Aggravated by Exposure				
<b>EMERGENCY FIRST AID PROCEDURES</b> - Seek medical assistance for further treatment, observation and support if necessary.				
Eye Contact	Flush with water for at least 15 minutes.			
Skin Contact	Wash affected area with soap and water.			
Inhalation	Remove affected person to fresh air.			
Ingestion	Drink at least 2 glasses of water.			

**Section 6 - Control and Protective Measures**

Respiratory Protection (Specify Type)	Use NIOSH approved dust mask or respirator.		
Protective Gloves	Light cotton gloves.	Eye Protection	Safety glasses.
<b>VENTILATION TO BE USED</b>	<input type="checkbox"/> Local Exhaust	<input type="checkbox"/> Mechanical (General)	<input type="checkbox"/> Special
	<input type="checkbox"/> Other (Specify)		
Other Protective Clothing and Equipment	None.		
Hygienic Work Practices	Avoid raising dust. Avoid contact with skin, eyes and clothing.		

**Section 7 - Precautions for Safe Handling and Use/Leak Procedures**

Steps to be Taken if Material Is Spilled Or Released	Sweep or vacuum up and place the spilled material in a waste disposal container. Avoid raising dust.		
Waste Disposal Methods	Dispose in an approved landfill according to federal, state and local regulations.		
Precautions to be Taken In Handling and Storage	Cover promptly to avoid blowing dust. Wash after handling.		
Other Precautions and/or Special Hazards	Keep in sealed containers away from moisture. The silica gel will readily adsorb moisture.		

\*Optional

Indicating Silica Gel

101 Christine Drive  
Belen, New Mexico 87002  
Phone: (505) 864-6691  
Fax: (505) 861-2355

**SÜD-CHEMIE**  
Performance Packaging



ISO 9002

MATERIAL SAFETY DATA SHEET -- September 28, 1998  
SORB-IT®  
Packaged Desiccant

### SECTION I -- PRODUCT IDENTIFICATION

<b>Trade Name and Synonyms:</b>	Silica Gel, Synthetic Amorphous Silica, Silicon, Dioxide
<b>Chemical Family:</b>	Synthetic Amorphous Silica
<b>Formula:</b>	SiO <sub>2</sub> .x H <sub>2</sub> O

### SECTION II -- HAZARDOUS INGREDIENTS

Components in the Solid Mixture

COMPONENT	CAS No	%	ACGIH/TLV (PPM)	OSHA-(PEL)
Amorphous Silica	63231-67-4	>99	PEL - 20 (RESPIRABLE), TLV - 5	LIMIT - NONE, HAZARD - IRRITANT "

Synthetic amorphous silica is not to be confused with crystalline silica such as quartz, cristobalite or tridymite or with diatomaceous earth or other naturally occurring forms of amorphous silica that frequently contain crystalline forms.

This product is in granular form and packed in bags for use as a desiccant. Therefore, no exposure to the product is anticipated under normal use of this product. Avoid inhaling desiccant dust.

### SECTION III -- PHYSICAL DATA

<b>Appearance and Odor:</b>	White granules; odorless.
<b>Melting Point:</b>	>1600 Deg C; >2900 Deg F
<b>Solubility in Water:</b>	Insoluble.
<b>Bulk Density:</b>	>40 lbs./cu. ft.
<b>Percent Volatile by Weight @ 1750 Deg F:</b>	<10%.

101 Christine Drive  
Belen, New Mexico 87002  
Phone: (505) 864-6691  
Fax: (505) 861-2355



---

MATERIAL SAFETY DATA SHEET -- September 28, 1998  
SORB-IT®  
Packaged Desiccant

---

#### SECTION IV -- FIRE EXPLOSION DATA

**Fire and Explosion Hazard** - Negligible fire and explosion hazard when exposed to heat or flame by reaction with incompatible substances.

**Flash Point** - Nonflammable.

**Firefighting Media** - Dry chemical, water spray, or foam. For larger fires, use water spray fog or foam.

**Firefighting** - Nonflammable solids, liquids, or gases: Cool containers that are exposed to flames with water from the side until well after fire is out. For massive fire in enclosed area, use unmanned hose holder or monitor nozzles; if this is impossible, withdraw from area and let fire burn. Withdraw immediately in case of rising sound from venting safety device or any discoloration of the tank due to fire.

#### SECTION V -- HEALTH HAZARD DATA

Health hazards may arise from inhalation, ingestion, and/or contact with the skin and/or eyes. Ingestion may result in damage to throat and esophagus and/or gastrointestinal disorders. Inhalation may cause burning to the upper respiratory tract and/or temporary or permanent lung damage. Prolonged or repeated contact with the skin, in absence of proper hygiene, may cause dryness, irritation, and/or dermatitis. Contact with eye tissue may result in irritation, burns, or conjunctivitis.

**First Aid (Inhalation)** - Remove to fresh air immediately. If breathing has stopped, give artificial respiration. Keep affected person warm and at rest. Get medical attention immediately.

**First Aid (Ingestion)** - If large amounts have been ingested, give emetics to cause vomiting. Stomach siphon may be applied as well. Milk and fatty acids should be avoided. Get medical attention immediately.

**First Aid (Eyes)** - Wash eyes immediately and carefully for 30 minutes with running water.

101 Christine Drive  
Belen, New Mexico 87002  
Phone: (505) 864-6691  
Fax: (505) 861-2355

**SÜD-CHEMIE**  
Performance Packaging



ISO 9002

---

MATERIAL SAFETY DATA SHEET -- September 28, 1998  
SORB-IT®  
Packaged Desiccant

---

**NOTE TO PHYSICIAN:** This product is a desiccant and generates heat as it adsorbs water. The used product can contain material of hazardous nature. Identify that material and treat accordingly.

#### SECTION VI -- REACTIVITY DATA

**Reactivity** - Silica gel is stable under normal temperatures and pressures in sealed containers. Moisture can cause a rise in temperature which may result in a burn.

#### SECTION VII -- SPILL OR LEAK PROCEDURES

Notify safety personnel of spills or leaks. Clean-up personnel need protection against inhalation of dusts or fumes. Eye protection is required. Vacuuming and/or wet methods of cleanup are preferred. Place in appropriate containers for disposal, keeping airborne particulates at a minimum.

#### SECTION VIII -- SPECIAL PROTECTION INFORMATION

**Respiratory Protection** - Provide a NIOSH/MSHA jointly approved respirator in the absence of proper environmental control. Contact your safety equipment supplier for proper mask type.

**Ventilation** - Provide general and/or local exhaust ventilation to keep exposures below the TLV. Ventilation used must be designed to prevent spots of dust accumulation or recycling of dusts.

**Protective Clothing** - Wear protective clothing, including long sleeves and gloves, to prevent repeated or prolonged skin contact.

**Eye Protection** - Chemical splash goggles designed in compliance with OSHA regulations are recommended. Consult your safety equipment supplier.

101 Christine Drive  
Belen, New Mexico 87002  
Phone: (505) 864-6691  
Fax: (505) 861-2355

**SÜD-CHEMIE**  
Performance Packaging



ISO 9002

---

MATERIAL SAFETY DATA SHEET -- September 28, 1998  
SORB-IT®  
Packaged Desiccant

---

**SECTION IX -- SPECIAL PRECAUTIONS**

Avoid breathing dust and prolonged contact with skin. Silica gel dust causes eye irritation and breathing dust may be harmful.

\* No Information Available

HMIS (Hazardous Materials Identification System) for this product is as follows:

Health Hazard	0
Flammability	0
Reactivity	0
Personal Protection	HMIS assigns choice of personal protective equipment to the customer, as the raw material supplier is unfamiliar with the condition of use.

---

The information contained herein is based upon data considered true and accurate. However, United Desiccants makes no warranties expressed or implied, as to the accuracy or adequacy of the information contained herein or the results to be obtained from the use thereof. This information is offered solely for the user's consideration, investigation and verification. Since the use and conditions of use of this information and the material described herein are not within the control of United Desiccants, United Desiccants assumes no responsibility for injury to the user or third persons. The material described herein is sold only pursuant to United Desiccants' Terms and Conditions of Sale, including those limiting warranties and remedies contained therein. It is the responsibility of the user to determine whether any use of the data and information is in accordance with applicable federal, state or local laws and regulations.





# BLZZRD™

## Portable Refrigerated Sampler

---

### *Appendix C General Safety Procedures*

The safety of the personnel who use the BLZZRD sampler is a critical consideration. The following procedures, applicable to working in and around manholes and sewers, are those used by Black & Veatch, a respected consulting firm, and are published here by their permission.

Field personnel must keep safety uppermost in their minds at all times. When working above ground, rules of common sense and safety prevail. However, when entering manholes, strict safety procedures must be observed. Failure to do so could jeopardize not only your own life, but also the lives of other crew members.

 **WARNING**

**The BLZZRD has not been approved for use in hazardous locations as defined by the National Electrical Code.**

 **CAUTION**

Before you install any sampler, you must take the proper safety precautions. The following discussion of safety procedures offers only general guidelines. Each situation in which you install a sampler varies. You must take into account the individual circumstances of each installation.

Additional safety considerations, other than those discussed here, may be required. Check applicable guidelines, codes, and regulations of federal, state, city, and county agencies.

#### **C.1 Hazards**

There are many hazards connected with entering manholes. Some of the most common hazards are:

- *Adverse Atmosphere*—The manhole may contain flammable or poisonous gases or the atmosphere may be deficient in oxygen. Forced ventilation may be necessary.
- *Deteriorated Rungs*—Manhole steps may be corroded and not strong enough to support a man. It may be difficult to inspect the rungs because of poor lighting.
- *Traffic*—Whenever manholes are located in the traveled way, barricades and warning devices are essential to direct traffic away from an open manhole.
- *Falling Objects*—Items placed near the manhole opening may fall and injure a worker in the manhole. All loose items should be kept away from the manhole opening.

This applies to hand tools as well as stones, gravel and other objects.

- *Sharp Edges*—Sharp edges of items in or near a manhole may cause cuts or bruises.
- *Lifting Injuries*—Unless proper tools are used to remove manhole covers, back injuries or injuries to hands or feet may result.

## C.2 Planning

Advance planning should include arrangements for test equipment, tools, ventilating equipment, protective clothing, traffic warning devices, ladders, safety harness, and adequate number of personnel. Hasty actions may result in serious injuries. Time spent in the manhole should be kept to a minimum.

## C.3 Adverse Atmospheres

[Refer to the table of Hazardous Gases at the end of this appendix.] Before workers enter a manhole, tests should be made for explosive atmosphere, presence of hydrogen sulfide, and oxygen deficiency. Combustible or toxic vapors may be heavier than air, so the tests on the atmosphere must be run at least  $\frac{3}{4}$  of the way down the manhole.

Whenever adverse atmosphere is encountered, forced ventilation must be used to create safe conditions. After the ventilating equipment has been operated for a few minutes, the atmosphere in the manhole should be retested before anyone enters the manhole.

When explosive conditions are encountered, the ventilating blower should be placed upwind to prevent igniting any gas that is emerging from the opening. When a gasoline engine blower is used, it must be located so that exhaust fumes cannot enter the manhole.

If testing equipment is not available, the manhole should be assumed to contain an unsafe atmosphere and forced ventilation must be provided. It should never be assumed that a manhole is safe just because there is no odor or the manhole has been entered previously.

## C.4 Entering Manholes

Since the top of the manhole is usually flush with the surrounding surface, there may not be anything for the person who is entering the manhole to grab on to steady himself. Persons who are entering manholes should not be permitted to carry anything in their hands as they enter the manhole, to ensure that their hands will be free to hold on or grab if they slip. A good method for entering a manhole is to sit on the surface facing the manhole steps or ladder, with the feet in the hole and the arms straddling the opening for support. As the body slides forward and downward, the feet can engage a rung, and the back can rest against the opposite side of the opening. If there is any doubt about the soundness of the manhole steps, a portable ladder should be used.

A person should never enter a manhole unless he is wearing personal safety equipment, including a safety harness and a hard hat. Two persons should be stationed at the surface continuously while anyone is working inside a manhole, to lift him out if he is overcome or injured. One man cannot lift an unconscious man out of a manhole.

The persons stationed at the surface should also function as guards to keep people and vehicles away from the manhole opening. To avoid a serious injury, a person should not be lifted out of manhole by his arm unless it is a dire emergency.

When more than one person must enter a manhole, the first person should reach the bottom and step off the ladder before the next one starts down. When two men climb at the same time, the upper one can cause the lower one to fall by slipping or stepping on his fingers.

#### **C.4.1 Traffic Protection**

In addition to traffic cones, markers, warning signs, and barricades, a vehicle or a heavy piece of equipment should be placed between the working area and oncoming traffic. Flashing warning signals should be used to alert drivers and pedestrians. Orange safety vests should be worn by personnel stationed at the surface when the manhole is located in a vehicular traffic area.

#### **C.4.2 Removing the Covers**

Manhole covers should be removed with a properly designed hook. Use of a pick ax, screwdriver, or small pry bar may result in injury. A suitable tool can be made from  $\frac{3}{4}$ -inch round or hex stock. Two inches of one end should be bent at a right angle and the other end should be formed into a D-handle wide enough to accommodate both hands. Even with this tool, care must be exercised to prevent the cover from being dropped on the toes. The 2-inch projection should be inserted into one of the holes in the cover, the handle grasped with both hands, and the cover lifted by straightening the legs which have been slightly bent at the knees.

#### **C.4.3 Other Precautions**

Other precautions which should be taken when entering a manhole are:

- Wear a hard hat.
- Wear coveralls or removable outer garment that can be readily removed when the work is completed.
- Wear boots or nonsparking safety shoes.
- Wear rubberized or waterproof gloves.
- Wear a safety harness with a stout rope attached.
- Do not smoke.
- Avoid touching yourself above the collar until you have cleaned your hands.

#### **C.4.4 Emergencies**

Every member of the crew should be instructed on procedures to be followed in cases of an emergency. It is the duty of each crew chief to have a list of emergency phone numbers, including the nearest hospital and ambulance service, police precinct, fire station, and rescue or general emergency number.

#### **C.4.5 Field Equipment**

The following equipment should be available for use:

Blowers	Gloves
Breathing apparatus	Hard Hats
Coverall	Harnesses
First aid kits	Manhole irons
Emergency flashers	Pick axes
Flashlight	Rain slickers
Mirror	Ropes
Gas detectors	Safety vests
Gas masks	Traffic cones
Waders	

#### **C.5 Lethal Atmospheres in Sewers**

The following is an article written by Dr. Richard D. Pomeroy, and published in the October 1980 issue of "Deeds & Data" of the WPCF. Dr. Pomeroy is particularly well known for his studies, over a period of nearly 50 years, in the field of the control of hydrogen sulfide and other odors in sewers and treatment plants. He has personally worked in a great many functioning sewers. In the earlier years he did so, he admits, with little knowledge of the grave hazards to which he exposed himself.

"It is gratifying that the subject of hazards to people working in sewers is receiving much more attention than in past years, and good safety procedures are prescribed in various publications on this subject. It is essential that people know and use correct procedures.

"It is less important to know just what the hazardous components of sewer atmospheres are, as safety precautions should in general be broadly applicable, but there should be a reasonable understanding of this subject. It is disturbing to see statements in print that do not reflect true conditions.

"One of the most common errors is the assumption that people have died from a lack of oxygen. The human body is able to function very well with substantially reduced oxygen concentrations. No one worries about going to Santa Fe, New Mexico, (elev. 2,100 meters), where the partial pressure of oxygen is equal to 16.2% (a normal atmosphere is about 21%) oxygen. "When first going there, a person may experience a little 'shortness of breath' following exercise.

"People in good health are not afraid to drive over the high passes in the Rocky Mountains. At Loveland Pass, oxygen pressure is 13.2% of a normal atmosphere. At the top of Mt.

Whitney, oxygen is equal to 12.2%. Many hikers go there, and to higher peaks as well. After adequate acclimation, they may climb to the top of Mt. Everest, where oxygen is equal to only 6.7%.

“The lowest oxygen concentrations that I have observed in a sewer atmosphere was 13 percent. It was in a sealed chamber, near sea level, upstream from an inverted siphon on a metropolitan trunk. A man would be foolish to enter the chamber. Without ventilation, he might die, but not from lack of oxygen.

“It seems unlikely that anyone has ever died in a sewer from suffocation, that is, a lack of oxygen. Deaths have often been attributed to ‘asphyxiation.’ This is a word which, according to the dictionary, is used to mean death from an atmosphere that does not support life. The word has sometimes been misinterpreted as meaning suffocation, which is only one kind of asphyxiation.

“In nearly all cases of death in sewers, the real killer is hydrogen sulfide. It is important that this fact be recognized. Many cities diligently test for explosive gases, which is very important, and they may measure the oxygen concentration which usually is unimportant, but they rarely measure H<sub>2</sub>S. Death has occurred where it is unlikely that there was any measurable reduction in the oxygen concentration. Waste water containing 2 mg per liter of dissolved sulfide, and at a pH of 7.0, can produce, in a chamber with high turbulence, a concentration of 300 PPM H<sub>2</sub>S, in the air. This is considered to be a lethal concentration. Many people have died from H<sub>2</sub>S, not only in sewers and industries, but also from swamps and from hot springs. In one resort area, at least five persons died from H<sub>2</sub>S poisoning before the people were ready to admit that H<sub>2</sub>S is not a therapeutic agent. Hardly a year passes in the US. without a sewer fatality from H<sub>2</sub>S as well as deaths elsewhere in the world.

“The presence of H<sub>2</sub>S in a sewer atmosphere is easily determined. A bellows-and-ampoule type of tester is very satisfactory for the purpose, even though it is only crudely quantitative. When using a tester of this type, do not bring the air to the ampoule by way of a tube, as this may change the H<sub>2</sub>S concentration. Hang the ampoule in the air to be tested, with a suction tube to the bulb or bellows.

“Lead acetate paper is very useful as a qualitative indicator. It cannot be used to estimate the amount of sulfide, but it will quickly turn black in an atmosphere containing only a tenth of a lethal concentration.

“Electrodes or other similar electrical indicating devices for H<sub>2</sub>S in air have been marketed. Some of them are known to be unreliable, and we know of none that have proved dependable. Do not use one unless you check it at frequent intervals against air containing known H<sub>2</sub>S concentra-

tions. A supposed safety device that is unreliable is worse than none at all.

“Remember that the nose fails, too, when it comes to sensing dangerous concentrations of H<sub>2</sub>S.

“Various other toxic gases have been mentioned in some publications. It is unlikely that any person has been asphyxiated in a sewer by any of those other gases, except possibly chlorine. The vapor of gasoline and other hydrocarbons is sometimes present in amounts that could cause discomfort and illness, but under that condition, the explosion hazard would be far more serious. The explosimeter tests, as well as the sense of smell, would warn of the danger. Pipelines in chemical plants might contain any number of harmful vapors. They, too, are sensed by smell and explosimeter tests if they get into the public sewer. Such occurrences are rare.

“The attempt to instill a sense of urgency about real hazards is diluted if a man is told to give attention to a long list of things that in fact are irrelevant.

“Be very careful to avoid high H<sub>2</sub>S concentrations, flammable atmospheres, and hazards of physical injuries. Remember that much H<sub>2</sub>S may be released by the stirring up of sludge in the bottom of a structure. Obey your senses in respect to irritating gases, such as chlorine (unconsciousness comes suddenly from breathing too much). Be cautious about strange odors. Do not determine percent oxygen in the air. There is a danger that the result will influence a man's thinking about the seriousness of the real hazards. Most important, use ample ventilation, and do not enter a potentially hazardous structure except in a good safety harness with two men at the top who can lift you out.”

**Table C-1 Hazardous Gases**

Gas	Chemical Formula	Common Properties	Specific Gravity or Vapor Density Air = 1	Physiological Effect*	Max Safe 60 Min. Exposure ppm	Max. Safe 8 Hour Exposure ppm	Explosive Range (% by vol. in air.) Limits lower/upper	Likely Location of Highest Concentration	Most Common Sources	Simplest and Cheapest Safe Method of Testing
Ammonia	NH <sub>3</sub>	Irritant and poisonous. Colorless with characteristic odor.	0.60	Causes throat and eye irritation at 0.05%, coughing at 0.17%. Short exposure at 0.5% to 1% fatal.	300 to 500	85	16 to 25	Near top. Concentrates in closed upper spaces	Sewers, chemical feed rooms	Detectable odor at low concentrations
Benzene	C <sub>6</sub> H <sub>6</sub>	Irritant, colorless anesthetic	2.77	Slight symptoms after several hours exposure at 0.16% to 0.32%. 2% rapidly fatal.	3,000 to 5,000	25	1.3 to 7.1	At bottom	Industrial wastes, varnish, solvents	Combustible gas indicator
Carbon Bisulfide	CS <sub>2</sub>	Nearly odorless when pure, colorless, anesthetic. Poisonous.	2.64	Very poisonous, irritating, vomiting, convulsions, psychic disturbance.	—	15	1.3 to 44.0	At bottom	An insecticide	Combustible gas indicator
Carbon Dioxide	CO <sub>2</sub>	Asphyxiant, Colorless, odorless. When breathed in large quantities, may cause acid taste. Non-flammable. Not generally present in dangerous amounts unless an oxygen deficiency exists.	1.53	Cannot be endured at 10% more than a few minutes, even if subject is at rest and oxygen content is normal. Acts on respiratory nerves.	40,000 to 60,000	5,000	— to —	At bottom; when heated may stratify at points above bottom.	Products of combustion, sewer gas, sludge. Also issues from carbonaceous strata.	Oxygen deficiency indicator
Carbon Monoxide	CO	Chemical asphyxiant. Colorless, odorless, tasteless. Flammable. Poisonous.	0.97	Combines with hemoglobin of blood. Unconsciousness in 30 min. at 0.2% to 0.25%. Fatal in 4 hours at 0.1%. Headache in few hours at 0.02%.	400	50	12.5 to 74.0	Near top, especially if present with illuminating gas.	Manufactured gas, flue gas, products of combustion, motor exhausts. Fires of almost any kind.	CO ampoules
Carbon Tetrachloride	CCl <sub>4</sub>	Heavy, ethereal odor.	5.3	Intestinal upset, loss of consciousness, possible renal damage, respiratory failure.	1,000 to 1,500	100	— to —	At bottom.	Industrial wastes, solvent, cleaning	Detectable odor at low concentrations
Chlorine	Cl <sub>2</sub>	Irritant. Yellow-green color. Choking odor detectable in very low concentrations. Non-flammable.	2.49	Irritates respiratory tract. Kills most animals in a very short time at 0.1%.	4	1	— to —	At bottom.	Chlorine cylinder and feed line leaks	Detectable odor at low concentrations
Formaldehyde	CH <sub>2</sub> O	Colorless, pungent suffocating odor.	1.07	Irritating to the nose.	—	10	7.0 to 73.0	Near bottom	Incomplete combustion of organics. Common air pollutant, fungicide.	Detectable odor
Gasoline	C <sub>8</sub> H <sub>12</sub> to C <sub>9</sub> H <sub>20</sub>	Volatile solvent. Colorless. Odor noticeable at 0.03%. Flammable.	3.0 to 4.0	Anesthetic effects when inhaled. Rapidly fatal at 2.4%. Dangerous for short exposure at 1.1 to 2.2%.	4,000 to 7,000	1,000	1.3 to 6.0	At bottom	Service stations, garages, storage tanks, houses.	1. Combustible gas indicator. 2. Oxygen deficiency indicator.**
Hydrogen	H <sub>2</sub>	Simple asphyxiant. Colorless, odorless, tasteless. Flammable	0.07	Acts mechanically to deprive tissues of oxygen. Does not support life.	—	—	4.0 to 74.0	At top	Manufactured gas, sludge digestion tank gas, electrolysis of water. Rarely from rock strata.	Combustible gas indicator
Hydrogen Cyanide	HCN	Faint odor of bitter almonds. Colorless gas	0.93	Slight symptoms appear upon exposure to 0.002% to 0.004%. 0.3% rapidly fatal.	—	10	6.0 to 40.0	Near top	Insecticide and rodenticide	Detector tube

**Table C-1 Hazardous Gases (Continued)**

Gas	Chemical Formula	Common Properties	Specific Gravity or Vapor Density Air = 1	Physiological Effect*	Max Safe 60 Min. Exposure ppm	Max. Safe 8 Hour Exposure ppm	Explosive Range (% by vol. in air.) Limits lower/upper	Likely Location of Highest Concentration	Most Common Sources	Simplest and Cheapest Safe Method of Testing
Hydrogen Sulfide	H <sub>2</sub> S	Irritant and poisonous volatile compound. Rotten egg odor in small concentrations. Exposure for 2 to 15 min. at 0.01% impairs sense of smell. Odor not evident at high concentrations. Colorless. Flammable.	1.19	Impairs sense of smell, rapidly as concentration increases. Death in few minutes at 0.2%. Exposure to 0.07 to 0.1% rapidly causes acute poisoning. Paralyzes respiratory center.	200 to 300	20	4.3 to 45.0	Near bottom, but may be above bottom if air is heated and highly humid.	Coal gas, petroleum, sewer gas. Fumes from blasting under some conditions. Sludge gas.	1. H <sub>2</sub> S Ampoule. 2. 5% by weight lead acetate solution.
Methane	CH <sub>4</sub>	Simple asphyxiant. Colorless, odorless, tasteless, flammable.	0.55	Acts mechanically to deprive tissues of oxygen. Does not support life.	Probably no limit, provided oxygen percent-age is sufficient for life.	—	5.0 to 15.0	At top, increasing to certain depth.	Natural gas, sludge gas, manufactured gas, sewer gas. Strata of sedimentary origin. In swamps or marshes.	1. Combustible gas indicator 2. Oxygen deficiency indicator.
Nitrogen	N <sub>2</sub>	Simple asphyxiant. Colorless, tasteless. Non-flammable. Principal constituent of air. (about 79%).	0.97	Physiologically inert.	—	—	—	Near top, but may be found near bottom.	Sewer gas, sludge gas. Also issues from some rock strata.	Oxygen deficiency indicator
Nitrogen Oxides	NO	Colorless	1.04	60 to 150 ppm cause irritation and coughing.	50	10	—	Near bottom	Industrial wastes. Common air pollutant.	NO <sub>2</sub> detector tube
	N <sub>2</sub> O	Colorless, sweet odor.	1.53	Asphyxiant.	—	—	—			
	NO <sub>2</sub>	Reddish-brown. Irritating odor. Deadly poison	1.58	100 ppm dangerous. 200 ppm fatal.	—	—	—			
Oxygen	O <sub>2</sub>	Colorless, odorless, tasteless. Supports combustion.	1.11	Normal air contains 20.8% of O <sub>2</sub> . Man can tolerate down to 12%. Minimum safe 8 hour exposure, 14 to 16%. Below 10%, dangerous to life. Below 5 to 7% probably fatal.	—	—	—	Variable at different levels.	Oxygen depletion from poor ventilation and absorption, or chemical consumption of oxygen.	Oxygen deficiency indicator
Ozone	O <sub>3</sub>	Irritant and poisonous. Strong electrical odor. Strong oxidizer. Colorless. At 1 ppm, strong sulfur-like odor.	1.66	Max. naturally occurring level is 0.04 ppm. 0.05 ppm causes irritation of eyes and nose. 1 to 10 ppm causes headache, nausea; can cause coma. Symptoms similar to radiation damage.	0.08	0.04	—	Near bottom	Where ozone is used for disinfection	Detectable odor at 0.015 ppm
Sludge Gas	—***	Mostly a simple asphyxiant. May be practically odorless, tasteless.	Variable	Will not support life.	No data. Would vary widely with composition.		5.3 to 19.3	Near top of structure	From digestion of sludge	See components
Sulfur Dioxide	SO <sub>2</sub>	Colorless, pungent odor. Suffocating, corrosive, poisonous, non-flammable.	2.26	Inflammation of the eyes. 400 to 500 ppm immediately fatal.	50 to 100	10	—	At bottom, can combine with water to form sulfurous acid.	Industrial waste, combustion, common air pollutant.	Detectable taste and odor at low concentration
Toluene	C <sub>6</sub> H <sub>12</sub> to C <sub>9</sub> H <sub>20</sub>	Colorless, benzene-like odor.	3.14	At 200-500 ppm, headache, nausea, bad taste, lassitude.	200	100	1.27 to 7.0	At bottom	Solvent	Combustible gas indicator
Turpentine	C <sub>10</sub> H <sub>16</sub>	Colorless, Characteristic odor.	4.84	Eye irritation. Headache, dizziness, nausea, irritation of the kidneys.	—	100	—	At bottom.	Solvent, used in paint	1. Detectable odor at low concentrations. 2. Combustible gas indicator.
Xylene	C <sub>8</sub> H <sub>10</sub>	Colorless, flammable	3.66	Narcotic in high concentrations. less toxic than benzene.	—	100	1.1 to 7.0	At bottom	Solvent	Combustible gas indicator

\* Percentages shown represent volume of gas in air.

\*\* For concentration over 0.3%.

\*\*\*Mostly methane and carbon dioxide with small amounts of hydrogen, nitrogen, hydrogen sulfide, and oxygen; occasionally traces of carbon monoxide.



# BLZZRD™

## Portable Refrigerated Sampler

---

### *Appendix D Replacement Parts*

#### D.1 Overview

Replacement parts are called out in the following illustrations. Refer to the tables to determine the part number for the item.

Replacement parts can be purchased on our online store:  
[store.teledyneisco.com/](http://store.teledyneisco.com/)

or by contacting Teledyne ISCO's Customer Service Department.

**Teledyne ISCO**

Customer Service Department

P.O. Box 82531

Lincoln, NE 68501 USA

Phone: (800) 228-4373

(402) 464-0231

FAX: (402) 465-3022

E-mail: [IscoInfo@teledyne.com](mailto:IscoInfo@teledyne.com)

## **BLZZRD 4 and 14 bottle parts**



60-2974-025  
Distributer arm



68-2740-027  
1.8 liter glass bottles and  
caps. Set of 8



60-2923-007  
Distributer arm nut



60-3004-193  
Cap with liner for 2.5 Gallon  
Glass bottle



68-2970-012  
950 mL poly bottles  
and cap. Set of 14

Photo Coming  
Soon

29-9048-401  
950 mL Cap with liner

## **BLZZRD 1 bottle parts**



60-9004-051  
Tube guide



68-2740-009  
2.5-gallon (9.5-liter)  
polyethylene round bottle  
with two caps



68-2700-005  
2.5-gallon (9.5-liter) glass  
round bottle with PTFE  
lined cap.



29-9001-303  
lid for 2.5 Gallon Poly bottle




60-3004-193  
Cap with liner for 2.5 Gallon  
Glass bottle

Visit our online store at: [store.teledyneisco.com](http://store.teledyneisco.com)

## BLZZRD Controller Related Replacement Parts

### Case Related Parts




60-2960-010  
Kit, BLZZRD Replacement  
Case Top



10-9080-300  
Draw Latch for Pump Case



60-2960-020  
Kit, BLZZRD Replacement  
Case Bottom



60-9004-442  
Kit, BLZZRD Dust Cap  
Replacement



60-9004-412  
Kit, 6712 BLZZRD Liquid  
Detector/pump Cover


### Pump related parts




60-9004-516  
Kit, BLZZRD-6712-  
Avalanche Pump Shaft  
Bushing and Seals



60-9004-515  
Kit, BLZZRD-6712-  
Avalanche Pump Gears  
and Plates



60-9004-511  
Kit, BLZZRD-6712-  
Avalanche Pump Roller  
Replacement



60-9004-046  
Motor Assembly



60-9004-411  
Kit, BLZZRD-6712-  
Avalanche Pump Gear Case



60-9004-203  
Pump Sensor Wiring  
Assembly

## Preventative Maintenance Related Parts



60-9004-413  
Preventative Maintenance  
Kit for BLZZRD-6712-  
Avalanche Controller



49-0001-300  
Humidity indicator card



099000200  
Desiccant Bag 8 Oz.

## Electronics Related Parts

Photo Coming  
Soon

60-2960-011  
CPU Replacement Kit



60-9004-408  
Replacement Connector  
Circuit Board Kit, BLZZRD-  
6712-Avalanche



34-0503-004  
Coin-style Lithium Battery  
(3 Volt / 560 mah)

Photo Coming  
Soon

60-2960-012  
Display/keypad  
Replacement Kit

Photo Coming  
Soon

60-2960-022  
Replacement Refrigeration  
Control Internal CBA



60-9004-443  
Replacement Module  
Harness Kit, BLZZRD-6712-  
Avalanche

## **BLZZRD Refrigerator Replacement Parts**

Photo Coming Soon	60-2960-013 Replacement Power Supply Kit	Photo Coming Soon	60-2960-016 Replacement Circulation Fan Internal Kit
Photo Coming Soon	60-2960-014 Replacement Temperature Sensor Kit	Photo Coming Soon	60-2960-017 Replacement Fan Compressor Kit
Photo Coming Soon	60-2960-015 Replacement Compressor Drive Module Kit	Photo Coming Soon	60-2960-018 Replacement Small Parts Package Kit

## **BLZZRD Top Cover and Controller Mounting Parts**



60-2974-012  
Top Cover Assembly for  
BLZZRD-Avalanche Sampler

Photo Coming  
Soon

60-2960-019  
Replacement Controller  
Mounting Center Section Kit



# BLZZRD™

## Portable Refrigerated Sampler

---

### *Appendix E Accessories*

#### E.1 Order Information

Prices available on request. Additional items appear in Appendix D, *Replacement Parts*. Many other items are available. To order any item, contact your sales representative or the factory.

**Teledyne ISCO**

P.O. Box 82531

Lincoln, NE 68501 USA

Phone: (800) 228-4373

(402) 464-0231

FAX: (402) 465-3022

E-mail: [IscoInfo@teledyne.com](mailto:IscoInfo@teledyne.com)

<input checked="" type="checkbox"/> <b>Note</b>
---

The part name listed on your order acknowledgment and invoice may be different than the item name listed here. When examining these documents, use the part number for reference.

<b>Table E-1 Accessories</b>	
<b>Item</b>	<b>Part Number</b>
<b>BLZZRD Controller Options</b>	
Program lock software for BLZZRD and 6700 Series Sampler. Factory-installed.	60-9003-532
6712 LTE North America - 6712 LTE North America cell phone modem module with magnetic mount antenna. Require customer supplied cell phone service SIM from Verizon or other using frequency bands B2, B4, B5, B13, B66 . Works with Hardware revision c.0.	60-9004-499
6712 LTE North America Non-Verizon Service - 6712 LTE North America cell phone modem module with magnetic mount antenna. Require customer supplied cell phone service SIM card (AT&T, Bell, Telus, Rogers or others using frequency bands B2 , B4, B5, B12, B14, B66 ). Works with Hardware revision c.0.	60-9004-517
6712 LTE Europe and International cellular modem module with magnetic mount antenna. Requires customer provided cell phone service SIM card (Orange, Telstra, Telus, Vodafone and others using frequency bands B1, B3, B7, B8, B20, GSM 900 and DCS 1800 ). Works with Hardware revision c.0.	60-9004-500
<b>BLZZRD Bottle Configurations</b>	
BLZZRD 14-bottle configuration. Includes 14 polypropylene 950-ml bottles with caps, two discharge tubes, bottle carrier and adapter.	68-2960-020
BLZZRD 4 -bottle configuration. Includes 4 polypropylene 5-liter bottles with caps, two discharge tubes and adapter.	68-2960-025
BLZZRD 4 -bottle configuration. Includes 4 glass 1.8-liter bottles with PTFE lined caps, distributor arm, one discharge tube, and adapter plate.	68-2960-026
BLZZRD 2.5 gal (10 liter) glass bottle with PTFE-lined caps. Includes two discharge tubes and adapter.	68-2960-021
BLZZRD 2.5 gal (10 liter) lightweight polyethylene bottle with a cap for sampling and a cap for transport. Includes two discharge tubes and adapter plate.	68-2960-024
BLZZRD 2.5 gal (10 liter) Nalgene bottle configuration. Includes two discharge tubes and adapter.	68-2960-022
BLZZRD 5 gal (20 liter) lightweight polyethylene bottle with two caps. Includes two discharge tubes and adapter plate.	68-2960-023
<b>BLZZRD Bottles and Accessories</b>	
BLZZRD Mobility Kit. Includes frame, pneumatic wheels, and handles.	60-2974-048
2.5-gallon (10-liter) polyethylene round bottle with two caps - quantity 1	68-2740-009
BLZZRD controller, includes pump tube.	60-2974-032
BLZZRD controller top cover.	60-2974-045
Connect cable for external 12 Vdc source for powering BLZZRD Sampler from an external 12 Vdc source, such as an automotive or deep-cycle marine battery. Terminates in heavy-duty battery clips.	60-5304-021
BLZZRD Refrigeration Module	60-2974-044
<b>Pump Tubes and Discharge Tubes</b>	
Discharge tube for 14 bottle and 4 bottle configurations (13 inches).	60-2973-054
Discharge tube for 2.5 gallon glass, Nalgene, and ProPak bottle (11 <sup>1</sup> / <sub>4</sub> inches).	60-2963-033
Silicone rubber pump tubing, for BLZZRD Sampler. Quantity 1.	60-9004-157



**Table E-1 Accessories (Continued)**

Item	Part Number
<b>Suction Line and Strainers</b>	
3/8 inch ID vinyl suction line with standard weighted polypropylene strainer, 25 foot length. Includes tubing coupler.	60-9004-379
3/8 inch ID PTFE suction line with protective jacket, 25 foot length	60-2703-114
3/8 inch standard weighted polypropylene strainer	60-9004-367
3/8 inch ID CPVC strainer	60-3704-066
3/8 inch stainless steel low flow strainer	69-2903-138
3/8 inch vinyl suction line - 100 feet. Includes tubing coupler.	68-1680-058
3/8 inch vinyl suction line - 500 feet. Includes two tubing couplers.	68-1680-059
Coupler for 3/8 inch vinyl suction line. Can be used with any Teledyne ISCO Sampler.	60-3709-002
<b>700 Series Modules</b>	
700 Series Modules plug into an BLZZRD or 6700 Series Sampler to expand monitoring capabilities. These environmentally-sealed modules can be interchanged in the field, and can be added to the sampler at any time.	
<b>701 pH/Temperature Module</b>	
701 Module with Stainless Steel double junction combination pH probe. Includes combination pH probe with built-in exposed temperature probe and 25 ft. cable. Also includes instruction manual. Probe mounting hardware sold separately.	68-6700-052
<b>701 pH/Temperature Accessories</b>	
Stainless Steel double-junction combination pH probe with built-in temperature probe. With 25 ft. cable.	60-9004-126
<b>720 Submerged Probe Flow Module</b>	
720 Flow Module submerged probe. Probe mounting hardware sold separately.	60-9004-030
<b>720 Submerged Probe Accessories</b>	
Submerged probe level sensor only, measures 10 ft. level range. With 25 ft. connect cable.	60-3224-002
Submerged Probe sensor extension cable, 25 ft. For additional vented separation distance between the sensor and the flow meter. More than one extension cable may be used provided that the total distance, including sensor cable length, does not exceed 75 ft.	60-3224-004
Submerged Probe sensor extension cable, 50 ft. For additional vented separation distance between the sensor and the flow meter. Total distance including sensor cable length cannot exceed 75 ft.	60-3224-005
Flume probe cap. For use in flumes built to accommodate a Teledyne ISCO probe.	60-2503-105
Sensor carrier for attaching submerged probe level sensor to Teledyne ISCO Standard Mounting Rings.	60-3204-005
<b>730 Bubbler Flow Module</b>	
730 Module with 1/8 inch x 25 ft. vinyl bubble line. Includes instruction manual and coupon for free Teledyne ISCO Open Channel Flow Measurement Handbook. Bubble tube mounting hardware sold separately.	68-6700-050

<b>Table E-1 Accessories (Continued)</b>	
<b>Item</b>	<b>Part Number</b>
<b>730 Bubbler Accessories</b>	
1/8 inch x 100 ft. vinyl tubing. For use as a flow meter bubble line or VOC discharge tubing.	60-1700-003
Stainless Steel Bubble Tube - 4 ft. length. For use with 1/8 inch ID vinyl bubble line.	60-1873-043
Sensor carrier for attaching bubble line to Teledyne ISCO Standard Mounting Rings.	60-3204-007
<b>750 Area Velocity Flow Module</b>	
750 Module and Low Profile Area Velocity Sensor with 25 ft. cable, measures 10 ft. level range. Includes instruction manual and coupon for free Teledyne ISCO Open Channel Flow Measurement Handbook. Probe mounting hardware sold separately.	60-9004-031
<b>750 Area Velocity Accessories</b>	
Low Profile Area Velocity Sensor with 25 ft. cable, measures 10 ft. level range.	60-3254-021
Area velocity sensor extension cable for Models 4150/4250 or BLZZRD/6712 with 750 module. Cut to length with maximum length of 50 ft.	60-5304-889
Sensor carrier for attaching Low Profile Area Velocity Sensor to Teledyne ISCO Standard Mounting Rings.	60-3204-029
<b>780 Analog Input Module</b>	
780 analog input module. Includes 10 ft. cable and instruction manual.	60-9004-040
<b>Other Parameter Measurement Equipment</b>	
674 Rain Gauge. Tips every 0.01 inch of rainfall. Tipping bucket rain gauge with 50 ft. cable and connector for use with a 4100 Series Flow Logger, 4200 Series Flow Meter, or 6700 Series Sampler. Includes instruction manual.	60-3284-001
674 Rain Gauge. Tips every 0.1 mm of rainfall. Tipping bucket rain gauge with 50 ft. cable and connector for use with a 4100 Series Flow Logger, 4200 Series Flow Meter, or 6700 Series Sampler. Includes instruction manual.	60-3284-006
Model 1640 Sampler Actuator. Includes instruction manual.	60-1644-000
4-20 mA input interface for 6700 Series, BLZZRD, or GLS Sampler. Converts an analog 4-20 mA flow proportional signal to a compatible flow proportional input signal.	60-3704-075
<b>SDI-12 Connect Cables for the BLZZRD</b>	
BLZZRD/6712 Universal 10 m (33 ft) cable. For SDI-12, rain, and I/O. Includes connector on one end and pigtailed on the other end.	60-9004-611

# BLZZRD™

## Portable Refrigerated Sampler

---

### *Appendix F Battery Selection Guide*

#### F.1 Introduction

This section provides information to assist with selecting a power source for a 12 Vdc battery-powered BLZZRD sampler. Teledyne ISCO recommends deep-cycle marine batteries, although any 12 Vdc power source will work provided that it can meet the current demands of the refrigeration system.

#### F.2 Determining the Power Consumption

The sampler requires continuous 12 Vdc power for several purposes:

- an idle current while the sampler is inhibited, waiting for the next sample, or after it has finished its sampling program
- logging temperature readings at one-minute intervals
- operating the pump
- cooling the samples.

The first three are negligible when compared to the energy required to cool the liquid as it is collected and to maintain the samples at 3 °C.

The amount of energy required to cool the samples correlates to ambient air temperature. The higher the air temperature, the more the sampler must operate the refrigeration system.

The equation below may be used as a guide for selecting a battery that will meet the requirements of your sampling program.

$$\left( \frac{V_c \times A_c}{6} \right) + (D_m \times A_m) = \text{Expected power consumption, Amp-hours}$$

Where:  $V_c$  = Total sample volume to cool, in liters  
 $A_c$  = Amp-hours to cool sample liquid from 20 °C (Fig F-1)  
 $D_m$  = Number of days sample volume must be maintained  
 $A_m$  = Amp-hour per day factor for maintenance (Fig F-2)

Before using the equation you must know:

- the total sample volume that will be collected
- the number of days that the sampler must keep the sample cooled to 3 °C
- and the average ambient temperature.

The first group in the equation calculates the amp-hours required to cool the sample liquid as it is collected. The total sample volume,  $V_c$ , can be determined by the programmed number of samples multiplied by the sample volume. If the sampler is programmed for continuous sampling,  $V_c$  would be the

programmed bottle volume, provided that bottle-full detection is being used. Use Figure F-1 to find  $A_c$  by determining the amp-hours for the average ambient temperature at your sampling site.

Once the sample liquid has been cooled, more energy is required to keep it cool. The second group in the equation may be used to calculate the amp-hours required each day to keep the sample at 3 °C. The number of days,  $D_m$ , is the days from the first collected sample to when the sampler is no longer expected to keep the sample cool. Use Figure F-2 to find  $A_m$ .

The following scenario illustrates how to use the equation. Consider an sampler that is inhibited while waiting for an external device to indicate that a measured condition, such as a rain event. The average ambient temperature is 30 °C (86 °F). When the external device clears the inhibit, the sampler will collect sample volumes of 250 mL, once an hour, for 24 hours. After running the program, it will take another day to retrieve the sampler and bring the composite sample to the lab.

From this example, we know that  $V_c$  is 6 liters,  $D_m$  is two days (one to collect the samples, another to retrieve it), and the ambient temperature is 30 °C. At this ambient temperature, Figure F-1 tells us  $A_c$  is 21, and Figure F-2 tells us  $A_m$  is 41. The equation would be:

$$\left(\frac{V_c \times A_c}{6}\right) + (D_m \times A_m) = \text{Expected power consumption, Amp-hours}$$
$$\left(\frac{6 \times 21}{6}\right) + (2 \times 41) = 103 \text{ Amp-hours}$$

Using the equation we find operating the sampler under these conditions would need a fully-charged battery with a capacity of at least 103 amp-hours.

As previously mentioned, standby current is negligible. In the example above, the sampler could be waiting several days for the rain event. Because the refrigeration system is not activated until the first sample is taken, the number of days the sampler is inhibited is generally of no concern.

Figure F-1 assumes that the liquid temperature at the sampling point is at room temperature (68 to 77 °F or 20 to 25 °C). Increase or decrease the expected battery capacity accordingly if the actual liquid temperature falls outside of this temperature range.

Other factors, such as high suction head heights (over 3 m or 10 feet) or the amount of direct sunlight will also affect the rated current demand on the battery.

 **Note**

The depth of discharge often affects the battery life—or more specifically—the number of charge/discharge cycles. A deep cycle battery that is discharged to 50% of its total capacity can be expected to deliver nearly twice the number of cycles than a battery discharged to a depth of 80%.




Figure F-1 Amp-hours required to cool 6 liters of sample liquid from 20 °C ( $A_c$ )




Figure F-2 Amp-hours per day required to maintain sample liquid at 3 °C ( $A_m$ )

